



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100721\
 Data File : BN016827.D
 Acq On : 08 Oct 2021 04:31
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100721.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Oct 08 01:39:08 2021






AutoFind: Scans 2444, 2445, 2446; Background Corrected with Scan 2437

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.6	51448	PASS
68	69	0.00	2	1.8	967	PASS
69	198	0.00	100	32.2	54173	PASS
70	69	0.00	2	0.2	119	PASS
127	198	10	80	44.6	74976	PASS
197	198	0.00	2	0.6	1013	PASS
198	198	100	100	100.0	168224	PASS
199	198	5	9	6.6	11080	PASS
275	198	10	60	29.8	50192	PASS
365	198	1	100	4.3	7165	PASS
441	198	0.01	100	20.5	34453	PASS
442	442	50	100	100.0	226539	PASS
443	442	15	24	20.2	45800	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100721\
Data File : BN016827.D
Acq On : 08 Oct 2021 04:31
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Oct 08 06:41:26 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 30 18:14:09 2021
Response via : Initial Calibration



TIC: BN016827.D\data.ms

(70) Pentachlorophenol (C)

16.651min (-0.023) 29301.17 ng




response 226499

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.98
264.00	61.60	63.04
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100721\
Data File : BN016827.D
Acq On : 08 Oct 2021 04:31
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Oct 08 06:41:26 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 30 18:14:09 2021
Response via : Initial Calibration



TIC: BN016827.D\data.ms

(77) Benzidine

19.269min (-0.023) 0.00 ng

response 1222328

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.19
183.00	13.20	11.83
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/7/2021	BNA_N	<u>BN016827.D</u>
Compound Name	Response	Retention Time
DDT	621993	20.51
DDD	9129	20.069
DDE	0	19.58
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
9129	631122	1.45