

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100725\
 Data File : BN038061.D
 Acq On : 08 Oct 2025 00:54
 Operator : RC/JU
 Sample : PB169992BL
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB169992BL

Quant Time: Oct 11 03:09:44 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.804	152	7319	0.400	ng	-0.02
7) Naphthalene-d8	10.605	136	19498	0.400	ng	-0.02
13) Acenaphthene-d10	14.451	164	8755	0.400	ng	-0.02
19) Phenanthrene-d10	17.210	188	14397	0.400	ng	-0.01
29) Chrysene-d12	21.384	240	8061	0.400	ng	#-0.02
35) Perylene-d12	23.703	264	8249	0.400	ng	#-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.370	112	6448	0.386	ng	-0.01
5) Phenol-d6	6.959	99	7839	0.357	ng	-0.01
8) Nitrobenzene-d5	8.950	82	6740	0.453	ng	-0.02
11) 2-Methylnaphthalene-d10	12.197	152	9519	0.351	ng	-0.02
14) 2,4,6-Tribromophenol	15.957	330	553	0.167	ng	-0.01
15) 2-Fluorobiphenyl	13.083	172	16370	0.457	ng	-0.01
27) Fluoranthene-d10	19.238	212	11085	0.306	ng	-0.01
31) Terphenyl-d14	19.838	244	8270	0.515	ng	-0.01

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100725\
Data File : BN038061.D
Acq On : 08 Oct 2025 00:54
Operator : RC/JU
Sample : PB169992BL
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB169992BL

Quant Time: Oct 11 03:09:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 24 11:00:01 2025
Response via : Initial Calibration

