

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100725\
 Data File : BN038060.D
 Acq On : 08 Oct 2025 00:17
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 08 02:03:30 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	154#	-0.02
2	1,4-Dioxane	0.406	0.417	-2.7	156#	-0.01
3	n-Nitrosodimethylamine	0.540	0.520	3.7	155#	0.00
4 S	2-Fluorophenol	0.913	0.868	4.9	156#	-0.01
5 S	Phenol-d6	1.199	1.101	8.2	151#	-0.01
6	bis(2-Chloroethyl)ether	1.113	1.083	2.7	154#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	152#	-0.02
8 S	Nitrobenzene-d5	0.305	0.297	2.6	155#	-0.02
9	Naphthalene	1.102	1.064	3.4	154#	-0.02
10	Hexachlorobutadiene	0.188	0.186	1.1	154#	-0.03
11 SURR	2-Methylnaphthalene-d10	0.556	0.498	10.4	151#	-0.02
12	2-Methylnaphthalene	0.720	0.665	7.6	151#	-0.03
13 I	Acenaphthene-d10	1.000	1.000	0.0	142	-0.02
14 S	2,4,6-Tribromophenol	0.152	0.111	27.0#	116	-0.02
15 S	2-Fluorobiphenyl	1.638	1.628	0.6	148	-0.02
16	Acenaphthylene	1.816	1.736	4.4	148	-0.02
17	Acenaphthene	1.294	1.201	7.2	142	-0.02
18	Fluorene	1.565	1.301	16.9	127	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	122	-0.02
20	4,6-Dinitro-2-methylphenol	0.057	0.034	40.4#	96	-0.01
21	4-Bromophenyl-phenylether	0.261	0.253	3.1	126	-0.02
22	Hexachlorobenzene	0.316	0.330	-4.4	132	-0.02
23	Atrazine	0.199	0.139	30.2#	99	-0.02
24	Pentachlorophenol	0.108	0.091	15.7	121	-0.01
25	Phenanthrene	1.316	1.245	5.4	124	-0.01
26	Anthracene	1.142	1.017	10.9	121	-0.02
27 SURR	Fluoranthene-d10	1.006	0.754	25.0#	104	-0.02
28	Fluoranthene	1.450	1.117	23.0	104	-0.02
29 I	Chrysene-d12	1.000	1.000	0.0	91	-0.02
30	Pyrene	1.579	1.713	-8.5	102	-0.02
31 S	Terphenyl-d14	0.797	0.800	-0.4	97	-0.02
32	Benzo(a)anthracene	1.398	1.328	5.0	93	-0.02
33	Chrysene	1.512	1.476	2.4	92	-0.02
34	Bis(2-ethylhexyl)phthalate	0.553	0.533	3.6	99	-0.02
35 I	Perylene-d12	1.000	1.000	0.0	95	-0.02
36	Indeno(1,2,3-cd)pyrene	1.828	2.039	-11.5	117	-0.04
37	Benzo(b)fluoranthene	1.725	1.691	2.0	98	-0.02
38	Benzo(k)fluoranthene	1.736	1.596	8.1	95	-0.03
39 C	Benzo(a)pyrene	1.364	1.359	0.4	105	-0.02
40	Dibenzo(a,h)anthracene	1.427	1.571	-10.1	115	-0.04
41	Benzo(g,h,i)perylene	1.499	1.675	-11.7	114	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4