

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100725\
 Data File : BN038060.D
 Acq On : 08 Oct 2025 00:17
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 08 02:03:30 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	154	-0.02
2	1,4-Dioxane	0.400	0.411	-2.7	156	-0.01
3	n-Nitrosodimethylamine	0.400	0.385	3.8	155	0.00
4 S	2-Fluorophenol	0.400	0.380	5.0	156	-0.01
5 S	Phenol-d6	0.400	0.367	8.3	151	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.389	2.8	154	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	152	-0.02
8 S	Nitrobenzene-d5	0.400	0.389	2.8	155	-0.02
9	Naphthalene	0.400	0.386	3.5	154	-0.02
10	Hexachlorobutadiene	0.400	0.395	1.3	154	-0.03
11 SURR	2-Methylnaphthalene-d10	0.400	0.359	10.3	151	-0.02
12	2-Methylnaphthalene	0.400	0.369	7.8	151	-0.03
13 I	Acenaphthene-d10	0.400	0.400	0.0	142	-0.02
14 S	2,4,6-Tribromophenol	0.400	0.293	26.8#	116	-0.02
15 S	2-Fluorobiphenyl	0.400	0.398	0.5	148	-0.02
16	Acenaphthylene	0.400	0.382	4.5	148	-0.02
17	Acenaphthene	0.400	0.371	7.3	142	-0.02
18	Fluorene	0.400	0.332	17.0	127	-0.02
19 I	Phenanthrene-d10	0.400	0.400	0.0	122	-0.02
20	4,6-Dinitro-2-methylphenol	0.400	0.337	15.8	96	-0.01
21	4-Bromophenyl-phenylether	0.400	0.388	3.0	126	-0.02
22	Hexachlorobenzene	0.400	0.418	-4.5	132	-0.02
23	Atrazine	0.400	0.281	29.8#	99	-0.02
24	Pentachlorophenol	0.400	0.337	15.8	121	-0.01
25	Phenanthrene	0.400	0.378	5.5	124	-0.01
26	Anthracene	0.400	0.356	11.0	121	-0.02
27 SURR	Fluoranthene-d10	0.400	0.300	25.0#	104	-0.02
28	Fluoranthene	0.400	0.308	23.0	104	-0.02
29 I	Chrysene-d12	0.400	0.400	0.0	91	-0.02
30	Pyrene	0.400	0.434	-8.5	102	-0.02
31 S	Terphenyl-d14	0.400	0.402	-0.5	97	-0.02
32	Benzo(a)anthracene	0.400	0.380	5.0	93	-0.02
33	Chrysene	0.400	0.390	2.5	92	-0.02
34	Bis(2-ethylhexyl)phthalate	0.400	0.385	3.8	99	-0.02
35 I	Perylene-d12	0.400	0.400	0.0	95	-0.02
36	Indeno(1,2,3-cd)pyrene	0.400	0.446	-11.5	117	-0.04
37	Benzo(b)fluoranthene	0.400	0.392	2.0	98	-0.02
38	Benzo(k)fluoranthene	0.400	0.368	8.0	95	-0.03
39 C	Benzo(a)pyrene	0.400	0.398	0.5	105	-0.02
40	Dibenzo(a,h)anthracene	0.400	0.440	-10.0	115	-0.04
41	Benzo(g,h,i)perylene	0.400	0.447	-11.7	114	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4