

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100725\
 Data File : BN038049.D
 Acq On : 07 Oct 2025 16:54
 Operator : RC/JU
 Sample : Q3236-02
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW7-SP201-20250926

Quant Time: Oct 07 17:20:17 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.804	152	8046	0.400	ng	-0.02
7) Naphthalene-d8	10.605	136	21585	0.400	ng	-0.02
13) Acenaphthene-d10	14.452	164	9666	0.400	ng	-0.02
19) Phenanthrene-d10	17.211	188	16187	0.400	ng	-0.01
29) Chrysene-d12	21.384	240	11198	0.400	ng	#-0.02
35) Perylene-d12	23.706	264	10794	0.400	ng	#-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.377	112	2867	0.156	ng	0.00
5) Phenol-d6	6.959	99	2073	0.086	ng	-0.01
8) Nitrobenzene-d5	8.950	82	7262	0.441	ng	-0.02
11) 2-Methylnaphthalene-d10	12.197	152	9905	0.330	ng	-0.02
14) 2,4,6-Tribromophenol	15.957	330	874	0.238	ng	-0.01
15) 2-Fluorobiphenyl	13.083	172	17228	0.435	ng	-0.01
27) Fluoranthene-d10	19.239	212	14300	0.351	ng	-0.01
31) Terphenyl-d14	19.838	244	12018	0.539	ng	-0.01
Target Compounds						Qvalue
34) Bis(2-ethylhexyl)phtha...	21.304	149	1397	0.090	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100725\
Data File : BN038049.D
Acq On : 07 Oct 2025 16:54
Operator : RC/JU
Sample : Q3236-02
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20250926


Quant Time: Oct 07 17:20:17 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 24 11:00:01 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.804 min Scan# 61
Delta R.T. -0.021 min
Lab File: BN038049.D
Acq: 07 Oct 2025 16:54

Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20250926







Tgt Ion:152 Resp: 8046
Ion Ratio Lower Upper
152 100
150 155.2 121.0 181.4
115 59.0 47.4 71.0



#4
2-Fluorophenol
Concen: 0.156 ng
RT: 5.377 min Scan# 318
Delta R.T. -0.007 min
Lab File: BN038049.D
Acq: 07 Oct 2025 16:54


Tgt Ion:112 Resp: 2867
Ion Ratio Lower Upper
112 100
64 55.6 44.6 66.8
63 31.7 24.9 37.3








#5
Phenol-d6
Concen: 0.086 ng
RT: 6.959 min Scan# 539
Delta R.T. -0.014 min
Lab File: BN038049.D
Acq: 07 Oct 2025 16:54

Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20250926







Tgt Ion:	99	Resp:	2073
Ion	Ratio	Lower	Upper
99	100		
42	24.2	17.2	25.8
71	39.4	27.3	40.9



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.605 min Scan# 939
Delta R.T. -0.021 min
Lab File: BN038049.D
Acq: 07 Oct 2025 16:54

Tgt Ion:	136	Resp:	21585
Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.0	13.4
54	7.2	5.8	8.8
68	5.6	4.3	6.5









#8
Nitrobenzene-d5
Concen: 0.441 ng
RT: 8.950 min Scan# 786
Delta R.T. -0.021 min
Lab File: BN038049.D
Acq: 07 Oct 2025 16:54




Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20250926


Tgt Ion: 82 Resp: 7262
Ion Ratio Lower Upper
82 100
128 41.9 34.8 52.2
54 56.1 42.2 63.2



#11
2-Methylnaphthalene-d10
Concen: 0.330 ng
RT: 12.197 min Scan# 1159
Delta R.T. -0.020 min
Lab File: BN038049.D
Acq: 07 Oct 2025 16:54

Tgt Ion: 152 Resp: 9905
Ion Ratio Lower Upper
152 100
151 21.5 17.4 26.2





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.452 min Scan# 14

Delta R.T. -0.021 min

Lab File: BN038049.D

Acq: 07 Oct 2025 16:54

Instrument :

BNA_N

ClientSampleId :

RW7-SP201-20250926





Tgt Ion:164 Resp: 9666

Ion Ratio Lower Upper

164 100

162 107.2 84.8 127.2

160 58.3 46.1 69.1



#14

2,4,6-Tribromophenol

Concen: 0.238 ng

RT: 15.957 min Scan# 1589

Delta R.T. -0.012 min

Lab File: BN038049.D

Acq: 07 Oct 2025 16:54




Tgt Ion:330 Resp: 874


Ion Ratio Lower Upper

330 100

332 97.3 77.2 115.8

141 45.5 37.2 55.8









#15
2-Fluorobiphenyl
Concen: 0.435 ng
RT: 13.083 min Scan# 1327
Delta R.T. -0.011 min
Lab File: BN038049.D
Acq: 07 Oct 2025 16:54




Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20250926


Tgt Ion:172	Resp:	17228
Ion Ratio	Lower	Upper
172	100	
171	34.6	28.6 43.0
170	22.9	19.7 29.5



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.211 min Scan# 1690
Delta R.T. -0.012 min
Lab File: BN038049.D
Acq: 07 Oct 2025 16:54

Tgt Ion:188	Resp:	16187
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0 0.0
80	8.3	7.6 11.4









#27
Fluoranthene-d10
Concen: 0.351 ng
RT: 19.239 min Scan# 1951
Delta R.T. -0.014 min
Lab File: BN038049.D
Acq: 07 Oct 2025 16:54




Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20250926


Tgt Ion:212 Resp: 14300
Ion Ratio Lower Upper
212 100
106 17.5 12.6 19.0
104 9.5 7.4 11.0



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.384 min Scan# 2270
Delta R.T. -0.018 min
Lab File: BN038049.D
Acq: 07 Oct 2025 16:54

Tgt Ion:240 Resp: 11198
Ion Ratio Lower Upper
240 100
120 20.2 11.4 17.2#
236 29.3 23.0 34.6









#31
 Terphenyl-d14
 Concen: 0.539 ng
 RT: 19.838 min Scan# 2080
 Delta R.T. -0.014 min
 Lab File: BN038049.D
 Acq: 07 Oct 2025 16:54




Instrument :
 BNA_N
 ClientSampleId :
 RW7-SP201-20250926


Tgt Ion:244 Resp: 12018
 Ion Ratio Lower Upper
 244 100
 212 9.3 7.6 11.4
 122 16.4 13.9 20.9



#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.090 ng
 RT: 21.304 min Scan# 2261
 Delta R.T. -0.018 min
 Lab File: BN038049.D
 Acq: 07 Oct 2025 16:54


Tgt Ion:149 Resp: 1397
 Ion Ratio Lower Upper
 149 100
 167 24.9 20.4 30.6
 279 6.0 2.4 3.6#





#35
Perylene-d12
Concen: 0.400 ng
RT: 23.706 min Scan# 2836
Delta R.T. -0.020 min
Lab File: BN038049.D
Acq: 07 Oct 2025 16:54

Instrument :
BNA_N
ClientSampleId :
RW7-SP201-20250926



Tgt Ion:264 Resp: 10794

Ion	Ratio	Lower	Upper
264	100		
260	27.4	21.5	32.3
265	96.6	53.8	80.6

