

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100824\
 Data File : BN034477.D
 Acq On : 08 Oct 2024 18:36
 Operator : RC/JU
 Sample : P4226-08
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW11-MW01I-20240924

Quant Time: Oct 09 00:55:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 26 14:44:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.380	152	5698	0.400	ng	-0.01
7) Naphthalene-d8	10.137	136	16454	0.400	ng	-0.01
13) Acenaphthene-d10	14.026	164	9269	0.400	ng	-0.02
19) Phenanthrene-d10	16.790	188	15260	0.400	ng	-0.01
29) Chrysene-d12	21.015	240	11081	0.400	ng	0.00
35) Perylene-d12	23.125	264	12648	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.025	112	2564	0.159	ng	0.00
5) Phenol-d6	6.593	99	1250	0.066	ng	0.00
8) Nitrobenzene-d5	8.536	82	3180	0.253	ng	0.00
11) 2-Methylnaphthalene-d10	11.738	152	7306	0.301	ng	-0.02
14) 2,4,6-Tribromophenol	15.549	330	1339	0.319	ng	0.00
15) 2-Fluorobiphenyl	12.647	172	10346	0.274	ng	-0.01
27) Fluoranthene-d10	18.836	212	15565	0.404	ng	0.00
31) Terphenyl-d14	19.454	244	10654	0.481	ng	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100824\
Data File : BN034477.D
Acq On : 08 Oct 2024 18:36
Operator : RC/JU
Sample : P4226-08
Misc :
ALS Vial : 48 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW11-MW01I-20240924

Quant Time: Oct 09 00:55:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 26 14:44:02 2024
Response via : Initial Calibration

