

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100918\
 Data File : BN002985.D
 Acq On : 08 Oct 2018 14:06
 Operator : MJ/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Oct 08 18:10:54 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN100918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 08 18:09:03 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	87	0.00
3	Naphthalene	1.026	1.095	-6.7	95	0.00
4 SURR	2-Methylnaphthalene-d10	0.620	0.671	-8.2	94	0.00
5	2-Methylnaphthalene	0.707	0.768	-8.6	94	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	84	0.00
7	Acenaphthylene	1.830	2.003	-9.5	94	0.00
8 C	Acenaphthene	1.389	1.538	-10.7	94	0.00
9	Fluorene	1.591	1.748	-9.9	93	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	86	0.00
11	Pentachlorophenol	0.039	0.034	12.8	110	0.00
12	Phenanthrene	1.115	1.213	-8.8	95	0.00
13	Anthracene	1.053	1.154	-9.6	96	0.00
14 SURR	Fluoranthene-d10	1.144	1.262	-10.3	98	0.00
15 C	Fluoranthene	1.308	1.445	-10.5	98	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
17	Pyrene	1.554	1.633	-5.1	99	0.00
18	Benzo(a)anthracene	1.287	1.402	-8.9	107	0.00
19	Chrysene	1.307	1.428	-9.3	107	0.00
20 I	Perylene-d12	1.000	1.000	0.0	98	0.00
21	Benzo(b)fluoranthene	1.441	1.578	-9.5	113	0.00
22	Benzo(k)fluoranthene	1.493	1.628	-9.0	110	0.00
23 C	Benzo(a)pyrene	1.291	1.418	-9.8	110	0.00
24	Indeno(1,2,3-cd)pyrene	1.484	1.551	-4.5	103	0.00
25	Dibenzo(a,h)anthracene	1.188	1.244	-4.7	103	0.00
26	Benzo(g,h,i)perylene	1.251	1.294	-3.4	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0