


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN100918\
Data File : BN002984.D
Acq On : 08 Oct 2018 12:17
Operator : MJ/SJ
Sample : DFTPP33
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP33

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN100918.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Oct 09 16:18:11 2018



AutoFind: Scans 2460, 2461, 2462; Background Corrected with Scan 2454

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.1	71952	PASS
68	69	0.00	2	1.2	955	PASS
69	198	0.00	100	35.6	77357	PASS
70	69	0.00	2	0.5	392	PASS
127	198	10	80	43.3	94232	PASS
197	198	0.00	2	0.4	848	PASS
198	198	100	100	100.0	217450	PASS
199	198	5	9	6.8	14868	PASS
275	198	10	60	26.5	57650	PASS
365	198	1	100	3.1	6739	PASS
441	198	0.01	100	13.6	29626	PASS
442	198	50	100	88.6	192645	PASS
443	442	15	24	19.5	37517	PASS

m/z	Abundance
38.00	737.0
39.00	5295.0
40.00	598.0
40.90	405.0
44.00	1824.0
49.00	421.0
50.00	16127.0
51.00	65304.0
52.00	3680.0
55.00	568.0
56.00	2253.0
57.00	4813.0
61.00	998.0
61.90	1044.0
63.00	2595.0
65.00	1850.0
69.00	70792.0
70.00	350.0
73.00	968.0
74.00	6546.0
75.00	10190.0
76.00	3985.0
77.00	79160.0
78.00	5728.0
79.00	4547.0
80.00	3537.0
81.00	5271.0
82.00	1302.0
83.00	1248.0
85.00	1152.0
85.90	1546.0
87.00	706.0
87.90	369.0
90.90	1185.0
91.90	1730.0
92.90	8105.0
93.90	559.0
96.00	574.0
98.00	6633.0
99.00	4768.0
100.00	409.0
101.00	2910.0
103.00	900.0
104.00	1801.0
105.00	1828.0
107.00	20728.0
108.00	3719.0
110.00	43792.0
111.00	6223.0
111.90	707.0
116.00	1144.0
117.00	15688.0
118.00	1224.0
121.90	1447.0
123.00	2323.0
123.90	851.0
125.00	983.0
127.00	81280.0
128.00	6393.0
129.00	30352.0
130.00	2437.0
131.00	465.0
132.00	330.0
133.90	846.0
135.00	2315.0
135.90	1001.0
137.00	1340.0
139.90	512.0
141.00	3485.0
141.90	1585.0
143.00	781.0
145.90	536.0
146.90	2084.0
148.00	3910.0
148.90	940.0
151.10	479.0
152.00	389.0
153.00	1362.0
154.00	825.0
155.00	1970.0
156.00	3585.0
157.00	586.0
157.90	861.0
159.00	563.0
160.10	1114.0
161.00	1711.0
161.90	499.0
163.10	391.0
164.90	1458.0
166.00	1211.0
167.00	7067.0

Instrument :
BNA_N
ClientSampleId :
DFTPP33

167.90	3488.0
168.90	682.0
170.80	354.0
171.90	788.0
172.90	1084.0
174.00	1556.0
175.00	2835.0
176.00	725.0
176.90	1675.0
177.90	616.0
178.90	5571.0
180.00	4090.0
181.00	1929.0
181.90	460.0
183.90	474.0
184.90	2302.0
186.00	21264.0
186.90	6060.0
188.00	573.0
189.00	1502.0
191.00	662.0
192.00	1763.0
193.00	2064.0
194.10	302.0
195.90	5155.0
197.90	169984.0
198.90	12056.0
200.00	851.0
201.60	864.0
203.00	839.0
204.00	5407.0
205.00	9145.0
206.00	41488.0
207.00	6729.0
208.00	1426.0
209.00	510.0
210.10	773.0
210.90	1749.0
214.80	405.0
215.90	980.0
216.90	10871.0
218.00	1196.0
221.00	7619.0
222.90	2511.0
224.00	21632.0
225.00	6175.0
227.00	8899.0
228.00	1240.0
228.90	1958.0
231.00	963.0
234.00	681.0
234.80	691.0
236.00	522.0
236.90	843.0
239.00	338.0
240.90	496.0
242.00	1124.0
243.00	1285.0
244.00	18032.0
245.00	2529.0
245.90	3128.0
246.80	650.0
248.90	628.0
253.10	331.0
255.00	94080.0
256.00	13799.0
257.00	1144.0
258.00	4578.0
258.90	767.0
264.90	1779.0
265.90	434.0
273.00	2626.0
274.00	7355.0
275.00	43712.0
276.00	5952.0
277.00	2930.0
278.00	505.0
281.10	318.0
283.00	496.0
284.00	362.0
285.00	483.0
292.90	564.0
296.00	9723.0
297.00	1637.0
303.00	1398.0
304.00	321.0
313.90	471.0
315.00	1071.0
316.00	964.0
323.00	3740.0
324.00	683.0
326.90	625.0
332.10	343.0
333.00	486.0
334.00	2303.0

Instrument :

BNA_N

ClientSampleId :

DFTPP33

335.00	530.0
340.90	430.0
345.90	771.0
352.00	1131.0
353.00	1046.0
354.00	1236.0
355.00	364.0
364.90	4544.0
365.90	636.0
372.00	1832.0
372.90	480.0
382.80	579.0
402.00	611.0
403.00	1098.0
404.00	475.0
421.00	1147.0
422.00	1062.0
423.10	7139.0
424.00	1775.0
441.00	17680.0
442.00	115088.0
443.00	23824.0
444.00	2227.0

Instrument :
BNA_N
ClientSampleId :
DFTPP33

m/z	Abundance
37.00	363.0
38.00	870.0
39.00	6515.0
40.00	618.0
41.00	392.0
44.00	1616.0
44.90	303.0
49.00	535.0
50.00	20640.0
51.00	84296.0
52.00	4417.0
55.10	491.0
56.00	2622.0
57.00	6278.0
61.00	1044.0
62.00	1364.0
63.00	3533.0
64.00	579.0
65.00	1730.0
68.00	1669.0
69.00	89080.0
70.00	521.0
73.00	930.0
74.00	9202.0
75.00	13598.0
76.00	4250.0
77.00	99680.0
78.00	7123.0
79.00	6658.0
79.90	5152.0
81.00	6943.0
82.00	1876.0
83.00	1679.0
85.00	1716.0
86.00	1592.0
87.00	1029.0
87.90	339.0
90.90	1875.0
91.90	1478.0
93.00	10142.0
94.00	738.0
94.80	308.0
95.90	556.0
98.00	7597.0
99.00	6398.0
100.00	544.0
101.00	3827.0
102.90	1546.0
104.00	2313.0
105.00	2062.0
106.10	767.0
107.00	30488.0
108.00	4788.0
109.10	956.0
110.00	59296.0
111.00	8814.0
111.90	1014.0
112.90	375.0
116.00	1631.0
117.00	19936.0
118.00	1536.0
119.80	356.0
122.00	2243.0
123.00	3253.0
124.00	1282.0
124.90	1511.0
127.00	110112.0
128.00	8259.0
129.00	39080.0
130.00	3958.0
130.90	777.0
131.90	405.0
132.90	314.0
133.90	1227.0
135.00	3478.0
135.90	1252.0
137.00	1708.0
139.90	635.0
141.00	5603.0
142.00	1725.0
142.90	1355.0
144.00	436.0
146.00	1151.0
146.90	2857.0
148.00	5569.0
149.00	1208.0
150.00	428.0
151.00	678.0
152.00	624.0
153.00	1894.0
154.00	1352.0

Instrument :
BNA_N
ClientSampleId :
DFTPP33

155.00	3054.0
156.00	4806.0
157.00	834.0
157.90	883.0
158.90	795.0
160.00	1812.0
161.00	2554.0
162.00	781.0
163.90	364.0
164.90	1796.0
166.00	1676.0
167.00	10252.0
168.00	4533.0
168.90	1005.0
170.00	400.0
171.10	575.0
171.90	849.0
172.90	1421.0
173.90	1954.0
175.00	4624.0
175.90	1260.0
176.90	2100.0
177.90	498.0
178.90	8517.0
180.00	5789.0
181.00	2468.0
181.90	410.0
183.90	602.0
185.00	4158.0
186.00	32552.0
187.00	8793.0
188.00	895.0
189.00	2082.0
189.90	356.0
191.00	1076.0
191.90	2514.0
193.00	2908.0
194.00	577.0
195.10	408.0
195.90	6875.0
197.10	1317.0
197.90	256576.0
198.90	17384.0
200.00	1369.0
201.40	1464.0
203.00	1842.0
204.00	8311.0
205.00	13974.0
206.00	63232.0
207.00	9051.0
207.90	2099.0
209.00	661.0
210.30	1007.0
211.00	2310.0
211.70	443.0
214.90	622.0
216.00	1155.0
216.90	16672.0
218.00	2216.0
221.00	11063.0
223.00	3555.0
224.00	33984.0
225.00	7851.0
225.90	1162.0
227.00	13288.0
227.90	1751.0
229.00	2398.0
231.00	1409.0
234.00	938.0
235.00	1043.0
235.90	628.0
236.90	1256.0
238.90	417.0
240.00	425.0
240.90	795.0
242.00	1360.0
243.00	1705.0
244.00	27704.0
245.00	3425.0
246.00	4525.0
247.00	1249.0
248.80	863.0
252.90	467.0
255.00	142464.0
256.00	20920.0
257.00	1636.0
258.00	6760.0
259.00	1073.0
265.00	3285.0
265.90	534.0
273.00	4453.0
274.00	11567.0
275.00	66088.0
276.00	9188.0
277.00	5340.0

Instrument :
BNA_N
ClientSampleId :
DFTPP33

277.90	835.0
283.00	739.0
284.00	420.0
284.90	747.0
293.00	1403.0
294.90	460.0
296.00	16488.0
297.00	2190.0
300.90	371.0
301.90	329.0
303.00	1865.0
304.00	412.0
310.00	331.0
314.00	893.0
315.00	1758.0
316.00	1198.0
321.00	610.0
322.00	346.0
323.00	5774.0
323.90	961.0
326.90	1183.0
328.00	585.0
332.90	495.0
334.00	3983.0
335.00	959.0
341.10	818.0
345.90	1537.0
352.00	1733.0
352.90	1006.0
354.00	1791.0
355.10	631.0
364.90	7427.0
365.90	1306.0
370.90	558.0
372.00	3195.0
373.00	802.0
382.90	703.0
389.90	360.0
391.10	394.0
392.00	368.0
401.90	1254.0
403.00	2035.0
404.00	976.0
421.00	1577.0
422.00	1568.0
423.00	11859.0
424.00	2987.0
424.90	347.0
441.00	32904.0
442.00	211328.0
443.00	42200.0
444.00	3937.0

Instrument :
BNA_N
ClientSampleId :
DFTPP33

m/z	Abundance
38.10	689.0
39.00	4778.0
40.00	614.0
41.00	375.0
44.00	1842.0
49.10	431.0
50.00	16440.0
51.00	66256.0
52.00	3149.0
55.00	581.0
55.90	1801.0
57.00	5385.0
61.00	983.0
62.00	979.0
63.00	2679.0
65.00	1429.0
68.10	1198.0
69.00	72200.0
69.90	307.0
73.00	915.0
74.00	6363.0
75.00	12362.0
76.00	4055.0
77.00	82864.0
78.00	5990.0
79.00	5156.0
80.00	4151.0
81.00	5673.0
81.90	1551.0
83.00	1457.0
84.90	1163.0
85.90	1554.0
87.00	699.0
91.00	1331.0
91.90	1389.0
93.00	8468.0
93.90	598.0
94.90	319.0
96.10	468.0
98.00	6698.0
99.00	4991.0
100.00	422.0
101.00	3453.0
102.90	960.0
104.00	2379.0
105.00	2151.0
106.10	671.0
107.00	24840.0
108.00	3571.0
109.10	758.0
110.00	47016.0
111.00	7171.0
111.90	1016.0
116.00	1476.0
116.90	17528.0
118.00	1319.0
119.90	387.0
122.00	2015.0
123.00	2625.0
124.00	1147.0
125.00	1237.0
127.00	91304.0
128.00	7627.0
128.90	34328.0
130.00	2518.0
130.80	628.0
131.70	406.0
133.10	311.0
134.00	901.0
135.00	2668.0
135.90	1229.0
137.00	1499.0
137.90	355.0
140.00	400.0
140.90	4581.0
141.90	1441.0
143.00	803.0
146.00	849.0
147.00	2266.0
147.90	4738.0
149.00	1028.0
150.00	315.0
151.10	811.0
152.90	1534.0
153.90	1262.0
155.00	2760.0
156.00	4639.0
157.10	824.0
157.90	1085.0
159.00	766.0
159.90	1465.0

Instrument :
BNA_N
ClientSampleId :
DFTPP33

161.00	2359.0
161.80	591.0
164.90	1679.0
166.00	1186.0
167.00	9317.0
168.00	4330.0
169.00	816.0
170.10	380.0
170.80	374.0
171.90	914.0
172.90	1033.0
173.90	1914.0
175.00	3861.0
176.00	964.0
176.90	2008.0
178.00	670.0
178.90	6532.0
180.00	5438.0
181.00	2692.0
182.00	418.0
183.80	572.0
185.00	3473.0
186.00	28040.0
187.00	8352.0
188.00	1122.0
189.00	1649.0
189.90	316.0
190.90	898.0
192.00	2478.0
193.00	2608.0
193.90	440.0
194.80	312.0
195.00	313.0
196.00	7020.0
196.90	1229.0
197.90	225792.0
198.90	15166.0
200.00	913.0
201.70	1260.0
202.90	1076.0
204.00	7506.0
205.00	12956.0
206.00	57552.0
207.00	7758.0
208.00	2005.0
208.90	867.0
210.90	2291.0
211.80	514.0
214.90	631.0
216.00	1289.0
216.90	14773.0
217.90	1873.0
221.00	9723.0
222.90	2963.0
224.00	31536.0
225.00	7930.0
226.00	898.0
227.00	11963.0
227.90	1420.0
229.00	2925.0
231.00	1305.0
234.00	823.0
235.00	849.0
235.90	652.0
236.90	957.0
239.00	498.0
241.00	964.0
242.00	1627.0
243.00	1754.0
244.00	25056.0
245.00	3281.0
246.00	4214.0
246.90	1026.0
249.00	993.0
252.90	473.0
255.00	139776.0
256.00	19936.0
257.00	1527.0
258.00	6974.0
258.90	1262.0
264.90	2835.0
265.90	408.0
271.00	351.0
273.00	3639.0
274.00	10840.0
275.00	63152.0
276.00	8435.0
277.00	5346.0
277.90	701.0
281.00	349.0
282.90	439.0
284.00	502.0
285.00	882.0
293.00	1212.0
296.00	16121.0

Instrument :

BNA_N

ClientSampleId :

DFTPP33

296.90	2189.0
301.90	326.0
303.00	1942.0
303.90	544.0
314.00	831.0
314.90	1620.0
316.00	1078.0
321.00	535.0
323.00	5767.0
324.00	1018.0
327.00	1086.0
327.90	518.0
331.90	438.0
333.00	621.0
334.00	3960.0
335.00	1122.0
340.90	707.0
345.90	1205.0
346.90	316.0
352.00	1823.0
353.00	1485.0
354.00	1683.0
354.90	494.0
365.00	8247.0
365.90	1287.0
371.00	533.0
372.00	3562.0
373.00	635.0
382.90	1132.0
390.00	382.0
401.90	1731.0
403.00	2571.0
403.90	907.0
421.00	2333.0
421.90	1715.0
423.00	12393.0
424.00	3410.0
425.00	308.0
441.00	38296.0
442.00	251520.0
443.00	46528.0
444.00	4258.0

Instrument :
BNA_N
ClientSampleId :
DFTPP33

Scan 2454 (17.422 min): BN002984.D
DFTPP33

m/z	Abundance
39.90	385.0
44.00	2005.0
55.10	314.0
72.90	458.0
207.00	822.0

Instrument :
BNA_N
ClientSampleId :
DFTPP33