

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100923\
 Data File : BN028187.D
 Acq On : 10 Oct 2023 08:02
 Operator : MA/JU
 Sample : 04622-01
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 I-1(0-5)

Quant Time: Oct 10 08:31:23 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN100223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 10 00:42:22 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.922	152	4885	0.400	ng	# 0.00
7) Naphthalene-d8	10.735	136	15477	0.400	ng	# 0.00
13) Acenaphthene-d10	14.587	164	25548	0.400	ng	# 0.02
19) Phenanthrene-d10	17.319	188	16314	0.400	ng	# 0.00
29) Chrysene-d12	21.515	240	23535	0.400	ng	# 0.00
35) Perylene-d12	23.978	264	26660	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.459	112	7067	0.485	ng	0.00
5) Phenol-d6	7.077	99	5867	0.319	ng	-0.01
8) Nitrobenzene-d5	9.112	82	5746	0.436	ng	-0.01
11) 2-Methylnaphthalene-d10	12.317	152	7620	0.332	ng	0.00
14) 2,4,6-Tribromophenol	16.065	330	1233	0.105	ng	-0.01
15) 2-Fluorobiphenyl	13.176	172	11381	0.124	ng	-0.01
27) Fluoranthene-d10	19.365	212	11682	0.285	ng	0.00
31) Terphenyl-d14	19.941	244	17106	0.312	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.386	88	677	0.124	ng	# 51
3) n-Nitrosodimethylamine	3.711	42	938	0.149	ng	# 1
6) bis(2-Chloroethyl)ether	7.445	93	470	0.032	ng	# 1
9) Naphthalene	10.799	128	14901144	372.626	ng	100
12) 2-Methylnaphthalene	12.392	142	2280903	81.536	ng	96
16) Acenaphthylene	14.288	152	1542776	13.632	ng	99
17) Acenaphthene	14.619	154	354702	4.917	ng	98
18) Fluorene	15.606	166	3093196	33.012	ng	99
20) 4,6-Dinitro-2-methylph...	16.772	198	1670	7.702	ng	# 16
23) Atrazine	16.810	200	582	0.078	ng	# 1
24) Pentachlorophenol	16.959	266	175	0.042	ng	# 61
25) Phenanthrene	17.368	178	12239858	278.292	ng	98
26) Anthracene	17.455	178	4096972	98.551	ng	# 95
28) Fluoranthene	19.407	202	9421711	185.617	ng	96
30) Pyrene	19.760	202	8108327	93.954	ng	97
32) Benzo(a)anthracene	21.497	228	4442288	55.470	ng	98
33) Chrysene	21.560	228	3672334	47.582	ng	96
34) Bis(2-ethylhexyl)phtha...	21.372	149	6859	0.130	ng	# 80
36) Indeno(1,2,3-cd)pyrene	26.565	276	1971797	21.988	ng	98
37) Benzo(b)fluoranthene	23.221	252	3996290	46.413	ng	# 89
39) Benzo(a)pyrene	24.031	252	744480	9.144	ng	91
40) Dibenzo(a,h)anthracene	26.571	278	580877	7.918	ng	94
41) Benzo(g,h,i)perylene	27.378	276	1850986	24.582	ng	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100923\
Data File : BN028187.D
Acq On : 10 Oct 2023 08:02
Operator : MA/JU
Sample : 04622-01
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
I-1(0-5)


Quant Time: Oct 10 08:31:23 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN100223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 10 00:42:22 2023
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.922 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BN028187.D
 Acq: 10 Oct 2023 08:02


Instrument :
 BNA_N
 ClientSampleId :
 I-1(0-5)





Tgt Ion:152 Resp: 4885
 Ion Ratio Lower Upper
 152 100
 150 152.3 125.6 188.4
 115 92.5 54.8 82.2#




#2
 1,4-Dioxane
 Concen: 0.124 ng
 RT: 3.386 min Scan# 70
 Delta R.T. 0.029 min
 Lab File: BN028187.D
 Acq: 10 Oct 2023 08:02



Tgt Ion: 88 Resp: 677
 Ion Ratio Lower Upper
 88 100
 43 0.0 24.6 36.8#
 58 30.1 53.8 80.6#









#3
n-Nitrosodimethylamine
Concen: 0.149 ng
RT: 3.711 min Scan# 116
Delta R.T. -0.000 min
Lab File: BN028187.D
Acq: 10 Oct 2023 08:02




Instrument :
BNA_N
ClientSampleId :
I-1(0-5)


Tgt Ion: 42 Resp: 938
Ion Ratio Lower Upper
42 100
74 456.3 115.9 173.9#
44 37.4 8.0 12.0#



#4
2-Fluorophenol
Concen: 0.485 ng
RT: 5.459 min Scan# 357
Delta R.T. -0.007 min
Lab File: BN028187.D
Acq: 10 Oct 2023 08:02

Tgt Ion: 112 Resp: 7067
Ion Ratio Lower Upper
112 100
64 0.0 45.0 67.6#
63 14.8 23.4 35.0#









#5
Phenol-d6
Concen: 0.319 ng
RT: 7.077 min Scan# 581
Delta R.T. -0.015 min
Lab File: BN028187.D
Acq: 10 Oct 2023 08:02




Instrument :
BNA_N
ClientSampleId :
I-1(0-5)


Tgt Ion: 99 Resp: 5867
Ion Ratio Lower Upper
99 100
42 46.8 14.6 22.0#
71 33.2 28.3 42.5



#6
bis(2-Chloroethyl)ether
Concen: 0.032 ng
RT: 7.445 min Scan# 632
Delta R.T. 0.079 min
Lab File: BN028187.D
Acq: 10 Oct 2023 08:02

Tgt Ion: 93 Resp: 470
Ion Ratio Lower Upper
93 100
63 0.0 58.8 88.2#
95 524.5 26.6 39.8#





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.735 min Scan# 91

Delta R.T. -0.000 min

Lab File: BN028187.D

Acq: 10 Oct 2023 08:02

Instrument :

BNA_N

ClientSampleId :

I-1(0-5)

Tgt Ion:136 Resp: 15477





Ion Ratio Lower Upper

136 100

137 13.8 9.4 14.0

54 9.9 7.0 10.6

68 10.4 6.4 9.6#



#8

Nitrobenzene-d5

Concen: 0.436 ng

RT: 9.112 min Scan# 827

Delta R.T. -0.011 min

Lab File: BN028187.D

Acq: 10 Oct 2023 08:02




Tgt Ion: 82 Resp: 5746


Ion Ratio Lower Upper

82 100

128 43.6 32.7 49.1

54 47.4 38.7 58.1





#9

Naphthalene

Concen: 372.626 ng

RT: 10.799 min Scan# 9

Delta R.T. 0.011 min

Lab File: BN028187.D

Acq: 10 Oct 2023 08:02

Instrument :

BNA_N

ClientSampleId :

I-1(0-5)





Tgt Ion:128 Resp:14901144

Ion Ratio Lower Upper

128 100

129 11.9 9.4 14.2

127 13.4 10.9 16.3



#11

2-Methylnaphthalene-d10

Concen: 0.332 ng

RT: 12.317 min Scan# 1266

Delta R.T. -0.008 min

Lab File: BN028187.D




Acq: 10 Oct 2023 08:02


Tgt Ion:152 Resp: 7620

Ion Ratio Lower Upper

152 100

151 22.3 16.5 24.7









#12
2-Methylnaphthalene
Concen: 81.536 ng
RT: 12.392 min Scan# 1286
Delta R.T. -0.008 min
Lab File: BN028187.D
Acq: 10 Oct 2023 08:02




Instrument :
BNA_N
ClientSampleId :
I-1(0-5)


Tgt Ion:142 Resp: 2280903
Ion Ratio Lower Upper
142 100
141 87.9 70.7 106.1
115 42.5 28.6 42.8



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.587 min Scan# 1596
Delta R.T. 0.021 min
Lab File: BN028187.D
Acq: 10 Oct 2023 08:02

Tgt Ion:164 Resp: 25548
Ion Ratio Lower Upper
164 100
162 13.9 80.8 121.2#
160 9.9 36.4 54.6#









#14
2,4,6-Tribromophenol
Concen: 0.105 ng
RT: 16.065 min Scan# 11
Delta R.T. -0.013 min
Lab File: BN028187.D
Acq: 10 Oct 2023 08:02




Instrument :
BNA_N
ClientSampleId :
I-1(0-5)


Tgt Ion:330 Resp: 1233
Ion Ratio Lower Upper
330 100
332 97.5 78.1 117.1
141 0.0 33.4 50.2#



#15
2-Fluorobiphenyl
Concen: 0.124 ng
RT: 13.176 min Scan# 1464
Delta R.T. -0.011 min
Lab File: BN028187.D
Acq: 10 Oct 2023 08:02

Tgt Ion:172 Resp: 11381
Ion Ratio Lower Upper
172 100
171 40.7 28.1 42.1
170 24.9 18.4 27.6





#16

Acenaphthylene

Concen: 13.632 ng

RT: 14.288 min Scan# 1568

Delta R.T. -0.011 min

Lab File: BN028187.D

Acq: 10 Oct 2023 08:02

Instrument :

BNA_N

ClientSampleId :

I-1(0-5)





Tgt Ion:152 Resp: 1542776

Ion Ratio Lower Upper

152 100

151 19.6 15.8 23.6

153 14.2 10.2 15.4



#17

Acenaphthene

Concen: 4.917 ng

RT: 14.619 min Scan# 1599

Delta R.T. -0.011 min

Lab File: BN028187.D

Acq: 10 Oct 2023 08:02




Tgt Ion:154 Resp: 354702


Ion Ratio Lower Upper

154 100

153 112.4 90.8 136.2

152 56.9 43.9 65.9





#18

Fluorene

Concen: 33.012 ng

RT: 15.606 min Scan# 1

Delta R.T. -0.013 min

Lab File: BN028187.D

Acq: 10 Oct 2023 08:02

Instrument :

BNA_N

ClientSampleId :

I-1(0-5)





Tgt Ion:166 Resp: 3093196

Ion Ratio Lower Upper

166 100

165 97.5 77.8 116.8

167 14.5 10.7 16.1



#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.319 min Scan# 1828

Delta R.T. -0.000 min

Lab File: BN028187.D

Acq: 10 Oct 2023 08:02




Tgt Ion:188 Resp: 16314


Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 38.2 10.2 15.4#









#20
4,6-Dinitro-2-methylphenol
Concen: 7.702 ng
RT: 16.772 min Scan# 1784
Delta R.T. -0.013 min
Lab File: BN028187.D
Acq: 10 Oct 2023 08:02




Instrument :
BNA_N
ClientSampleId :
I-1(0-5)


Tgt Ion:198 Resp: 1670
Ion Ratio Lower Upper
198 100
51 215.4 180.7 271.1
105 553.4 222.2 333.2#



#23
Atrazine
Concen: 0.078 ng
RT: 16.810 min Scan# 1787
Delta R.T. 0.025 min
Lab File: BN028187.D
Acq: 10 Oct 2023 08:02

Tgt Ion:200 Resp: 582
Ion Ratio Lower Upper
200 100
173 265.3 24.7 37.1#
215 109.4 39.0 58.6#





#24

Pentachlorophenol

Concen: 0.042 ng

RT: 16.959 min Scan# 1

Delta R.T. -0.013 min

Lab File: BN028187.D

Acq: 10 Oct 2023 08:02

Instrument :

BNA_N

ClientSampleId :

I-1(0-5)





Tgt Ion:266 Resp: 175

Ion Ratio Lower Upper

266 100

264 49.1 52.7 79.1#

268 115.4 56.2 84.2#



#25

Phenanthrene

Concen: 278.292 ng

RT: 17.368 min Scan# 1832

Delta R.T. -0.000 min

Lab File: BN028187.D

Acq: 10 Oct 2023 08:02




Tgt Ion:178 Resp:12239858


Ion Ratio Lower Upper

178 100

176 19.4 15.3 22.9

179 16.7 12.4 18.6





#26

Anthracene

Concen: 98.551 ng

RT: 17.455 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN028187.D

Acq: 10 Oct 2023 08:02

Instrument :

BNA_N

ClientSampleId :

1-1(0-5)





Tgt Ion:178 Resp: 4096972

Ion Ratio Lower Upper

178 100

176 18.2 14.7 22.1

179 19.6 12.2 18.2#



#27

Fluoranthene-d10

Concen: 0.285 ng

RT: 19.365 min Scan# 2108

Delta R.T. 0.009 min

Lab File: BN028187.D

Acq: 10 Oct 2023 08:02




Tgt Ion:212 Resp: 11682


Ion Ratio Lower Upper

212 100

106 25.5 11.3 16.9#

104 21.3 6.4 9.6#









#28
Fluoranthene
Concen: 185.617 ng
RT: 19.407 min Scan# 2117
Delta R.T. 0.023 min
Lab File: BN028187.D
Acq: 10 Oct 2023 08:02




Instrument :
BNA_N
ClientSampleId :
I-1(0-5)


Tgt Ion:202 Resp: 9421711
Ion Ratio Lower Upper
202 100
101 14.3 9.8 14.6
203 18.5 13.8 20.8



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.515 min Scan# 2518
Delta R.T. -0.000 min
Lab File: BN028187.D
Acq: 10 Oct 2023 08:02

Tgt Ion:240 Resp: 23535
Ion Ratio Lower Upper
240 100
120 38.1 13.8 20.6#
236 39.6 22.6 33.8#









#30
Pyrene
Concen: 93.954 ng
RT: 19.760 min Scan# 2196
Delta R.T. 0.009 min
Lab File: BN028187.D
Acq: 10 Oct 2023 08:02




Instrument :
BNA_N
ClientSampleId :
I-1(0-5)


Tgt Ion:202 Resp: 8108327
Ion Ratio Lower Upper
202 100
200 21.5 16.5 24.7
203 20.1 14.7 22.1



#31
Terphenyl-d14
Concen: 0.312 ng
RT: 19.941 min Scan# 2232
Delta R.T. -0.000 min
Lab File: BN028187.D
Acq: 10 Oct 2023 08:02

Tgt Ion:244 Resp: 17106
Ion Ratio Lower Upper
244 100
212 20.3 6.6 9.8#
122 32.7 11.4 17.0#





#32

Benzo(a)anthracene

Concen: 55.470 ng

RT: 21.497 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN028187.D

Acq: 10 Oct 2023 08:02

Instrument :

BNA_N

ClientSampleId :

I-1(0-5)





Tgt Ion:228 Resp: 4442288

Ion Ratio Lower Upper

228 100

226 28.4 22.0 33.0

229 21.1 16.0 24.0



#33

Chrysene

Concen: 47.582 ng

RT: 21.560 min Scan# 2523

Delta R.T. 0.009 min

Lab File: BN028187.D

Acq: 10 Oct 2023 08:02




Tgt Ion:228 Resp: 3672334


Ion Ratio Lower Upper

228 100

226 30.8 23.8 35.8

229 23.9 16.6 25.0









#34
Bis(2-ethylhexyl)phthalate
Concen: 0.130 ng
RT: 21.372 min Scan# 21
Delta R.T. 0.009 min
Lab File: BN028187.D
Acq: 10 Oct 2023 08:02




Instrument :
BNA_N
ClientSampleId :
I-1(0-5)


Tgt Ion:149 Resp: 6859
Ion Ratio Lower Upper
149 100
167 39.4 22.1 33.1#
279 4.7 3.4 5.0



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.978 min Scan# 3154
Delta R.T. 0.003 min
Lab File: BN028187.D
Acq: 10 Oct 2023 08:02

Tgt Ion:264 Resp: 26660
Ion Ratio Lower Upper
264 100
260 30.0 20.2 30.4
265 160.3 54.8 82.2#





#36

Indeno(1,2,3-cd)pyrene

Concen: 21.988 ng

RT: 26.565 min Scan# 4055

Delta R.T. 0.009 min

Lab File: BN028187.D

Acq: 10 Oct 2023 08:02

Instrument :

BNA_N

ClientSampleId :

I-1(0-5)





Tgt Ion:276 Resp: 1971797

Ion Ratio Lower Upper

276 100

138 23.8 20.9 31.3

277 24.7 19.8 29.6



#37

Benzo(b)fluoranthene

Concen: 46.413 ng

RT: 23.221 min Scan# 2895

Delta R.T. 0.009 min

Lab File: BN028187.D

Acq: 10 Oct 2023 08:02




Tgt Ion:252 Resp: 3996290


Ion Ratio Lower Upper

252 100

253 22.7 22.9 34.3#

125 12.4 13.8 20.8#





#39

Benzo(a)pyrene

Concen: 9.144 ng

RT: 24.031 min Scan# 31

Delta R.T. 0.167 min

Lab File: BN028187.D

Acq: 10 Oct 2023 08:02

Instrument :

BN028187.D

ClientSampleId :

1-1(0-5)





Tgt Ion:252 Resp: 744480

Ion Ratio Lower Upper

252 100

253 25.0 25.0 37.4

125 17.6 16.6 24.8



#40

Dibenzo(a,h)anthracene

Concen: 7.918 ng

RT: 26.571 min Scan# 4041

Delta R.T. -0.006 min

Lab File: BN028187.D

Acq: 10 Oct 2023 08:02




Tgt Ion:278 Resp: 580877


Ion Ratio Lower Upper

278 100

139 20.3 17.7 26.5

279 26.8 25.2 37.8





#41

Benzo(g,h,i)perylene

Concen: 24.582 ng

RT: 27.378 min Scan# 41

Delta R.T. 0.012 min

Lab File: BN028187.D


Acq: 10 Oct 2023 08:02

Instrument :

BNA_N

ClientSampleId :

I-1(0-5)



Tgt Ion:276 Resp: 1850986

Ion Ratio Lower Upper

276 100

277 24.2 22.1 33.1

138 24.3 20.8 31.2

