

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100923\
 Data File : BN028173.D
 Acq On : 09 Oct 2023 23:38
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 10 00:45:16 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN100223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 10 00:42:22 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	103	0.00
2	1,4-Dioxane	0.400	0.436	-9.0	110	0.00
3	n-Nitrosodimethylamine	0.400	0.398	0.5	102	0.00
4 S	2-Fluorophenol	0.400	0.342	14.5	94	0.00
5 S	Phenol-d6	0.400	0.319	20.3	90	0.00
6	bis(2-Chloroethyl)ether	0.400	0.352	12.0	91	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	89	0.00
8 S	Nitrobenzene-d5	0.400	0.361	9.8	84	0.00
9	Naphthalene	0.400	0.403	-0.8	90	0.00
10	Hexachlorobutadiene	0.400	0.398	0.5	89	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.366	8.5	82	0.00
12	2-Methylnaphthalene	0.400	0.367	8.3	81	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	74	0.00
14 S	2,4,6-Tribromophenol	0.400	0.250	37.5#	49	0.00
15 S	2-Fluorobiphenyl	0.400	0.413	-3.2	77	0.00
16	Acenaphthylene	0.400	0.363	9.3	67	0.00
17	Acenaphthene	0.400	0.403	-0.8	73	0.00
18	Fluorene	0.400	0.388	3.0	70	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	65	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.411	-2.7	66	0.00
21	4-Bromophenyl-phenylether	0.400	0.395	1.3	64	0.00
22	Hexachlorobenzene	0.400	0.418	-4.5	66	0.00
23	Atrazine	0.400	0.327	18.3	52	0.00
24	Pentachlorophenol	0.400	0.295	26.3#	56	0.00
25	Phenanthrene	0.400	0.406	-1.5	65	0.00
26	Anthracene	0.400	0.363	9.3	58	0.00
27 SURR	Fluoranthene-d10	0.400	0.383	4.3	59	0.00
28	Fluoranthene	0.400	0.388	3.0	60	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	59	0.00
30	Pyrene	0.400	0.389	2.8	59	0.00
31 S	Terphenyl-d14	0.400	0.366	8.5	58	0.00
32	Benzo(a)anthracene	0.400	0.368	8.0	53	0.00
33	Chrysene	0.400	0.425	-6.2	62	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.285	28.8#	41	0.00
35 I	Perylene-d12	0.400	0.400	0.0	66	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.455	-13.7	81	0.00
37	Benzo(b)fluoranthene	0.400	0.398	0.5	65	0.00
38	Benzo(k)fluoranthene	0.400	0.403	-0.8	66	0.00
39 C	Benzo(a)pyrene	0.400	0.389	2.8	64	0.00
40	Dibenzo(a,h)anthracene	0.400	0.455	-13.7	81	0.00
41	Benzo(g,h,i)perylene	0.400	0.482	-20.5	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4