

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN101018\
 Data File : BN003050.D
 Acq On : 10 Oct 2018 04:55
 Operator : MJ/SJ
 Sample : J5115-15
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 JLC67

Manual Integrations
 APPROVED

Sohil
 10/10/2018 4:58:20 PM

Quant Time: Oct 10 09:07:16 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN100918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 10 08:36:11 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	1924	0.40	ng/ul	-0.02
2) Naphthalene-d8	10.39	136	7829	0.40	ng/ul	-0.01
6) Acenaphthene-d10	14.26	164	4327	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.01	188	8864m	0.40	ng/ul	-0.02
16) Chrysene-d12	21.22	240	5141m	0.40	ng/ul	-0.01
20) Perylene-d12	23.43	264	5139m	0.40	ng/ul	-0.01
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.00	152	2983	0.25	ng/ul	-0.01
14) Fluoranthene-d10	19.05	212	5644m	0.22	ng/ul	-0.01
Target Compounds					Qvalue	
3) Naphthalene	10.44	128	2775	0.138	ng/ul	97
5) 2-Methylnaphthalene	12.08	142	711	0.051	ng/ul	100
7) Acenaphthylene	13.98	152	2674	0.135	ng/ul#	90
8) Acenaphthene	14.32	153	1928	0.128	ng/ul	97
9) Fluorene	15.32	166	3170	0.184	ng/ul#	93
12) Phenanthrene	17.06	178	48756	1.973	ng/ul	99
13) Anthracene	17.15	178	15214m	0.652	ng/ul	
15) Fluoranthene	19.08	202	201304m	6.944	ng/ul	
17) Pyrene	19.45	202	203601m	10.197	ng/ul	
18) Benzo(a)anthracene	21.21	228	87287m	5.276	ng/ul	
19) Chrysene	21.26	228	90825m	5.407	ng/ul	
21) Benzo(b)fluoranthene	22.77	252	107872m	5.829	ng/ul	
22) Benzo(k)fluoranthene	22.81	252	38134m	1.988	ng/ul	
23) Benzo(a)pyrene	23.33	252	78605m	4.739	ng/ul	
24) Indeno(1,2,3-cd)pyrene	25.64	276	45479m	2.386	ng/ul	
25) Dibenzo(a,h)anthracene	25.64	278	12300m	0.806	ng/ul	
26) Benzo(g,h,i)perylene	26.31	276	42894m	2.669	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN101018\
Data File : BN003050.D
Acq On : 10 Oct 2018 04:55
Operator : MJ/SJ
Sample : J5115-15
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
JLC67

Manual Integrations
APPROVED

Sohil
10/10/2018 4:58:20 PM

Quant Time: Oct 10 09:07:16 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN100918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 10 08:36:11 2018
Response via : Initial Calibration

