

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN101018\
 Data File : BN003052.D
 Acq On : 10 Oct 2018 06:05
 Operator : MJ/SJ
 Sample : J5115-20
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 JLC63

Manual Integrations
 APPROVED

Sohil
 10/10/2018 4:58:22 PM

Quant Time: Oct 10 09:17:18 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN100918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 10 08:36:11 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	1984	0.40	ng/ul	-0.02
2) Naphthalene-d8	10.39	136	8020	0.40	ng/ul	-0.01
6) Acenaphthene-d10	14.26	164	4563	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.01	188	9735m	0.40	ng/ul	-0.02
16) Chrysene-d12	21.22	240	6078m	0.40	ng/ul	-0.01
20) Perylene-d12	23.43	264	6699m	0.40	ng/ul	-0.01
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.00	152	3382	0.27	ng/ul	-0.02
14) Fluoranthene-d10	19.05	212	7573m	0.27	ng/ul	-0.01
Target Compounds				Qvalue		
3) Naphthalene	10.44	128	38127	1.853	ng/ul	96
5) 2-Methylnaphthalene	12.07	142	9791	0.691	ng/ul	97
7) Acenaphthylene	13.98	152	5786	0.277	ng/ul#	96
8) Acenaphthene	14.32	153	74679	4.714	ng/ul	98
9) Fluorene	15.32	166	63974	3.526	ng/ul	98
11) Pentachlorophenol	16.71	266	121m	0.128	ng/ul	
12) Phenanthrene	17.06	178	1050321m	38.704	ng/ul	
13) Anthracene	17.15	178	207806m	8.107	ng/ul	
15) Fluoranthene	19.09	202	1768415m	55.547	ng/ul	
17) Pyrene	19.45	202	1478936m	62.649	ng/ul	
18) Benzo(a)anthracene	21.21	228	654341m	33.455	ng/ul	
19) Chrysene	21.26	228	681160m	34.297	ng/ul	
21) Benzo(b)fluoranthene	22.78	252	813034m	33.700	ng/ul	
22) Benzo(k)fluoranthene	22.82	252	285303m	11.407	ng/ul	
23) Benzo(a)pyrene	23.34	252	576277m	26.653	ng/ul	
24) Indeno(1,2,3-cd)pyrene	25.64	276	419243m	16.873	ng/ul	
25) Dibenzo(a,h)anthracene	25.64	278	100446m	5.051	ng/ul	
26) Benzo(g,h,i)perylene	26.32	276	413655m	19.744	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN101018\
Data File : BN003052.D
Acq On : 10 Oct 2018 06:05
Operator : MJ/SJ
Sample : J5115-20
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
JLC63

Manual Integrations
APPROVED

Sohil
10/10/2018 4:58:22 PM

Quant Time: Oct 10 09:17:18 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN100918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 10 08:36:11 2018
Response via : Initial Calibration

