


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN101018\
Data File : BN003028.D
Acq On : 09 Oct 2018 16:00
Operator : MJ/SJ
Sample : DFTPP34
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP34

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN100918.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Oct 10 08:36:11 2018



AutoFind: Scans 2460, 2461, 2462; Background Corrected with Scan 2453

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.8	80437	PASS
68	69	0.00	2	1.8	1589	PASS
69	198	0.00	100	34.7	87939	PASS
70	69	0.00	2	0.7	606	PASS
127	198	10	80	42.9	108736	PASS
197	198	0.00	2	0.2	477	PASS
198	198	100	100	100.0	253269	PASS
199	198	5	9	6.8	17291	PASS
275	198	10	60	27.5	69752	PASS
365	198	1	100	3.3	8356	PASS
441	198	0.01	100	14.5	36800	PASS
442	198	50	100	94.6	239466	PASS
443	442	15	24	19.0	45610	PASS

m/z	Abundance
38.10	985.0
39.00	6518.0
40.00	792.0
41.00	629.0
43.10	458.0
43.90	1888.0
49.00	520.0
50.00	21320.0
51.00	83352.0
52.10	4491.0
53.00	367.0
55.00	997.0
56.00	2647.0
57.00	6031.0
61.00	1193.0
62.00	1264.0
63.00	3243.0
63.90	577.0
65.00	1902.0
67.10	365.0
68.10	1551.0
69.00	90216.0
69.90	521.0
73.00	1030.0
74.00	9407.0
75.00	13823.0
76.00	4484.0
77.00	102360.0
78.00	7323.0
79.00	6780.0
80.00	5197.0
81.00	7865.0
81.90	1752.0
82.90	2138.0
85.00	1577.0
86.00	2180.0
86.90	952.0
87.80	336.0
91.00	1533.0
92.00	1447.0
93.00	11155.0
94.00	834.0
95.10	609.0
96.00	690.0
97.20	444.0
98.00	7902.0
99.00	6012.0
100.00	610.0
101.00	3833.0
103.00	1103.0
103.90	2450.0
105.00	2243.0
106.00	831.0
107.00	28640.0
108.00	4698.0
109.00	1269.0
110.00	56336.0
111.00	7824.0
112.00	820.0
116.00	1581.0
117.00	20208.0
117.90	1452.0
118.90	347.0
120.00	493.0
122.00	1896.0
123.00	2877.0
124.00	1405.0
125.00	1375.0
127.00	103840.0
128.00	8052.0
129.00	38992.0
130.00	2956.0
131.00	767.0
132.10	466.0
133.00	356.0
133.90	1171.0
134.90	3045.0
136.00	1103.0
137.10	1550.0
139.90	613.0
141.00	5150.0
141.90	1735.0
142.80	1135.0
145.90	996.0
146.90	2756.0
148.00	5200.0
148.90	1365.0
149.90	379.0
151.00	971.0
152.00	676.0
153.00	1377.0

Instrument :
BNA_N
ClientSampleId :
DFTPP34

154.00	1356.0
155.00	2380.0
156.00	4313.0
157.10	932.0
158.00	1278.0
158.90	781.0
160.00	1627.0
160.90	2600.0
162.00	820.0
164.90	2121.0
165.90	1619.0
167.00	9446.0
168.00	4107.0
169.00	784.0
170.00	409.0
171.10	599.0
171.90	1177.0
173.00	1162.0
174.00	2628.0
175.00	3803.0
176.10	1280.0
176.90	2299.0
178.10	475.0
178.90	7065.0
180.00	5270.0
180.90	2625.0
182.10	509.0
183.90	447.0
185.00	3450.0
186.00	28736.0
187.00	8628.0
188.00	1198.0
189.00	1409.0
191.10	696.0
192.00	2414.0
193.00	2334.0
194.00	505.0
196.00	7123.0
197.00	1432.0
197.90	226112.0
198.90	16760.0
200.00	1210.0
201.40	1456.0
202.90	1370.0
204.00	7512.0
205.00	12557.0
206.00	55608.0
207.00	8654.0
207.90	2000.0
208.90	547.0
211.10	2815.0
214.80	314.0
215.20	344.0
216.00	1078.0
216.90	14034.0
217.90	1714.0
221.00	10102.0
221.60	2747.0
223.10	2991.0
224.00	30936.0
225.00	8067.0
225.90	912.0
227.00	12091.0
228.00	1759.0
229.00	2572.0
230.00	418.0
231.00	1117.0
233.90	856.0
235.00	920.0
236.00	538.0
237.10	1029.0
238.70	426.0
239.00	318.0
240.00	426.0
240.90	650.0
241.90	1548.0
243.00	1623.0
244.00	24568.0
245.00	3261.0
245.90	4284.0
246.90	867.0
249.00	987.0
252.90	655.0
255.00	124296.0
256.00	18696.0
257.00	1419.0
258.00	6174.0
259.00	978.0
264.90	2176.0
265.90	307.0
273.00	3727.0
274.00	9612.0
275.00	59296.0
276.00	8067.0
277.00	4259.0

Instrument :
BNA_N
ClientSampleId :
DFTPP34

278.00	691.0
283.00	501.0
284.10	379.0
285.00	605.0
292.90	988.0
293.90	316.0
295.00	466.0
296.00	13892.0
297.00	2262.0
302.00	363.0
302.90	1519.0
304.00	541.0
313.90	767.0
314.90	1514.0
315.90	974.0
320.90	510.0
323.00	4660.0
324.00	773.0
326.90	952.0
327.90	526.0
332.00	518.0
332.90	516.0
334.00	3034.0
334.90	894.0
341.00	644.0
341.90	367.0
345.90	1000.0
346.90	320.0
352.00	1351.0
353.00	810.0
353.90	1658.0
355.00	353.0
365.00	6644.0
366.00	914.0
371.00	368.0
372.00	2753.0
372.90	532.0
382.90	632.0
389.90	354.0
400.90	341.0
402.00	1074.0
403.00	1909.0
403.90	625.0
421.10	1588.0
422.00	1265.0
423.00	9583.0
424.10	2329.0
441.00	26304.0
442.00	168256.0
443.00	33592.0
444.00	3111.0

Instrument :
BNA_N
ClientSampleId :
DFTPP34

m/z	Abundance
36.90	304.0
38.10	1089.0
39.00	6980.0
39.90	888.0
41.10	641.0
44.00	1919.0
49.00	766.0
50.00	22920.0
51.00	95688.0
52.00	5035.0
55.00	851.0
56.00	3158.0
57.00	6869.0
57.90	409.0
61.00	1008.0
62.10	1269.0
63.00	3705.0
64.00	835.0
64.90	2056.0
67.00	398.0
68.10	1923.0
69.00	104680.0
70.00	686.0
73.20	907.0
74.00	9934.0
75.00	15568.0
76.00	5355.0
77.00	112208.0
78.00	8005.0
79.00	7431.0
79.90	5895.0
81.00	9057.0
81.90	2304.0
83.00	1956.0
83.90	342.0
85.00	1887.0
85.90	2156.0
87.00	958.0
87.90	369.0
90.90	2145.0
92.00	1913.0
93.00	11872.0
94.00	927.0
95.00	764.0
96.00	908.0
96.90	457.0
98.00	9080.0
99.00	7233.0
100.00	580.0
101.00	4738.0
101.90	351.0
103.00	1498.0
104.00	3102.0
105.00	2760.0
105.90	902.0
107.00	33928.0
108.00	5566.0
109.00	1410.0
110.00	65720.0
111.00	9494.0
111.90	1336.0
112.90	417.0
115.90	1955.0
117.00	23376.0
117.90	2005.0
120.00	540.0
121.90	2312.0
123.00	3508.0
124.00	1559.0
125.00	1910.0
127.00	126256.0
128.00	9451.0
129.00	47136.0
130.00	3615.0
131.00	904.0
131.80	465.0
133.10	329.0
134.00	1312.0
135.00	4005.0
136.00	1573.0
136.90	2441.0
137.90	453.0
138.90	470.0
140.00	707.0
141.00	6303.0
142.00	2228.0
142.90	1396.0
144.00	465.0
144.80	393.0
146.00	1108.0
147.00	3395.0

Instrument :
BNA_N
ClientSampleId :
DFTPP34

148.00	6752.0
149.00	1351.0
150.00	437.0
151.00	784.0
151.90	566.0
153.00	1757.0
154.00	1524.0
155.00	3832.0
156.00	5216.0
157.00	1335.0
157.90	1066.0
158.90	981.0
160.00	1950.0
160.90	2744.0
161.90	794.0
163.10	414.0
164.10	510.0
164.90	2483.0
166.00	1801.0
167.00	11347.0
167.90	5525.0
168.90	836.0
169.80	451.0
170.90	506.0
171.90	1222.0
172.90	1515.0
174.00	2815.0
175.00	4874.0
176.00	1525.0
176.90	2337.0
177.90	859.0
178.90	9820.0
180.00	7098.0
181.00	3303.0
182.00	544.0
184.00	862.0
184.90	5206.0
186.00	37064.0
187.00	11251.0
188.00	1509.0
189.00	2322.0
189.90	441.0
190.90	1306.0
191.90	3292.0
193.00	3427.0
193.90	1086.0
195.00	612.0
196.00	8681.0
197.90	304640.0
198.90	20152.0
199.90	1410.0
201.50	1838.0
202.90	2095.0
204.00	10579.0
205.00	16696.0
206.00	75600.0
207.00	10598.0
207.90	2425.0
208.90	957.0
210.20	933.0
210.90	2940.0
211.80	709.0
214.90	678.0
215.90	1458.0
216.90	19648.0
217.90	2564.0
221.00	14519.0
223.00	4131.0
224.00	41640.0
225.00	10657.0
226.00	1242.0
226.90	15340.0
228.00	2066.0
228.90	3507.0
230.00	389.0
231.10	1501.0
233.90	968.0
234.90	1246.0
235.90	850.0
237.00	1749.0
238.90	726.0
239.90	536.0
241.00	1095.0
242.00	2468.0
243.00	2118.0
244.00	33408.0
245.00	4058.0
245.90	6080.0
246.90	1160.0
247.90	323.0
249.00	1284.0
250.80	309.0
252.90	543.0
254.10	1410.0
255.00	168320.0

Instrument :
BNA_N
ClientSampleId :
DFTPP34

256.00	24640.0
257.00	1904.0
257.90	9307.0
258.90	1538.0
260.90	322.0
264.90	3250.0
270.90	333.0
271.80	414.0
273.00	4851.0
274.00	13534.0
275.00	82568.0
276.00	10677.0
276.90	5844.0
277.90	1107.0
283.00	798.0
284.00	535.0
285.00	1080.0
292.00	345.0
293.00	1556.0
295.00	486.0
296.00	19272.0
297.00	2751.0
302.00	425.0
303.00	2322.0
303.90	644.0
314.10	892.0
315.00	2615.0
316.00	1051.0
320.90	778.0
322.10	373.0
323.00	6605.0
324.00	1394.0
326.90	1108.0
328.00	875.0
331.90	416.0
332.90	657.0
334.00	4849.0
335.00	1022.0
340.90	887.0
345.90	1333.0
351.90	2250.0
353.00	1746.0
353.90	2270.0
354.90	527.0
364.90	9135.0
366.00	1468.0
371.10	732.0
372.00	4073.0
373.00	1128.0
383.00	1001.0
384.00	401.0
390.00	481.0
390.90	395.0
391.90	330.0
400.90	378.0
402.00	1587.0
402.90	2413.0
404.00	809.0
421.00	2204.0
422.00	2261.0
423.00	14772.0
424.00	3243.0
425.00	381.0
441.10	42840.0
442.00	273472.0
443.00	54928.0
444.00	4402.0

Instrument :
BNA_N
ClientSampleId :
DFTPP34

m/z	Abundance
38.10	820.0
39.00	3849.0
40.00	502.0
41.10	464.0
43.00	443.0
43.90	2106.0
48.90	566.0
50.00	14225.0
51.00	62272.0
52.00	3328.0
55.00	671.0
56.00	2204.0
57.00	4649.0
60.90	985.0
62.00	964.0
63.00	2599.0
64.00	407.0
65.00	1416.0
67.10	333.0
68.00	1294.0
69.00	70168.0
70.00	611.0
72.90	957.0
74.00	6482.0
75.00	10305.0
76.00	3636.0
77.00	80264.0
78.00	5652.0
79.00	5337.0
79.90	3942.0
81.00	6156.0
81.90	1304.0
83.00	1310.0
83.90	443.0
84.90	1207.0
86.00	1570.0
86.90	780.0
87.90	374.0
91.00	1333.0
92.00	1483.0
93.00	8764.0
94.00	642.0
95.00	584.0
96.10	616.0
97.00	644.0
98.00	6283.0
98.90	4853.0
99.90	408.0
101.00	3002.0
103.00	1233.0
103.90	2252.0
104.90	2213.0
106.00	570.0
107.00	22976.0
108.00	3618.0
109.00	1295.0
110.00	47424.0
111.00	7503.0
112.10	590.0
116.00	1409.0
117.00	16752.0
117.90	1419.0
119.90	331.0
121.90	1453.0
123.00	2266.0
124.00	1067.0
124.90	1097.0
127.00	96112.0
128.00	7423.0
129.00	36168.0
130.00	3122.0
131.10	497.0
132.10	376.0
133.90	1076.0
135.00	2802.0
135.90	1211.0
136.90	1559.0
138.00	468.0
139.90	391.0
141.00	4463.0
141.90	1502.0
142.90	941.0
144.80	488.0
145.90	777.0
147.00	2515.0
148.00	4617.0
148.90	1128.0
149.90	363.0
151.00	645.0
151.50	447.0
153.00	1572.0

Instrument :
BNA_N
ClientSampleId :
DFTPP34

154.00	1056.0
155.00	2773.0
156.00	3769.0
157.20	813.0
158.00	989.0
158.90	721.0
159.90	1450.0
161.00	2051.0
161.90	804.0
165.00	1793.0
166.00	1582.0
167.00	9075.0
168.00	3612.0
168.90	786.0
171.00	585.0
171.90	1125.0
173.00	967.0
174.00	1964.0
175.00	3909.0
176.00	1125.0
176.90	1705.0
177.90	564.0
179.00	6190.0
179.90	5333.0
180.90	2412.0
181.90	438.0
184.00	589.0
184.90	3490.0
186.00	28904.0
187.00	7650.0
188.00	953.0
189.00	1979.0
189.90	311.0
191.10	945.0
192.00	2478.0
193.00	2336.0
194.00	644.0
194.90	392.0
196.00	6147.0
197.90	229056.0
198.90	14963.0
199.90	1212.0
201.40	1331.0
202.90	1283.0
204.00	7460.0
205.00	12876.0
206.00	59256.0
207.00	7535.0
207.90	2007.0
208.90	737.0
210.20	1099.0
210.90	2328.0
214.90	467.0
215.90	1374.0
217.00	15006.0
218.00	1785.0
221.00	11129.0
222.90	3455.0
224.00	32184.0
225.00	8123.0
226.00	823.0
227.00	11937.0
228.00	1882.0
229.00	2756.0
231.00	1194.0
233.90	864.0
234.90	996.0
235.90	570.0
236.90	803.0
239.00	559.0
240.00	507.0
240.90	873.0
242.00	1888.0
243.00	1624.0
244.00	26912.0
245.00	2978.0
245.90	4816.0
246.90	936.0
248.90	1098.0
251.90	325.0
253.00	483.0
255.00	141312.0
256.00	20528.0
257.00	1396.0
258.00	6940.0
258.80	1044.0
264.90	2964.0
266.00	480.0
272.10	366.0
273.00	4386.0
274.00	10973.0
275.00	67392.0
276.00	8932.0
277.00	5233.0
278.00	835.0

Instrument :
BNA_N
ClientSampleId :
DFTPP34

283.00	789.0
283.90	434.0
284.90	1023.0
292.00	393.0
293.00	916.0
296.00	17352.0
297.00	2091.0
301.90	426.0
303.00	2185.0
304.00	924.0
313.90	732.0
315.00	1509.0
315.90	1254.0
321.10	615.0
321.90	337.0
323.00	5924.0
324.00	1002.0
327.00	1256.0
327.90	458.0
331.90	506.0
332.90	438.0
334.00	3824.0
335.00	1196.0
341.00	542.0
345.90	1338.0
347.00	363.0
351.90	2123.0
353.00	1535.0
354.00	1934.0
355.00	335.0
364.90	9289.0
366.00	1007.0
370.80	532.0
371.10	512.0
372.00	3617.0
372.90	794.0
383.00	1133.0
383.80	420.0
389.90	533.0
390.80	366.0
400.90	430.0
402.00	1455.0
403.00	2359.0
404.00	700.0
421.00	2154.0
422.00	1760.0
423.00	14764.0
424.00	3488.0
425.10	380.0
441.00	41256.0
442.00	276672.0
443.00	48312.0
444.00	4722.0

Instrument :

BNA_N

ClientSampleId :

DFTPP34

m/z	Abundance
39.80	343.0
43.00	344.0
44.00	1929.0
55.00	471.0
68.90	415.0
73.10	474.0
81.00	370.0
109.10	302.0
206.90	895.0

Instrument :
BNA_N
ClientSampleId :
DFTPP34