




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101023\
Data File : BN028211.D
Acq On : 11 Oct 2023 00:05
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 11 06:05:12 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 11 06:03:31 2023
Response via : Initial Calibration






TIC: BN028211.D\data.ms

(70) Pentachlorophenol (C)**16.957min (+ 0.006) 25683.00 ng****response 459341**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	58.96
264.00	61.60	60.13
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101023\
Data File : BN028211.D
Acq On : 11 Oct 2023 00:05
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 11 06:05:12 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 11 06:03:31 2023
Response via : Initial Calibration



TIC: BN028211.D\data.ms

(77) Benzidine**19.598min (+ 0.006) 564044.77 ng****response 2834063**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.10
183.00	13.20	12.18
0.00	0.00	0.00