

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN101118\
 Data File : BN003079.D
 Acq On : 11 Oct 2018 00:02
 Operator : MJ/SJ
 Sample : J5157-14
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AD6

Manual Integrations
 APPROVED

Sohil
 10/11/2018 5:45:50 PM

Quant Time: Oct 11 08:22:30 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN100918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 11 08:10:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	1569	0.40	ng/ul	-0.02
2) Naphthalene-d8	10.39	136	6726	0.40	ng/ul	-0.02
6) Acenaphthene-d10	14.26	164	4053	0.40	ng/ul	-0.01
10) Phenanthrene-d10	17.01	188	8766m	0.40	ng/ul	-0.02
16) Chrysene-d12	21.23	240	6927m	0.40	ng/ul	0.00
20) Perylene-d12	23.44	264	4684m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.01	152	4354	0.42	ng/ul	-0.01
14) Fluoranthene-d10	19.06	212	10424m	0.42	ng/ul	0.00
Target Compounds					Ovalue	
13) Anthracene	17.15	178	639m	0.028	ng/ul	
15) Fluoranthene	19.09	202	1618m	0.056	ng/ul	
17) Pyrene	19.45	202	2004	0.074	ng/ul#	89
18) Benzo(a)anthracene	21.22	228	931	0.042	ng/ul#	79
19) Chrysene	21.27	228	2024	0.089	ng/ul	94
21) Benzo(b)fluoranthene	22.78	252	2859m	0.169	ng/ul	
22) Benzo(k)fluoranthene	22.82	252	1039m	0.059	ng/ul	
23) Benzo(a)pyrene	23.34	252	1563	0.103	ng/ul#	75
24) Indeno(1,2,3-cd)pyrene	25.69	276	980	0.056	ng/ul	93
26) Benzo(g,h,i)perylene	26.36	276	1001	0.068	ng/ul#	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN101118\
Data File : BN003079.D
Acq On : 11 Oct 2018 00:02
Operator : MJ/SJ
Sample : J5157-14
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AD6

Manual Integrations
APPROVED

Sohil
10/11/2018 5:45:50 PM

Quant Time: Oct 11 08:22:30 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN100918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 11 08:10:30 2018
Response via : Initial Calibration

