


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101118\
 Data File : BN003063.D
 Acq On : 10 Oct 2018 14:04
 Operator : MJ/SJ
 Sample : DFTPP35
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN100918.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Oct 11 08:10:30 2018



AutoFind: Scans 2460, 2461, 2462; Background Corrected with Scan 2453

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.8	70194	PASS
68	69	0.00	2	1.0	778	PASS
69	198	0.00	100	34.3	78170	PASS
70	69	0.00	2	0.6	487	PASS
127	198	10	80	42.7	97208	PASS
197	198	0.00	2	0.2	369	PASS
198	198	100	100	100.0	227861	PASS
199	198	5	9	6.8	15593	PASS
275	198	10	60	27.6	62853	PASS
365	198	1	100	3.0	6728	PASS
441	198	0.01	100	14.9	33922	PASS
442	198	50	100	96.5	219861	PASS
443	442	15	24	19.3	42378	PASS

m/z	Abundance
37.10	450.0
38.00	891.0
39.00	5874.0
39.90	671.0
40.90	368.0
44.00	2039.0
45.00	335.0
49.10	553.0
50.00	19664.0
51.00	78368.0
52.00	4197.0
54.90	715.0
56.00	2802.0
57.00	5748.0
58.00	311.0
61.00	1298.0
61.90	1320.0
63.00	3079.0
64.00	308.0
65.00	1668.0
68.00	1440.0
69.00	84856.0
70.00	362.0
73.00	956.0
74.00	8186.0
75.00	13087.0
76.00	4240.0
77.00	94472.0
78.00	6139.0
79.00	5863.0
80.00	4651.0
81.00	6645.0
82.00	1673.0
82.90	1815.0
83.90	474.0
85.00	1396.0
85.90	1724.0
87.00	943.0
88.00	432.0
91.00	1406.0
91.90	1657.0
93.00	9636.0
94.00	707.0
95.90	622.0
98.00	6654.0
99.00	6029.0
100.00	457.0
101.00	3760.0
103.00	1358.0
103.90	2096.0
104.90	2285.0
106.10	493.0
107.00	26816.0
108.00	4237.0
109.10	1091.0
110.00	52848.0
111.00	8378.0
111.90	828.0
116.10	1630.0
117.00	19000.0
118.00	1518.0
120.00	445.0
121.90	1955.0
122.90	3028.0
124.00	1618.0
125.00	1234.0
127.00	101616.0
128.00	7361.0
129.00	36400.0
130.00	3343.0
131.00	736.0
132.10	302.0
134.00	1090.0
135.00	3305.0
136.00	1261.0
137.00	1346.0
138.10	327.0
140.00	636.0
141.00	4577.0
142.00	1519.0
142.90	809.0
144.90	363.0
146.00	1255.0
146.90	2896.0
147.90	4777.0
148.90	1212.0
150.00	349.0
151.10	619.0
152.00	309.0
152.90	1595.0
153.90	942.0

Instrument :
BNA_N
ClientSampleId :

155.00	3110.0
156.00	4488.0
157.00	1152.0
157.80	1039.0
159.00	511.0
160.00	1364.0
161.00	2153.0
162.00	813.0
165.00	1527.0
166.00	1320.0
167.00	8836.0
167.90	4591.0
169.10	875.0
170.10	382.0
171.90	720.0
172.90	1281.0
173.90	1918.0
175.00	3729.0
176.00	1206.0
176.90	1894.0
177.80	743.0
178.90	7111.0
180.00	4929.0
181.00	2340.0
181.80	469.0
184.00	654.0
185.00	3406.0
186.00	27248.0
187.00	7668.0
187.90	751.0
188.90	1774.0
191.00	748.0
191.90	2115.0
193.00	2556.0
194.00	622.0
194.90	321.0
195.90	5658.0
197.90	223360.0
198.90	16310.0
200.00	1201.0
201.50	931.0
202.90	1458.0
204.00	7779.0
205.00	12070.0
206.00	54768.0
207.00	7697.0
208.00	1873.0
208.90	565.0
210.10	967.0
211.00	2015.0
214.90	536.0
215.90	1173.0
217.00	14402.0
218.00	1873.0
221.10	8893.0
221.70	1934.0
223.00	3429.0
224.00	29264.0
225.00	7776.0
226.10	703.0
227.00	10864.0
228.00	1505.0
229.00	2529.0
230.10	530.0
230.90	1151.0
232.90	309.0
233.90	708.0
235.00	1052.0
235.90	570.0
236.90	1228.0
238.80	507.0
240.00	324.0
241.00	705.0
242.00	1286.0
243.00	1405.0
244.00	22808.0
245.00	3372.0
245.90	4273.0
246.90	890.0
248.90	699.0
252.90	466.0
255.00	119432.0
256.00	18848.0
256.90	1458.0
258.00	6084.0
259.00	970.0
264.00	360.0
265.00	2563.0
265.80	506.0
273.00	3457.0
274.00	9211.0
275.00	57288.0
276.00	7988.0
277.00	4501.0
277.90	823.0

Instrument :
BNA_N
ClientSampleId :

283.10	503.0
283.90	349.0
285.10	659.0
292.90	1144.0
294.00	386.0
296.00	13185.0
296.90	2080.0
301.90	303.0
303.10	1320.0
304.10	407.0
314.00	448.0
314.90	1510.0
316.00	915.0
321.00	453.0
323.00	4868.0
323.90	1164.0
327.00	1028.0
327.90	606.0
331.90	390.0
333.00	469.0
333.90	3228.0
335.00	660.0
340.90	495.0
345.90	909.0
352.00	1634.0
353.00	1068.0
354.00	1607.0
355.00	372.0
364.90	5655.0
365.90	842.0
371.10	423.0
372.00	2568.0
373.00	615.0
382.90	597.0
389.90	491.0
401.90	1277.0
402.90	1674.0
404.00	585.0
421.00	1613.0
421.90	1284.0
423.00	9321.0
424.00	2565.0
425.00	375.0
441.00	25504.0
442.00	167616.0
443.00	32712.0
444.00	2868.0

Instrument :
BNA_N
ClientSampleId :

m/z	Abundance
38.00	991.0
39.00	6199.0
40.00	583.0
41.00	343.0
44.00	1657.0
49.00	771.0
50.00	21152.0
51.00	84176.0
52.00	4634.0
55.00	463.0
56.00	2985.0
57.00	6248.0
61.00	974.0
62.00	1460.0
63.00	3880.0
64.00	520.0
65.00	1844.0
69.00	93800.0
69.90	754.0
73.00	977.0
74.00	8851.0
75.00	14540.0
76.00	4678.0
77.00	107432.0
78.00	7033.0
79.00	6342.0
80.00	5019.0
81.00	7459.0
82.00	1818.0
83.00	1463.0
85.00	1178.0
85.90	2135.0
86.90	799.0
88.00	311.0
90.90	1801.0
92.00	1864.0
92.90	11550.0
93.90	891.0
95.90	574.0
98.00	8381.0
99.00	6040.0
100.00	657.0
101.00	3922.0
103.00	1638.0
103.90	2718.0
104.90	2518.0
107.00	31104.0
108.00	4883.0
110.00	59688.0
111.00	8295.0
112.00	884.0
112.90	366.0
116.00	1657.0
117.00	21768.0
117.90	1518.0
119.90	537.0
121.90	1638.0
122.90	2875.0
124.00	1395.0
125.00	1497.0
127.00	115784.0
128.00	9160.0
129.00	41928.0
129.90	3451.0
130.90	538.0
132.10	460.0
133.90	1000.0
135.00	3978.0
135.90	1344.0
137.00	1747.0
137.80	472.0
140.00	689.0
141.00	5942.0
141.90	1831.0
142.90	1393.0
144.10	453.0
146.10	975.0
147.00	3120.0
147.90	5849.0
149.10	1295.0
150.00	445.0
151.10	889.0
151.80	506.0
152.90	1535.0
153.90	1612.0
155.00	2914.0
155.90	4934.0
157.10	1068.0
158.00	867.0
158.80	727.0
160.00	1856.0

Instrument :
BNA_N
ClientSampleId :

161.00	2490.0
162.00	749.0
163.90	383.0
165.00	2021.0
166.00	1895.0
167.00	11026.0
168.00	4804.0
169.00	1022.0
170.00	422.0
170.90	565.0
172.00	1214.0
173.00	1226.0
174.00	2478.0
175.00	4113.0
176.00	1464.0
177.00	1849.0
177.90	735.0
178.90	7890.0
180.00	5768.0
181.00	2954.0
181.90	497.0
183.90	659.0
185.00	4403.0
186.00	34336.0
187.00	10077.0
187.90	995.0
189.00	1887.0
189.90	399.0
190.90	1234.0
192.00	2515.0
193.00	2991.0
193.90	889.0
194.90	406.0
196.00	7199.0
197.90	273664.0
198.90	18192.0
199.90	1204.0
201.60	1223.0
202.90	1587.0
204.00	9046.0
205.00	14918.0
206.00	69096.0
207.00	9066.0
208.00	1572.0
208.90	738.0
210.20	1039.0
210.90	2607.0
214.90	725.0
216.10	1703.0
216.90	18384.0
217.90	2194.0
219.00	343.0
220.00	337.0
221.00	12656.0
223.00	3902.0
224.00	36552.0
225.00	9007.0
226.00	1252.0
226.90	13873.0
228.00	2294.0
229.00	3528.0
229.80	403.0
231.00	1258.0
233.90	953.0
234.80	942.0
236.00	573.0
237.00	1341.0
237.90	365.0
238.90	491.0
240.90	753.0
242.00	1950.0
243.00	2480.0
244.00	28104.0
245.00	3820.0
245.90	5549.0
247.00	1001.0
249.00	1162.0
252.80	520.0
255.00	160256.0
256.00	23640.0
257.00	1902.0
257.90	7438.0
258.90	1505.0
265.00	3674.0
265.90	416.0
271.00	375.0
271.90	374.0
273.00	4481.0
274.00	11728.0
275.00	76336.0
276.00	10067.0
277.00	4864.0
277.90	882.0
282.90	840.0
283.90	362.0

Instrument :
BNA_N
ClientSampleId :

285.00	1135.0
293.00	1353.0
293.90	381.0
296.00	18248.0
297.00	2476.0
297.90	419.0
301.00	354.0
301.90	369.0
303.00	2245.0
304.00	596.0
307.90	350.0
313.80	889.0
315.00	2214.0
316.00	1174.0
317.00	344.0
321.00	690.0
321.90	415.0
323.00	6493.0
324.00	1123.0
326.90	1324.0
328.00	691.0
331.90	447.0
332.90	602.0
334.00	3989.0
335.00	1296.0
341.00	761.0
345.90	1487.0
352.00	2392.0
353.00	1261.0
354.00	2441.0
355.10	443.0
364.90	8331.0
365.90	962.0
371.00	771.0
372.00	3404.0
373.10	1144.0
383.00	1236.0
383.80	474.0
389.90	610.0
391.00	468.0
391.90	366.0
401.90	1996.0
403.00	2762.0
403.90	955.0
421.00	2045.0
422.00	1658.0
423.00	15110.0
424.00	3918.0
424.80	405.0
441.00	39208.0
442.00	261568.0
443.00	50592.0
444.00	4735.0

Instrument :
BNA_N
ClientSampleId :

m/z	Abundance
38.00	621.0
39.00	4031.0
40.00	735.0
44.00	1939.0
48.90	324.0
50.00	12430.0
51.00	48040.0
52.00	2978.0
54.90	328.0
55.90	1567.0
57.00	3855.0
61.00	812.0
62.00	879.0
63.00	2263.0
64.00	394.0
65.00	1011.0
68.10	894.0
69.00	55856.0
69.90	345.0
72.90	584.0
74.00	5457.0
75.00	8798.0
76.10	2897.0
77.00	63848.0
78.00	4150.0
79.00	3645.0
80.00	3326.0
81.00	4040.0
82.00	1028.0
83.00	867.0
85.00	918.0
86.00	1443.0
87.00	535.0
90.90	971.0
91.90	1307.0
93.00	6681.0
93.90	501.0
96.00	389.0
98.00	5369.0
99.00	3944.0
99.90	385.0
100.90	2518.0
102.90	939.0
103.90	1632.0
104.90	1691.0
106.00	586.0
107.00	18920.0
108.00	3046.0
110.00	38232.0
111.00	5533.0
111.90	749.0
116.00	1010.0
117.00	13441.0
118.00	948.0
120.00	441.0
122.00	1082.0
122.90	1903.0
123.90	829.0
125.00	941.0
127.00	74224.0
128.00	5423.0
129.00	27840.0
129.90	2337.0
131.00	596.0
134.00	895.0
135.00	2436.0
135.90	853.0
137.00	1216.0
139.90	357.0
141.00	3756.0
142.00	1168.0
142.90	709.0
146.00	756.0
147.00	1498.0
147.90	4134.0
148.90	851.0
150.10	398.0
151.00	543.0
152.00	458.0
152.90	1243.0
154.00	788.0
154.90	1944.0
156.10	2341.0
156.80	722.0
157.90	650.0
159.00	593.0
159.80	1173.0
161.00	1930.0
162.00	487.0
164.90	1699.0
166.00	1171.0

Instrument :
BNA_N
ClientSampleId :

167.00	6941.0
167.90	3012.0
169.00	423.0
170.20	310.0
171.10	349.0
172.00	558.0
173.00	1081.0
174.00	1664.0
174.90	3187.0
175.90	1015.0
176.90	1756.0
177.90	486.0
178.90	5840.0
179.90	4210.0
181.00	1843.0
183.90	591.0
185.00	2910.0
186.00	23512.0
187.00	7207.0
187.90	586.0
188.90	1220.0
191.00	823.0
192.00	1853.0
192.90	2139.0
194.00	451.0
194.90	352.0
196.00	5714.0
196.80	1109.0
197.90	186560.0
198.90	12278.0
200.00	900.0
201.30	790.0
203.00	1114.0
204.00	6202.0
205.00	10713.0
206.00	46160.0
207.00	6464.0
208.00	1720.0
208.90	614.0
210.00	906.0
211.10	2049.0
214.80	564.0
216.00	999.0
216.90	11985.0
218.00	1322.0
221.00	9717.0
221.70	2171.0
222.90	2764.0
224.00	27352.0
225.00	6498.0
226.10	722.0
227.00	10371.0
227.90	1353.0
229.00	2479.0
229.90	348.0
230.90	1031.0
234.00	717.0
234.90	632.0
236.00	450.0
237.00	1004.0
238.90	335.0
239.90	367.0
240.90	751.0
242.00	1526.0
243.00	1627.0
244.00	20968.0
245.00	2746.0
245.90	4109.0
246.90	709.0
248.90	670.0
251.90	323.0
253.00	601.0
254.00	790.0
255.00	116904.0
256.00	16960.0
256.90	941.0
258.00	5670.0
258.90	966.0
264.90	2480.0
265.80	347.0
273.00	3408.0
274.00	9266.0
275.00	54936.0
276.00	7130.0
277.00	3820.0
277.80	628.0
283.10	661.0
285.00	761.0
293.00	1261.0
296.00	13588.0
297.00	1773.0
303.00	1518.0
304.00	593.0
313.90	674.0
315.00	1570.0

Instrument :
BNA_N
ClientSampleId :

315.90	887.0
320.90	486.0
323.00	4476.0
324.00	947.0
327.00	934.0
328.00	539.0
332.90	510.0
333.90	3292.0
335.00	1067.0
340.90	796.0
345.90	1024.0
351.90	1636.0
352.90	1052.0
354.00	1788.0
354.90	531.0
365.00	6198.0
366.00	962.0
370.80	513.0
372.00	3113.0
373.00	697.0
383.00	852.0
389.90	362.0
402.00	1425.0
403.00	2012.0
404.00	716.0
421.10	1589.0
422.00	1677.0
423.00	13278.0
424.00	2564.0
425.00	308.0
441.00	37056.0
442.00	230400.0
443.00	43832.0
444.00	3755.0

Instrument :

BNA_N

ClientSampleId :

Scan 2453 (17.416 min): BN003063.D
DFTPP35

m/z	Abundance
40.00	417.0
43.90	2125.0
207.00	915.0

Instrument :
BNA_N
ClientSampleId :