

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN101119\
 Data File : BN008108.D
 Acq On : 12 Oct 2019 11:45
 Operator : JU
 Sample : PB123769BL
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK69

Manual Integrations
 APPROVED

mohammad
 10/15/2019 7:58:32 AM

Quant Time: Oct 13 03:05:21 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN101019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 11 19:05:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	2473	0.40	ng/ul	0.00
2) Naphthalene-d8	10.39	136	11883	0.40	ng/ul	-0.01
6) Acenaphthene-d10	14.26	164	7975	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.01	188	18559m	0.40	ng/ul	0.00
16) Chrysene-d12	21.21	240	19478m	0.40	ng/ul	0.00
20) Perylene-d12	23.40	264	20956	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.99	152	5792	0.29	ng/ul	0.00
14) Fluoranthene-d10	19.05	212	19378m	0.30	ng/ul	0.00
Target Compounds						
7) Acenaphthylene	13.98	152	1830	0.048	ng/ul#	Ovalue 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN101119\
Data File : BN008108.D
Acq On : 12 Oct 2019 11:45
Operator : JU
Sample : PB123769BL
Misc :
ALS Vial : 41 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SBLK69

Manual Integrations
APPROVED

mohammad
10/15/2019 7:58:32 AM

Quant Time: Oct 13 03:05:21 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN101019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 11 19:05:54 2019
Response via : Initial Calibration





#7

Acenaphthylene

Concen: 0.048 ng/ul

RT: 13.98 min Scan# 2993

Delta R.T. -0.01 min

Lab File: BN008108.D

Acq: 12 Oct 2019 11:45

Instrument :

BNA_N

ClientSampleId :

SBLK69

Tot Ion:	152	Resp:	1830
Ion Ratio	Lower	Upper	
152	100		
151	28.8	16.2	24.4#
153	29.2	11.0	16.4#

Manual Integrations
APPROVED

mohammad
10/15/2019 7:58:32 AM

