




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101121\
Data File : BN016880.D
Acq On : 11 Oct 2021 11:42
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 11 13:04:36 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 30 18:14:09 2021
Response via : Initial Calibration



TIC: BN016880.D\data.ms

(70) Pentachlorophenol (C)




16.651min (-0.024) 34327.99 ng m

response 355874

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	66.54
264.00	61.60	64.09
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101121\
Data File : BN016880.D
Acq On : 11 Oct 2021 11:42
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 11 13:04:36 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 30 18:14:09 2021
Response via : Initial Calibration



TIC: BN016880.D\data.ms

(77) Benzidine

19.275min (-0.018) 0.00 ng

response 1832815

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.08
183.00	13.20	11.52
0.00	0.00	0.00