

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101123\
 Data File : BN028232.D
 Acq On : 11 Oct 2023 14:48
 Operator : MA/JU
 Sample : PB156212BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB156212BL

Quant Time: Oct 11 15:16:02 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN100223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 10 00:42:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.922	152	2754	0.400	ng	0.00
7) Naphthalene-d8	10.746	136	8852	0.400	ng	0.01
13) Acenaphthene-d10	14.566	164	5237	0.400	ng	0.00
19) Phenanthrene-d10	17.331	188	12351	0.400	ng	# 0.01
29) Chrysene-d12	21.524	240	12601	0.400	ng	# 0.00
35) Perylene-d12	24.001	264	6930	0.400	ng	# 0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.459	112	3606	0.439	ng	0.00
5) Phenol-d6	7.091	99	4271	0.412	ng	0.00
8) Nitrobenzene-d5	9.134	82	3050	0.405	ng	0.01
11) 2-Methylnaphthalene-d10	12.328	152	5467	0.417	ng	0.00
14) 2,4,6-Tribromophenol	16.077	330	1209	0.504	ng	0.00
15) 2-Fluorobiphenyl	13.197	172	7553	0.400	ng	0.01
27) Fluoranthene-d10	19.360	212	14507	0.468	ng	0.00
31) Terphenyl-d14	19.946	244	12353	0.420	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101123\
Data File : BN028232.D
Acq On : 11 Oct 2023 14:48
Operator : MA/JU
Sample : PB156212BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB156212BL

Quant Time: Oct 11 15:16:02 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN100223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 10 00:42:22 2023
Response via : Initial Calibration

