

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101123\
 Data File : BN028238.D
 Acq On : 11 Oct 2023 18:26
 Operator : MA/JU
 Sample : 04810-05
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-RW10A-GW-640-642

Quant Time: Oct 12 04:05:27 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN100223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 12 04:03:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.922	152	2380	0.400	ng	0.00
7) Naphthalene-d8	10.735	136	8295	0.400	ng	# 0.00
13) Acenaphthene-d10	14.566	164	5235	0.400	ng	0.00
19) Phenanthrene-d10	17.331	188	11917	0.400	ng	# 0.01
29) Chrysene-d12	21.524	240	12729	0.400	ng	0.00
35) Perylene-d12	23.999	264	13747	0.400	ng	0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.459	112	927	0.131	ng	0.00
5) Phenol-d6	7.098	99	736	0.082	ng	0.00
8) Nitrobenzene-d5	9.134	82	1886	0.267	ng	0.01
11) 2-Methylnaphthalene-d10	12.324	152	3886	0.316	ng	0.00
14) 2,4,6-Tribromophenol	16.077	330	795	0.332	ng	0.00
15) 2-Fluorobiphenyl	13.187	172	5760	0.305	ng	0.00
27) Fluoranthene-d10	19.360	212	10791	0.361	ng	0.00
31) Terphenyl-d14	19.946	244	10709	0.361	ng	0.00
Target Compounds						
28) Fluoranthene	19.393	202	1122	0.030	ng	# 64
30) Pyrene	19.760	202	1108	0.024	ng	# 79
32) Benzo(a)anthracene	21.506	228	1346	0.031	ng	# 47
33) Chrysene	21.560	228	1577	0.038	ng	# 57
34) Bis(2-ethylhexyl)phtha...	21.381	149	9582	0.336	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101123\
Data File : BN028238.D
Acq On : 11 Oct 2023 18:26
Operator : MA/JU
Sample : 04810-05
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-RW10A-GW-640-642

Quant Time: Oct 12 04:05:27 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN100223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 12 04:03:12 2023
Response via : Initial Calibration

