

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101123\
 Data File : BN028239.D
 Acq On : 11 Oct 2023 19:02
 Operator : MA/JU
 Sample : 04810-09
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-RW10A-GW-678-680

Quant Time: Oct 12 04:05:40 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN100223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 12 04:03:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.922	152	2816	0.400	ng	0.00
7) Naphthalene-d8	10.735	136	9606	0.400	ng	# 0.00
13) Acenaphthene-d10	14.566	164	6059	0.400	ng	0.00
19) Phenanthrene-d10	17.331	188	13588	0.400	ng	# 0.01
29) Chrysene-d12	21.524	240	13832	0.400	ng	0.00
35) Perylene-d12	23.996	264	14165	0.400	ng	0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.459	112	1215	0.145	ng	0.00
5) Phenol-d6	7.106	99	975	0.092	ng	0.01
8) Nitrobenzene-d5	9.134	82	2473	0.303	ng	0.01
11) 2-Methylnaphthalene-d10	12.324	152	4563	0.321	ng	0.00
14) 2,4,6-Tribromophenol	16.078	330	1237	0.446	ng	0.00
15) 2-Fluorobiphenyl	13.187	172	6802	0.312	ng	0.00
27) Fluoranthene-d10	19.360	212	13621	0.399	ng	0.00
31) Terphenyl-d14	19.946	244	12110	0.375	ng	0.00
Target Compounds						
28) Fluoranthene	19.388	202	1241	0.029	ng	# 64
30) Pyrene	19.760	202	1194	0.024	ng	# 77
32) Benzo(a)anthracene	21.507	228	1589	0.034	ng	# 55
33) Chrysene	21.560	228	1704	0.038	ng	# 66
34) Bis(2-ethylhexyl)phtha...	21.372	149	12304	0.397	ng	99
36) Indeno(1,2,3-cd)pyrene	26.604	276	1358	0.029	ng	# 85
41) Benzo(g,h,i)perylene	27.413	276	1256	0.031	ng	# 9

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101123\
Data File : BN028239.D
Acq On : 11 Oct 2023 19:02
Operator : MA/JU
Sample : 04810-09
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-RW10A-GW-678-680

Quant Time: Oct 12 04:05:40 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN100223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 12 04:03:12 2023
Response via : Initial Calibration

