

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101123\
 Data File : BN028251.D
 Acq On : 12 Oct 2023 02:51
 Operator : MA/JU
 Sample : PB156246BL
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB156246BL

Quant Time: Oct 12 04:09:23 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN100223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 12 04:03:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.922	152	2153	0.400	ng	0.00
7) Naphthalene-d8	10.746	136	7098	0.400	ng	# 0.01
13) Acenaphthene-d10	14.566	164	4431	0.400	ng	0.00
19) Phenanthrene-d10	17.331	188	10612	0.400	ng	0.01
29) Chrysene-d12	21.524	240	10763	0.400	ng	0.00
35) Perylene-d12	23.990	264	7481	0.400	ng	# 0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.459	112	3008	0.468	ng	0.00
5) Phenol-d6	7.105	99	3249	0.401	ng	0.01
8) Nitrobenzene-d5	9.144	82	2338	0.387	ng	0.02
11) 2-Methylnaphthalene-d10	12.339	152	4511	0.429	ng	0.01
14) 2,4,6-Tribromophenol	16.090	330	825	0.407	ng	0.01
15) 2-Fluorobiphenyl	13.197	172	6136	0.384	ng	0.01
27) Fluoranthene-d10	19.360	212	12727	0.478	ng	0.00
31) Terphenyl-d14	19.946	244	10864	0.433	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101123\
Data File : BN028251.D
Acq On : 12 Oct 2023 02:51
Operator : MA/JU
Sample : PB156246BL
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB156246BL

Quant Time: Oct 12 04:09:23 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN100223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 12 04:03:12 2023
Response via : Initial Calibration

