




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN101218\
Data File : BN003112.D
Acq On : 11 Oct 2018 20:36
Operator : MJ/SJ
Sample : J5368-05
Misc :
ALS Vial : 14 Sample Multiplier: 1

Quant Time: Oct 12 06:45:05 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN100918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 12 05:29:36 2018
Response via : Initial Calibration



TIC: BN003112.D

(12) Phenanthrene

17.057min (-0.013) 0.06ng/ul m

response 1441

Ion	Exp%	Act%
178.00	100	100
179.00	16.10	21.17#
176.00	19.30	22.46
0.00	0.00	0.00