




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN101218\  
Data File : BN003128.D  
Acq On : 12 Oct 2018 06:00  
Operator : MJ/SJ  
Sample : J5234-16  
Misc :  
ALS Vial : 29 Sample Multiplier: 1

Quant Time: Oct 12 07:52:06 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN100918.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Oct 12 07:27:48 2018  
Response via : Initial Calibration



TIC: BN003128.D

(16) Chrysene-d12 (I)

21.306min (+0.074) 0.40ng/ul

response 954

| Ion    | Exp%  | Act%    |
|--------|-------|---------|
| 240.00 | 100   | 100     |
| 120.00 | 16.40 | 58.99#  |
| 236.00 | 28.10 | 111.57# |
| 0.00   | 0.00  | 0.00    |