

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101223\
 Data File : BN028268.D
 Acq On : 12 Oct 2023 19:35
 Operator : MA/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Oct 13 01:47:48 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN101223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 13 01:45:19 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.886	152	2668	0.400	ng	0.00
7) Naphthalene-d8	10.703	136	8554	0.400	ng	-0.01
13) Acenaphthene-d10	14.534	164	5414	0.400	ng	-0.01
19) Phenanthrene-d10	17.294	188	11839	0.400	ng	#-0.01
29) Chrysene-d12	21.497	240	12521	0.400	ng	0.00
35) Perylene-d12	23.958	264	14426	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.423	112	23764	3.440	ng	0.00
5) Phenol-d6	7.048	99	30661	3.513	ng	0.00
8) Nitrobenzene-d5	9.080	82	23781	3.491	ng	-0.02
11) 2-Methylnaphthalene-d10	12.290	152	41457	3.356	ng	0.00
14) 2,4,6-Tribromophenol	16.040	330	8300	3.495	ng	-0.01
15) 2-Fluorobiphenyl	13.154	172	60255	3.283	ng	-0.01
27) Fluoranthene-d10	19.333	212	101247	3.275	ng	0.00
31) Terphenyl-d14	19.922	244	81287	3.320	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.299	88	9554	2.927	ng	# 95
3) n-Nitrosodimethylamine	3.653	42	11323	3.022	ng	# 99
6) bis(2-Chloroethyl)ether	7.329	93	25620	3.125	ng	97
9) Naphthalene	10.757	128	70814	3.044	ng	97
10) Hexachlorobutadiene	10.970	225	11974	2.965	ng	# 99
12) 2-Methylnaphthalene	12.366	142	50670	3.107	ng	99
16) Acenaphthylene	14.266	152	78464	3.365	ng	100
17) Acenaphthene	14.598	154	49203	3.061	ng	99
18) Fluorene	15.581	166	64994	2.819	ng	96
20) 4,6-Dinitro-2-methylph...	15.705	198	5025	5.227	ng	# 56
21) 4-Bromophenyl-phenylether	16.475	248	20029	3.311	ng	# 78
22) Hexachlorobenzene	16.561	284	20015	2.951	ng	99
23) Atrazine	16.760	200	18503	3.961	ng	# 95
24) Pentachlorophenol	16.934	266	9951	3.537	ng	95
25) Phenanthrene	17.343	178	103347	3.031	ng	99
26) Anthracene	17.430	178	100860	3.387	ng	100
28) Fluoranthene	19.365	202	126225	2.978	ng	100
30) Pyrene	19.732	202	130900	2.939	ng	100
32) Benzo(a)anthracene	21.480	228	136806	3.127	ng	97
33) Chrysene	21.533	228	133765	2.855	ng	97
34) Bis(2-ethylhexyl)phtha...	21.354	149	99652	4.541	ng	# 98
36) Indeno(1,2,3-cd)pyrene	26.530	276	179206	2.956	ng	99
37) Benzo(b)fluoranthene	23.192	252	140409	2.522	ng	# 88
38) Benzo(k)fluoranthene	23.241	252	143518	2.528	ng	# 88
39) Benzo(a)pyrene	23.844	252	140718	2.752	ng	# 86
40) Dibenzo(a,h)anthracene	26.539	278	147571	2.976	ng	# 90
41) Benzo(g,h,i)perylene	27.334	276	150422	2.782	ng	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101223\
Data File : BN028268.D
Acq On : 12 Oct 2023 19:35
Operator : MA/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Quant Time: Oct 13 01:47:48 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN101223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 13 01:45:19 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.886 min Scan# 694
Delta R.T. -0.007 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 2668
Ion Ratio Lower Upper
152 100
150 155.4 121.6 182.4
115 71.2 54.3 81.5



#2
1,4-Dioxane
Concen: 2.927 ng
RT: 3.299 min Scan# 58
Delta R.T. -0.000 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35

Tgt Ion: 88 Resp: 9554
Ion Ratio Lower Upper
88 100
43 26.3 26.6 39.8#
58 68.1 55.4 83.0









#3
n-Nitrosodimethylamine
Concen: 3.022 ng
RT: 3.653 min Scan# 108
Delta R.T. -0.007 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 11323
Ion Ratio Lower Upper
42 100
74 140.6 112.6 168.8
44 5.6 7.8 11.6#



#4
2-Fluorophenol
Concen: 3.440 ng
RT: 5.423 min Scan# 352
Delta R.T. -0.000 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35

Tgt Ion: 112 Resp: 23764
Ion Ratio Lower Upper
112 100
64 58.3 46.9 70.3
63 30.5 25.0 37.4









#5
Phenol-d6
Concen: 3.513 ng
RT: 7.048 min Scan# 51
Delta R.T. -0.007 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 99 Resp: 30661
Ion Ratio Lower Upper
99 100
42 19.1 15.9 23.9
71 35.3 28.6 42.8



#6
bis(2-Chloroethyl)ether
Concen: 3.125 ng
RT: 7.329 min Scan# 616
Delta R.T. -0.000 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35

Tgt Ion: 93 Resp: 25620
Ion Ratio Lower Upper
93 100
63 75.4 62.6 93.8
95 33.0 27.2 40.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.703 min Scan# 91
Delta R.T. -0.011 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	8554
Ion Ratio	Lower	Upper	
136	100		
137	12.0	9.8	14.8
54	8.6	6.2	9.4
68	8.1	5.7	8.5



#8
Nitrobenzene-d5
Concen: 3.491 ng
RT: 9.080 min Scan# 824
Delta R.T. -0.021 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35

Tgt Ion:	82	Resp:	23781
Ion Ratio	Lower	Upper	
82	100		
128	35.4	36.8	55.2#
54	49.9	37.4	56.2









#9
Naphthalene
Concen: 3.044 ng
RT: 10.757 min Scan# 981
Delta R.T. -0.000 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:128 Resp: 70814
Ion Ratio Lower Upper
128 100
129 11.1 9.8 14.6
127 13.0 11.4 17.0



#10
Hexachlorobutadiene
Concen: 2.965 ng
RT: 10.970 min Scan# 1001
Delta R.T. -0.000 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35

Tgt Ion:225 Resp: 11974
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.9 51.6 77.4









#11
2-Methylnaphthalene-d10
Concen: 3.356 ng
RT: 12.290 min Scan# 1259
Delta R.T. -0.008 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 41457
Ion Ratio Lower Upper
152 100
151 20.4 16.1 24.1



#12
2-Methylnaphthalene
Concen: 3.107 ng
RT: 12.366 min Scan# 1279
Delta R.T. -0.008 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35

Tgt Ion:142 Resp: 50670
Ion Ratio Lower Upper
142 100
141 88.5 71.1 106.7
115 35.6 29.9 44.9





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.534 min Scan# 1

Delta R.T. -0.011 min

Lab File: BN028268.D

Acq: 12 Oct 2023 19:35

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:164 Resp: 5414

Ion Ratio Lower Upper

164 100

162 104.6 81.1 121.7

160 48.6 36.6 54.8



#14

2,4,6-Tribromophenol

Concen: 3.495 ng

RT: 16.040 min Scan# 1725

Delta R.T. -0.012 min

Lab File: BN028268.D

Acq: 12 Oct 2023 19:35




Tgt Ion:330 Resp: 8300


Ion Ratio Lower Upper

330 100

332 98.0 74.1 111.1

141 41.5 32.1 48.1









#15
2-Fluorobiphenyl
Concen: 3.283 ng
RT: 13.154 min Scan# 1462
Delta R.T. -0.011 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:172 Resp: 60255
Ion Ratio Lower Upper
172 100
171 35.5 28.3 42.5
170 23.5 18.7 28.1



#16
Acenaphthylene
Concen: 3.365 ng
RT: 14.266 min Scan# 1566
Delta R.T. -0.000 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35

Tgt Ion:152 Resp: 78464
Ion Ratio Lower Upper
152 100
151 19.6 15.6 23.4
153 13.1 10.6 16.0









#17
Acenaphthene
Concen: 3.061 ng
RT: 14.598 min Scan# 1597
Delta R.T. -0.000 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 49203
Ion Ratio Lower Upper
154 100
153 112.7 90.7 136.1
152 53.8 43.6 65.4



#18
Fluorene
Concen: 2.819 ng
RT: 15.581 min Scan# 1688
Delta R.T. -0.012 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35

Tgt Ion:166 Resp: 64994
Ion Ratio Lower Upper
166 100
165 94.7 79.3 118.9
167 13.5 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.294 min Scan# 11
Delta R.T. -0.012 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:188 Resp: 11839
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 14.4 7.1 10.7#



#20
4,6-Dinitro-2-methylphenol
Concen: 5.227 ng
RT: 15.705 min Scan# 1698
Delta R.T. -0.025 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35

Tgt Ion:198 Resp: 5025
Ion Ratio Lower Upper
198 100
51 63.5 96.1 144.1#
105 60.3 75.4 113.0#









#21
4-Bromophenyl-phenylether
Concen: 3.311 ng
RT: 16.475 min Scan# 1760
Delta R.T. -0.012 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:248 Resp: 20029
Ion Ratio Lower Upper
248 100
250 94.1 80.5 120.7
141 85.7 38.9 58.3#



#22
Hexachlorobenzene
Concen: 2.951 ng
RT: 16.561 min Scan# 1767
Delta R.T. -0.012 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35

Tgt Ion:284 Resp: 20015
Ion Ratio Lower Upper
284 100
142 39.6 32.1 48.1
249 31.6 25.5 38.3





#23

Atrazine

Concen: 3.961 ng

RT: 16.760 min Scan# 1783

Delta R.T. -0.000 min

Lab File: BN028268.D

Acq: 12 Oct 2023 19:35

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:200 Resp: 18503

Ion Ratio Lower Upper

200 100

173 21.0 21.6 32.4#

215 49.4 39.0 58.6



#24

Pentachlorophenol

Concen: 3.537 ng

RT: 16.934 min Scan# 1797

Delta R.T. -0.012 min

Lab File: BN028268.D

Acq: 12 Oct 2023 19:35




Tgt Ion:266 Resp: 9951


Ion Ratio Lower Upper

266 100

264 62.6 51.5 77.3

268 64.9 56.6 84.8









#25
Phenanthrene
Concen: 3.031 ng
RT: 17.343 min Scan# 1830
Delta R.T. -0.000 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:178 Resp: 103347
Ion Ratio Lower Upper
178 100
176 18.9 15.3 22.9
179 15.2 12.5 18.7



#26
Anthracene
Concen: 3.387 ng
RT: 17.430 min Scan# 1837
Delta R.T. -0.012 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35

Tgt Ion:178 Resp: 100860
Ion Ratio Lower Upper
178 100
176 18.2 14.6 22.0
179 15.3 12.2 18.4









#27
Fluoranthene-d10
Concen: 3.275 ng
RT: 19.333 min Scan# 2101
Delta R.T. -0.005 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:212 Resp: 101247
Ion Ratio Lower Upper
212 100
106 14.7 11.5 17.3
104 8.3 6.7 10.1



#28
Fluoranthene
Concen: 2.978 ng
RT: 19.365 min Scan# 2108
Delta R.T. -0.000 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35

Tgt Ion:202 Resp: 126225
Ion Ratio Lower Upper
202 100
101 12.4 10.1 15.1
203 17.2 13.9 20.9





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.497 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN028268.D

Acq: 12 Oct 2023 19:35

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:240 Resp: 12521

Ion Ratio Lower Upper

240 100

120 13.7 13.4 20.0

236 28.1 22.7 34.1



#30

Pyrene

Concen: 2.939 ng

RT: 19.732 min Scan# 2187

Delta R.T. -0.000 min

Lab File: BN028268.D

Acq: 12 Oct 2023 19:35




Tgt Ion:202 Resp: 130900


Ion Ratio Lower Upper

202 100

200 20.7 16.7 25.1

203 18.0 14.5 21.7









#31
 Terphenyl-d14
 Concen: 3.320 ng
 RT: 19.922 min Scan# 211
 Delta R.T. -0.000 min
 Lab File: BN028268.D
 Acq: 12 Oct 2023 19:35




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:244 Resp: 81287
 Ion Ratio Lower Upper
 244 100
 212 7.9 6.6 10.0
 122 13.6 11.7 17.5



#32
 Benzo(a)anthracene
 Concen: 3.127 ng
 RT: 21.480 min Scan# 2514
 Delta R.T. -0.000 min
 Lab File: BN028268.D
 Acq: 12 Oct 2023 19:35

Tgt Ion:228 Resp: 136806
 Ion Ratio Lower Upper
 228 100
 226 26.8 22.9 34.3
 229 19.7 16.6 24.8





#33

Chrysene

Concen: 2.855 ng

RT: 21.533 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN028268.D

Acq: 12 Oct 2023 19:35

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:228 Resp: 133765

Ion Ratio Lower Upper

228 100

226 30.0 25.2 37.8

229 19.6 16.6 25.0



#34

Bis(2-ethylhexyl)phthalate

Concen: 4.541 ng

RT: 21.354 min Scan# 2500

Delta R.T. -0.000 min

Lab File: BN028268.D

Acq: 12 Oct 2023 19:35




Tgt Ion:149 Resp: 99652


Ion Ratio Lower Upper

149 100

167 27.3 22.6 34.0

279 3.9 4.2 6.2#









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.958 min Scan# 3147
Delta R.T. -0.003 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:264 Resp: 14426




Ion	Ratio	Lower	Upper
264	100		
260	26.4	20.7	31.1
265	53.9	38.3	57.5




#36
Indeno(1,2,3-cd)pyrene
Concen: 2.956 ng
RT: 26.530 min Scan# 4027
Delta R.T. -0.006 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35

Tgt Ion:276 Resp: 179206

Ion	Ratio	Lower	Upper
276	100		
138	27.2	21.3	31.9
277	25.2	20.0	30.0









#37
Benzo(b)fluoranthene
Concen: 2.522 ng
RT: 23.192 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 140409
Ion Ratio Lower Upper
252 100
253 22.6 21.8 32.8
125 12.2 15.6 23.4#



#38
Benzo(k)fluoranthene
Concen: 2.528 ng
RT: 23.241 min Scan# 2902
Delta R.T. -0.009 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35

Tgt Ion:252 Resp: 143518
Ion Ratio Lower Upper
252 100
253 22.2 22.0 33.0
125 12.0 15.3 22.9#









#39
Benzo(a)pyrene
Concen: 2.752 ng
RT: 23.844 min Scan# 3110
Delta R.T. -0.006 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 140718
Ion Ratio Lower Upper
252 100
253 22.5 22.7 34.1#
125 13.2 17.0 25.6#



#40
Dibenzo(a,h)anthracene
Concen: 2.976 ng
RT: 26.539 min Scan# 4030
Delta R.T. -0.012 min
Lab File: BN028268.D
Acq: 12 Oct 2023 19:35

Tgt Ion:278 Resp: 147571
Ion Ratio Lower Upper
278 100
139 18.6 18.8 28.2#
279 24.3 23.5 35.3





#41
 Benzo(g,h,i)perylene
 Concen: 2.782 ng
 RT: 27.334 min Scan# 41
 Delta R.T. -0.009 min
 Lab File: BN028268.D
 Acq: 12 Oct 2023 19:35

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Tgt Ion:	276	Resp:	150422
Ion	Ratio	Lower	Upper
276	100		
277	24.4	21.7	32.5
138	23.7	21.7	32.5

