

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101223\
 Data File : BN028270.D
 Acq On : 12 Oct 2023 20:47
 Operator : MA/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN101223

Quant Time: Oct 13 02:19:23 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN101223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 13 02:17:11 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	77	0.00
2	1,4-Dioxane	0.400	0.374	6.5	71	0.00
3	n-Nitrosodimethylamine	0.400	0.398	0.5	74	0.00
4 S	2-Fluorophenol	0.400	0.362	9.5	74	0.00
5 S	Phenol-d6	0.400	0.354	11.5	71	0.00
6	bis(2-Chloroethyl)ether	0.400	0.355	11.3	69	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	75	0.00
8 S	Nitrobenzene-d5	0.400	0.360	10.0	71	0.00
9	Naphthalene	0.400	0.401	-0.3	74	0.01
10	Hexachlorobutadiene	0.400	0.397	0.8	74	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.390	2.5	72	0.00
12	2-Methylnaphthalene	0.400	0.395	1.3	72	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	72	0.00
14 S	2,4,6-Tribromophenol	0.400	0.357	10.8	68	0.00
15 S	2-Fluorobiphenyl	0.400	0.392	2.0	72	0.00
16	Acenaphthylene	0.400	0.390	2.5	72	0.00
17	Acenaphthene	0.400	0.405	-1.3	71	0.00
18	Fluorene	0.400	0.403	-0.8	71	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	73	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.403	-0.8	52	0.00
21	4-Bromophenyl-phenylether	0.400	0.396	1.0	70	0.00
22	Hexachlorobenzene	0.400	0.410	-2.5	73	0.00
23	Atrazine	0.400	0.386	3.5	68	0.00
24	Pentachlorophenol	0.400	0.308	23.0	68	0.01
25	Phenanthrene	0.400	0.399	0.3	71	0.00
26	Anthracene	0.400	0.390	2.5	71	0.00
27 SURR	Fluoranthene-d10	0.400	0.404	-1.0	71	0.00
28	Fluoranthene	0.400	0.401	-0.3	72	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	74	0.00
30	Pyrene	0.400	0.401	-0.3	72	0.00
31 S	Terphenyl-d14	0.400	0.386	3.5	71	0.00
32	Benzo(a)anthracene	0.400	0.393	1.8	71	0.00
33	Chrysene	0.400	0.415	-3.7	76	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.392	2.0	66	0.00
35 I	Perylene-d12	0.400	0.400	0.0	76	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.392	2.0	78	0.00
37	Benzo(b)fluoranthene	0.400	0.387	3.3	75	0.00
38	Benzo(k)fluoranthene	0.400	0.398	0.5	77	0.00
39 C	Benzo(a)pyrene	0.400	0.391	2.3	75	0.00
40	Dibenzo(a,h)anthracene	0.400	0.391	2.3	77	0.00
41	Benzo(g,h,i)perylene	0.400	0.393	1.8	78	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN101223