

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN101318\
 Data File : BN003159.D
 Acq On : 13 Oct 2018 02:53
 Operator : MJ/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD0.410

Quant Time: Oct 13 04:58:26 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN100918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 13 04:51:37 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	68	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	75	0.00
3	Naphthalene	1.026	1.094	-6.6	82	0.00
4 SURR	2-Methylnaphthalene-d10	0.620	0.685	-10.5	84	0.00
5	2-Methylnaphthalene	0.707	0.777	-9.9	83	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	77	0.00
7	Acenaphthylene	1.830	1.967	-7.5	84	0.00
8 C	Acenaphthene	1.389	1.526	-9.9	85	0.00
9	Fluorene	1.591	1.778	-11.8	86	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	80	0.00
11	Pentachlorophenol	0.039	0.054	-38.5#	163#	0.00
12	Phenanthrene	1.115	1.216	-9.1	88	0.00
13	Anthracene	1.053	1.140	-8.3	88	0.00
14 SURR	Fluoranthene-d10	1.144	1.259	-10.1	91	0.00
15 C	Fluoranthene	1.308	1.440	-10.1	91	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	81	0.00
17	Pyrene	1.554	1.788	-15.1	91	0.00
18	Benzo(a)anthracene	1.287	1.411	-9.6	91	0.00
19	Chrysene	1.307	1.455	-11.3	92	0.00
20 I	Perylene-d12	1.000	1.000	0.0	91	0.00
21	Benzo(b)fluoranthene	1.441	1.448	-0.5	96	0.00
22	Benzo(k)fluoranthene	1.493	1.551	-3.9	97	0.00
23 C	Benzo(a)pyrene	1.291	1.421	-10.1	103	0.00
24	Indeno(1,2,3-cd)pyrene	1.484	1.615	-8.8	100	0.00
25	Dibenzo(a,h)anthracene	1.188	1.297	-9.2	100	0.00
26	Benzo(g,h,i)perylene	1.251	1.362	-8.9	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0