

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101321\
 Data File : BN016907.D
 Acq On : 12 Oct 2021 20:35
 Operator : CG/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Oct 13 01:54:08 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 13 01:50:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.689	152	20501	0.400	ng	0.00
7) Naphthalene-d8	10.445	136	88462	0.400	ng	#-0.01
13) Acenaphthene-d10	14.294	164	47201	0.400	ng	0.00
19) Phenanthrene-d10	17.030	188	87687	0.400	ng	#-0.01
29) Chrysene-d12	21.227	240	90363	0.400	ng	# 0.00
35) Perylene-d12	23.471	264	89484	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.291	112	81094	1.544	ng	0.00
5) Phenol-d6	6.868	99	91984	1.562	ng	0.00
8) Nitrobenzene-d5	8.822	82	100409	1.680	ng	0.00
11) 2-Methylnaphthalene-d10	12.038	152	210008	1.639	ng	0.00
14) 2,4,6-Tribromophenol	15.791	330	39532	1.727	ng	0.00
15) 2-Fluorobiphenyl	12.924	172	294804	1.632	ng	0.00
27) Fluoranthene-d10	19.068	212	390171	1.696	ng	0.00
31) Terphenyl-d14	19.679	244	328212	1.596	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.281	88	24336	1.586	ng	93
3) n-Nitrosodimethylamine	3.604	42	28303	1.689	ng	98
6) bis(2-Chloroethyl)ether	7.126	93	93031	1.617	ng	99
9) Naphthalene	10.495	128	378069	1.568	ng	100
10) Hexachlorobutadiene	10.787	225	74235	1.641	ng	# 100
12) 2-Methylnaphthalene	12.114	142	243675	1.616	ng	99
16) Acenaphthylene	14.015	152	349549	1.736	ng	100
17) Acenaphthene	14.358	154	218224	1.631	ng	98
18) Fluorene	15.347	166	270383	1.674	ng	100
20) 4,6-Dinitro-2-methylph...	15.436	198	25677	2.090	ng	# 87
21) 4-Bromophenyl-phenylether	16.239	248	89822	1.709	ng	96
22) Hexachlorobenzene	16.348	284	98861	1.652	ng	99
23) Atrazine	16.519	200	67604	1.763	ng	99
24) Pentachlorophenol	16.689	266	40616	1.693	ng	100
25) Phenanthrene	17.078	178	422414	1.677	ng	100
26) Anthracene	17.164	178	386569	1.746	ng	100
28) Fluoranthene	19.103	202	473176	1.707	ng	100
30) Pyrene	19.460	202	474793	1.602	ng	100
32) Benzo(a)anthracene	21.206	228	479405	1.657	ng	97
33) Chrysene	21.259	228	483594	1.629	ng	97
34) Bis(2-ethylhexyl)phtha...	21.164	149	237951	1.498	ng	99
36) Indeno(1,2,3-cd)pyrene	25.740	276	580977	1.656	ng	99
37) Benzo(b)fluoranthene	22.797	252	498673	1.667	ng	100
38) Benzo(k)fluoranthene	22.841	252	490034	1.659	ng	# 97
39) Benzo(a)pyrene	23.372	252	457369	1.679	ng	98
40) Dibenzo(a,h)anthracene	25.761	278	475346	1.660	ng	98
41) Benzo(g,h,i)perylene	26.435	276	498855	1.654	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101321\
Data File : BN016907.D
Acq On : 12 Oct 2021 20:35
Operator : CG/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Quant Time: Oct 13 01:54:08 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 13 01:50:22 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.689 min Scan# 500
 Delta R.T. -0.000 min
 Lab File: BN016907.D
 Acq: 12 Oct 2021 20:35




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:152 Resp: 20501
 Ion Ratio Lower Upper
 152 100
 150 154.1 122.6 183.8
 115 57.6 46.1 69.1



#2
 1,4-Dioxane
 Concen: 1.586 ng
 RT: 3.281 min Scan# 22
 Delta R.T. -0.000 min
 Lab File: BN016907.D
 Acq: 12 Oct 2021 20:35

Tgt Ion: 88 Resp: 24336
 Ion Ratio Lower Upper
 88 100
 43 27.8 24.3 36.5
 58 75.0 65.2 97.8









#3
n-Nitrosodimethylamine
Concen: 1.689 ng
RT: 3.604 min Scan# 57
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	42	Resp:	28303
Ion	Ratio	Lower	Upper
42	100		
74	145.0	114.4	171.6
44	6.2	5.7	8.5



#4
2-Fluorophenol
Concen: 1.544 ng
RT: 5.291 min Scan# 240
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35

Tgt Ion:	112	Resp:	81094
Ion	Ratio	Lower	Upper
112	100		
64	55.8	45.0	67.4
63	28.1	22.6	33.8









#5
Phenol-d6
Concen: 1.562 ng
RT: 6.868 min Scan# 411
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35




Instrument :
BNA_N
ClientSampled :
SSTDICC1.6


Tgt Ion: 99 Resp: 91984
Ion Ratio Lower Upper
99 100
42 17.5 14.2 21.4
71 30.6 24.5 36.7



#6
bis(2-Chloroethyl)ether
Concen: 1.617 ng
RT: 7.126 min Scan# 439
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35

Tgt Ion: 93 Resp: 93031
Ion Ratio Lower Upper
93 100
63 74.6 60.3 90.5
95 32.7 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.445 min Scan# 736
Delta R.T. -0.013 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	136	Resp:	88462
Ion Ratio	Lower	Upper	
136	100		
137	10.8	8.7	13.1
54	9.1	3.7	5.5#
68	7.4	3.2	4.8#



#8
Nitrobenzene-d5
Concen: 1.680 ng
RT: 8.822 min Scan# 608
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35

Tgt Ion:	82	Resp:	100409
Ion Ratio	Lower	Upper	
82	100		
128	36.4	28.3	42.5
54	51.7	43.0	64.4









#9
Naphthalene
Concen: 1.568 ng
RT: 10.495 min Scan# 740
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:128 Resp: 378069
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.1 10.6 16.0



#10
Hexachlorobutadiene
Concen: 1.641 ng
RT: 10.787 min Scan# 763
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35

Tgt Ion:225 Resp: 74235
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.1 51.1 76.7









#11
2-Methylnaphthalene-d10
Concen: 1.639 ng
RT: 12.038 min Scan# 939
Delta R.T. -0.004 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 210008
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.4



#12
2-Methylnaphthalene
Concen: 1.616 ng
RT: 12.114 min Scan# 956
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35

Tgt Ion:142 Resp: 243675
Ion Ratio Lower Upper
142 100
141 87.5 70.5 105.7
115 32.4 26.0 39.0









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.294 min Scan# 1185
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:164 Resp: 47201
Ion Ratio Lower Upper
164 100
162 101.2 81.4 122.0
160 45.3 37.0 55.4



#14
2,4,6-Tribromophenol
Concen: 1.727 ng
RT: 15.791 min Scan# 1303
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35

Tgt Ion:330 Resp: 39532
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 30.1 24.4 36.6









#15
2-Fluorobiphenyl
Concen: 1.632 ng
RT: 12.924 min Scan# 1077
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:172 Resp: 294804
Ion Ratio Lower Upper
172 100
171 34.0 27.4 41.0
170 22.1 18.1 27.1



#16
Acenaphthylene
Concen: 1.736 ng
RT: 14.015 min Scan# 1163
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35

Tgt Ion:152 Resp: 349549
Ion Ratio Lower Upper
152 100
151 19.4 15.5 23.3
153 13.1 10.3 15.5









#17
Acenaphthene
Concen: 1.631 ng
RT: 14.358 min Scan# 1190
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 218224
Ion Ratio Lower Upper
154 100
153 112.3 91.2 136.8
152 54.0 44.4 66.6



#18
Fluorene
Concen: 1.674 ng
RT: 15.347 min Scan# 1268
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35

Tgt Ion:166 Resp: 270383
Ion Ratio Lower Upper
166 100
165 97.3 78.1 117.1
167 13.4 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.030 min Scan# 1404
Delta R.T. -0.012 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:188 Resp: 87687
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 13.6 4.9 7.3#



#20
4,6-Dinitro-2-methylphenol
Concen: 2.090 ng
RT: 15.436 min Scan# 1275
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35

Tgt Ion:198 Resp: 25677
Ion Ratio Lower Upper
198 100
182 8.5 11.0 16.6#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 1.709 ng
RT: 16.239 min Scan# 1339
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:248 Resp: 89822
Ion Ratio Lower Upper
248 100
250 98.5 78.2 117.2
141 46.3 43.1 64.7



#22
Hexachlorobenzene
Concen: 1.652 ng
RT: 16.348 min Scan# 1348
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35

Tgt Ion:284 Resp: 98861
Ion Ratio Lower Upper
284 100
142 31.7 25.1 37.7
249 29.6 23.2 34.8





#23

Atrazine

Concen: 1.763 ng

RT: 16.519 min Scan# 1362

Delta R.T. -0.000 min

Lab File: BN016907.D

Acq: 12 Oct 2021 20:35





Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6

Tgt Ion:	200	Resp:	67604
Ion Ratio	Lower	Upper	
200	100		
173	22.0	18.0	27.0
215	49.8	40.0	60.0



#24

Pentachlorophenol

Concen: 1.693 ng




RT: 16.689 min Scan# 1376


Delta R.T. -0.000 min

Lab File: BN016907.D

Acq: 12 Oct 2021 20:35

Tgt Ion:	266	Resp:	40616
Ion Ratio	Lower	Upper	
266	100		
264	62.6	50.5	75.7
268	63.9	51.2	76.8









#25
Phenanthrene
Concen: 1.677 ng
RT: 17.078 min Scan# 1408
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:178 Resp: 422414
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.4
179 15.3 12.3 18.5



#26
Anthracene
Concen: 1.746 ng
RT: 17.164 min Scan# 1415
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35

Tgt Ion:178 Resp: 386569
Ion Ratio Lower Upper
178 100
176 18.2 14.6 21.8
179 15.4 12.2 18.4









#27
Fluoranthene-d10
Concen: 1.696 ng
RT: 19.068 min Scan# 1697
Delta R.T. -0.005 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:212 Resp: 390171
Ion Ratio Lower Upper
212 100
106 13.5 10.7 16.1
104 7.4 5.8 8.8



#28
Fluoranthene
Concen: 1.707 ng
RT: 19.103 min Scan# 1704
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35

Tgt Ion:202 Resp: 473176
Ion Ratio Lower Upper
202 100
101 11.3 9.0 13.4
203 17.4 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.227 min Scan# 2001
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:240 Resp: 90363
Ion Ratio Lower Upper
240 100
120 8.5 8.6 12.8#
236 26.0 21.0 31.4



#30
Pyrene
Concen: 1.602 ng
RT: 19.460 min Scan# 1776
Delta R.T. -0.005 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35

Tgt Ion:202 Resp: 474793
Ion Ratio Lower Upper
202 100
200 20.6 16.5 24.7
203 17.5 14.0 21.0









#31
 Terphenyl-d14
 Concen: 1.596 ng
 RT: 19.679 min Scan# 1820
 Delta R.T. -0.000 min
 Lab File: BN016907.D
 Acq: 12 Oct 2021 20:35




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:244 Resp: 328212
 Ion Ratio Lower Upper
 244 100
 212 7.1 5.7 8.5
 122 11.7 9.8 14.6



#32
 Benzo(a)anthracene
 Concen: 1.657 ng
 RT: 21.206 min Scan# 1999
 Delta R.T. -0.011 min
 Lab File: BN016907.D
 Acq: 12 Oct 2021 20:35

Tgt Ion:228 Resp: 479405
 Ion Ratio Lower Upper
 228 100
 226 27.7 20.6 31.0
 229 19.3 15.8 23.8









#33
Chrysene
Concen: 1.629 ng
RT: 21.259 min Scan# 2004
Delta R.T. -0.011 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:228 Resp: 483594
Ion Ratio Lower Upper
228 100
226 30.3 22.6 34.0
229 19.2 16.0 24.0



#34
Bis(2-ethylhexyl)phthalate
Concen: 1.498 ng
RT: 21.164 min Scan# 1995
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35

Tgt Ion:149 Resp: 237951
Ion Ratio Lower Upper
149 100
167 28.0 22.1 33.1
279 4.3 3.4 5.0









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.471 min Scan# 2423
Delta R.T. -0.003 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:264	Resp:	89484
Ion Ratio	Lower	Upper
264	100	
260	24.3	19.4 29.2
265	23.2	18.6 27.8



#36
Indeno(1,2,3-cd)pyrene
Concen: 1.656 ng
RT: 25.740 min Scan# 3086
Delta R.T. -0.003 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35

Tgt Ion:276	Resp:	580977
Ion Ratio	Lower	Upper
276	100	
138	24.3	20.2 30.4
277	25.2	20.2 30.4









#37
Benzo(b)fluoranthene
Concen: 1.667 ng
RT: 22.797 min Scan# 2226
Delta R.T. -0.003 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 498673
Ion Ratio Lower Upper
252 100
253 22.0 17.7 26.5
125 9.6 7.8 11.8



#38
Benzo(k)fluoranthene
Concen: 1.659 ng
RT: 22.841 min Scan# 2239
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35

Tgt Ion:252 Resp: 490034
Ion Ratio Lower Upper
252 100
253 22.1 18.1 27.1
125 9.8 9.8 14.8#









#39
Benzo(a)pyrene
Concen: 1.679 ng
RT: 23.372 min Scan# 2394
Delta R.T. -0.003 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 457369
Ion Ratio Lower Upper
252 100
253 21.8 18.1 27.1
125 10.6 9.0 13.4



#40
Dibenzo(a,h)anthracene
Concen: 1.660 ng
RT: 25.761 min Scan# 3092
Delta R.T. -0.003 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35

Tgt Ion:278 Resp: 475346
Ion Ratio Lower Upper
278 100
139 16.0 13.8 20.6
279 23.4 19.0 28.6





#41
Benzo(g,h,i)perylene
Concen: 1.654 ng
RT: 26.435 min Scan# 3289
Delta R.T. -0.000 min
Lab File: BN016907.D
Acq: 12 Oct 2021 20:35

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:276 Resp: 498855
Ion Ratio Lower Upper
276 100
277 23.8 19.1 28.7
138 20.7 16.6 24.8

