

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101321\
 Data File : BN016937.D
 Acq On : 13 Oct 2021 21:05
 Operator : CG/JU
 Sample : PB139115BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB139115BS

Quant Time: Oct 14 02:10:47 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 13 02:19:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.688	152	20352	0.400	ng	0.00
7) Naphthalene-d8	10.445	136	89681	0.400	ng	#-0.01
13) Acenaphthene-d10	14.294	164	45840	0.400	ng	0.00
19) Phenanthrene-d10	17.030	188	88085	0.400	ng	#-0.01
29) Chrysene-d12	21.227	240	89575	0.400	ng	0.00
35) Perylene-d12	23.471	264	89113	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.291	112	15553	0.337	ng	0.00
5) Phenol-d6	6.868	99	17305	0.336	ng	0.00
8) Nitrobenzene-d5	8.822	82	17861	0.313	ng	0.00
11) 2-Methylnaphthalene-d10	12.038	152	41594	0.331	ng	0.00
14) 2,4,6-Tribromophenol	15.791	330	5909	0.284	ng	0.00
15) 2-Fluorobiphenyl	12.924	172	59827	0.345	ng	0.00
27) Fluoranthene-d10	19.073	212	74333	0.336	ng	0.00
31) Terphenyl-d14	19.678	244	65647	0.334	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.272	88	4582	0.337	ng	97
3) n-Nitrosodimethylamine	3.604	42	5342	0.342	ng	99
6) bis(2-Chloroethyl)ether	7.126	93	19180	0.348	ng	100
9) Naphthalene	10.495	128	78498	0.337	ng	99
10) Hexachlorobutadiene	10.787	225	15459	0.335	ng	# 100
12) 2-Methylnaphthalene	12.114	142	50381	0.341	ng	99
16) Acenaphthylene	14.015	152	60017	0.323	ng	100
17) Acenaphthene	14.358	154	43049	0.338	ng	99
18) Fluorene	15.347	166	51635	0.340	ng	100
20) 4,6-Dinitro-2-methylph...	15.436	198	1399	0.271	ng	# 92
21) 4-Bromophenyl-phenylether	16.239	248	16498	0.320	ng	97
22) Hexachlorobenzene	16.348	284	20045	0.335	ng	99
23) Atrazine	16.519	200	10310	0.324	ng	100
24) Pentachlorophenol	16.701	266	4765	0.359	ng	100
25) Phenanthrene	17.078	178	84446	0.337	ng	100
26) Anthracene	17.164	178	71357	0.333	ng	100
28) Fluoranthene	19.098	202	96724	0.357	ng	100
30) Pyrene	19.465	202	96222	0.338	ng	100
32) Benzo(a)anthracene	21.217	228	90006	0.334	ng	100
33) Chrysene	21.270	228	102828	0.353	ng	99
34) Bis(2-ethylhexyl)phtha...	21.164	149	31708	0.376	ng	99
36) Indeno(1,2,3-cd)pyrene	25.740	276	101084	0.300	ng	100
37) Benzo(b)fluoranthene	22.796	252	100303	0.351	ng	100
38) Benzo(k)fluoranthene	22.841	252	100397	0.354	ng	99
39) Benzo(a)pyrene	23.375	252	88507	0.339	ng	100
40) Dibenzo(a,h)anthracene	25.764	278	83190	0.304	ng	99
41) Benzo(g,h,i)perylene	26.435	276	81818	0.280	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101321\
Data File : BN016937.D
Acq On : 13 Oct 2021 21:05
Operator : CG/JU
Sample : PB139115BS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB139115BS

Quant Time: Oct 14 02:10:47 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 13 02:19:57 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.688 min Scan# 500
 Delta R.T. -0.001 min
 Lab File: BN016937.D
 Acq: 13 Oct 2021 21:05




Instrument :
 BNA_N
 ClientSampleId :
 PB139115BS


Tgt Ion:152 Resp: 20352
 Ion Ratio Lower Upper
 152 100
 150 154.1 122.6 183.8
 115 57.7 46.1 69.1



#2
 1,4-Dioxane
 Concen: 0.337 ng
 RT: 3.272 min Scan# 21
 Delta R.T. -0.009 min
 Lab File: BN016937.D
 Acq: 13 Oct 2021 21:05

Tgt Ion: 88 Resp: 4582
 Ion Ratio Lower Upper
 88 100
 43 29.6 24.3 36.5
 58 78.2 65.2 97.8









#3
n-Nitrosodimethylamine
Concen: 0.342 ng
RT: 3.604 min Scan# 57
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:	42	Resp:	5342
Ion	Ratio	Lower	Upper
42	100		
74	143.6	114.4	171.6
44	7.9	5.7	8.5



#4
2-Fluorophenol
Concen: 0.337 ng
RT: 5.291 min Scan# 240
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05

Tgt Ion:	112	Resp:	15553
Ion	Ratio	Lower	Upper
112	100		
64	56.6	45.0	67.4
63	28.6	22.6	33.8









#5
Phenol-d6
Concen: 0.336 ng
RT: 6.868 min Scan# 411
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion: 99 Resp: 17305
Ion Ratio Lower Upper
99 100
42 17.7 14.2 21.4
71 30.6 24.5 36.7



#6
bis(2-Chloroethyl)ether
Concen: 0.348 ng
RT: 7.126 min Scan# 439
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05

Tgt Ion: 93 Resp: 19180
Ion Ratio Lower Upper
93 100
63 75.0 60.3 90.5
95 32.7 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.445 min Scan# 736
Delta R.T. -0.012 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:	136	Resp:	89681
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.7	13.1
54	8.5	3.7	5.5#
68	6.8	3.2	4.8#



#8
Nitrobenzene-d5
Concen: 0.313 ng
RT: 8.822 min Scan# 608
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05

Tgt Ion:	82	Resp:	17861
Ion	Ratio	Lower	Upper
82	100		
128	37.7	28.3	42.5
54	52.0	43.0	64.4









#9
Naphthalene
Concen: 0.337 ng
RT: 10.495 min Scan# 740
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:	128	Resp:	78498
Ion Ratio	Lower	Upper	
128	100		
129	11.1	8.7	13.1
127	13.0	10.6	16.0



#10
Hexachlorobutadiene
Concen: 0.335 ng
RT: 10.787 min Scan# 763
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05

Tgt Ion:	225	Resp:	15459
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	64.2	51.1	76.7









#11
2-Methylnaphthalene-d10
Concen: 0.331 ng
RT: 12.038 min Scan# 939
Delta R.T. -0.005 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:152 Resp: 41594
Ion Ratio Lower Upper
152 100
151 20.5 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.341 ng
RT: 12.114 min Scan# 956
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05

Tgt Ion:142 Resp: 50381
Ion Ratio Lower Upper
142 100
141 87.5 70.5 105.7
115 32.4 26.0 39.0









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.294 min Scan# 1185
Delta R.T. 0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:164 Resp: 45840
Ion Ratio Lower Upper
164 100
162 101.2 81.4 122.0
160 45.1 37.0 55.4



#14
2,4,6-Tribromophenol
Concen: 0.284 ng
RT: 15.791 min Scan# 1303
Delta R.T. 0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05

Tgt Ion:330 Resp: 5909
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 30.4 24.4 36.6









#15
2-Fluorobiphenyl
Concen: 0.345 ng
RT: 12.924 min Scan# 1077
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:	172	Resp:	59827
Ion Ratio	Lower	Upper	
172	100		
171	33.9	27.4	41.0
170	22.0	18.1	27.1



#16
Acenaphthylene
Concen: 0.323 ng
RT: 14.015 min Scan# 1163
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05

Tgt Ion:	152	Resp:	60017
Ion Ratio	Lower	Upper	
152	100		
151	19.3	15.5	23.3
153	13.0	10.3	15.5









#17
Acenaphthene
Concen: 0.338 ng
RT: 14.358 min Scan# 1190
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:154 Resp: 43049
Ion Ratio Lower Upper
154 100
153 113.2 91.2 136.8
152 54.8 44.4 66.6



#18
Fluorene
Concen: 0.340 ng
RT: 15.347 min Scan# 1268
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05

Tgt Ion:166 Resp: 51635
Ion Ratio Lower Upper
166 100
165 97.3 78.1 117.1
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.030 min Scan# 1404
Delta R.T. -0.012 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:188 Resp: 88085
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 12.4 4.9 7.3#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.271 ng
RT: 15.436 min Scan# 1275
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05

Tgt Ion:198 Resp: 1399
Ion Ratio Lower Upper
198 100
182 17.1 11.0 16.6#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.320 ng
RT: 16.239 min Scan# 1339
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion	Ratio	Lower	Upper
248	100		
250	99.0	78.2	117.2
141	49.4	43.1	64.7



#22
Hexachlorobenzene
Concen: 0.335 ng
RT: 16.348 min Scan# 1348
Delta R.T. 0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05

Tgt Ion	Ratio	Lower	Upper
284	100		
142	31.2	25.1	37.7
249	28.6	23.2	34.8









#23
Atrazine
Concen: 0.324 ng
RT: 16.519 min Scan# 1362
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:200 Resp: 10310
Ion Ratio Lower Upper
200 100
173 22.7 18.0 27.0
215 50.0 40.0 60.0



#24
Pentachlorophenol
Concen: 0.359 ng
RT: 16.701 min Scan# 1377
Delta R.T. 0.012 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05

Tgt Ion:266 Resp: 4765
Ion Ratio Lower Upper
266 100
264 62.9 50.5 75.7
268 64.3 51.2 76.8









#25
Phenanthrene
Concen: 0.337 ng
RT: 17.078 min Scan# 1408
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:178 Resp: 84446
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.4
179 15.6 12.3 18.5



#26
Anthracene
Concen: 0.333 ng
RT: 17.164 min Scan# 1415
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05

Tgt Ion:178 Resp: 71357
Ion Ratio Lower Upper
178 100
176 18.2 14.6 21.8
179 15.4 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.336 ng
RT: 19.073 min Scan# 1698
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:212 Resp: 74333
Ion Ratio Lower Upper
212 100
106 13.5 10.7 16.1
104 7.4 5.8 8.8



#28
Fluoranthene
Concen: 0.357 ng
RT: 19.098 min Scan# 1703
Delta R.T. -0.005 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05

Tgt Ion:202 Resp: 96724
Ion Ratio Lower Upper
202 100
101 11.3 9.0 13.4
203 17.3 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.227 min Scan# 2001
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:240 Resp: 89575
Ion Ratio Lower Upper
240 100
120 10.0 8.6 12.8
236 26.0 21.0 31.4



#30
Pyrene
Concen: 0.338 ng
RT: 19.465 min Scan# 1777
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05

Tgt Ion:202 Resp: 96222
Ion Ratio Lower Upper
202 100
200 20.5 16.5 24.7
203 17.5 14.0 21.0









#31
 Terphenyl-d14
 Concen: 0.334 ng
 RT: 19.678 min Scan# 1820
 Delta R.T. -0.000 min
 Lab File: BN016937.D
 Acq: 13 Oct 2021 21:05




Instrument :
 BNA_N
 ClientSampleId :
 PB139115BS


Tgt Ion:244	Resp:	65647
Ion Ratio	Lower	Upper
244	100	
212	7.1	5.7 8.5
122	12.2	9.8 14.6



#32
 Benzo(a)anthracene
 Concen: 0.334 ng
 RT: 21.217 min Scan# 2000
 Delta R.T. -0.000 min
 Lab File: BN016937.D
 Acq: 13 Oct 2021 21:05

Tgt Ion:228	Resp:	90006
Ion Ratio	Lower	Upper
228	100	
226	25.9	20.6 31.0
229	20.0	15.8 23.8









#33
Chrysene
Concen: 0.353 ng
RT: 21.270 min Scan# 2005
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:228	Resp: 102828
Ion Ratio	Lower Upper
228 100	
226 27.9	22.6 34.0
229 20.2	16.0 24.0



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.376 ng
RT: 21.164 min Scan# 1995
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05

Tgt Ion:149	Resp: 31708
Ion Ratio	Lower Upper
149 100	
167 27.2	22.1 33.1
279 4.2	3.4 5.0









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.471 min Scan# 2423
Delta R.T. -0.004 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05

Instrument :
BNA_N
ClientSampleId :
PB139115BS

Tgt Ion:264 Resp: 89113




Ion	Ratio	Lower	Upper
264	100		
260	24.6	19.4	29.2
265	22.8	18.6	27.8




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.300 ng
RT: 25.740 min Scan# 3086
Delta R.T. -0.004 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05

Tgt Ion:276 Resp: 101084


Ion	Ratio	Lower	Upper
276	100		
138	25.2	20.2	30.4
277	25.4	20.2	30.4







#37
Benzo(b)fluoranthene
Concen: 0.351 ng
RT: 22.796 min Scan# 2226
Delta R.T. -0.004 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05


Instrument :
BNA_N
ClientSampleId :
PB139115BS





Tgt Ion:252 Resp: 100303
Ion Ratio Lower Upper
252 100
253 22.0 17.7 26.5
125 10.1 7.8 11.8




#38
Benzo(k)fluoranthene
Concen: 0.354 ng
RT: 22.841 min Scan# 2239
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05



Tgt Ion:252 Resp: 100397
Ion Ratio Lower Upper
252 100
253 22.3 18.1 27.1
125 11.3 9.8 14.8









#39
Benzo(a)pyrene
Concen: 0.339 ng
RT: 23.375 min Scan# 2395
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:252 Resp: 88507
Ion Ratio Lower Upper
252 100
253 22.4 18.1 27.1
125 11.2 9.0 13.4



#40
Dibenzo(a,h)anthracene
Concen: 0.304 ng
RT: 25.764 min Scan# 3093
Delta R.T. -0.000 min
Lab File: BN016937.D
Acq: 13 Oct 2021 21:05

Tgt Ion:278 Resp: 83190
Ion Ratio Lower Upper
278 100
139 16.7 13.8 20.6
279 24.3 19.0 28.6





#41
 Benzo(g,h,i)perylene
 Concen: 0.280 ng
 RT: 26.435 min Scan# 3289
 Delta R.T. -0.000 min
 Lab File: BN016937.D
 Acq: 13 Oct 2021 21:05

Instrument :
 BNA_N
 ClientSampleId :
 PB139115BS

Tgt Ion:	276	Resp:	81818
Ion Ratio	Lower	Upper	
276	100		
277	23.8	19.1	28.7
138	20.8	16.6	24.8

