

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101321\
 Data File : BN016912.D
 Acq On : 13 Oct 2021 00:12
 Operator : CG/JU
 Sample : PB139838BL
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB139838BL

Quant Time: Oct 13 06:19:20 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 13 02:19:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.688	152	20818	0.40	ng	0.00
7) Naphthalene-d8	10.457	136	90857	0.40	ng	0.00
13) Acenaphthene-d10	14.294	164	45865	0.40	ng	0.00
19) Phenanthrene-d10	17.042	188	82315	0.40	ng	0.00
29) Chrysene-d12	21.227	240	70535	0.40	ng	0.00
35) Perylene-d12	23.471	264	74183	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.328	112	17340	0.37	ng	0.04
5) Phenol-d6	6.886	99	14806	0.28	ng	0.02
8) Nitrobenzene-d5	8.835	82	19147	0.33	ng	0.01
11) 2-Methylnaphthalene-d10	12.043	152	50524	0.40	ng	0.00
14) 2,4,6-Tribromophenol	15.791	330	3668	0.18	ng	0.00
15) 2-Fluorobiphenyl	12.924	172	64327	0.37	ng	0.00
27) Fluoranthene-d10	19.073	212	81206	0.39	ng	0.00
31) Terphenyl-d14	19.678	244	66893	0.43	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101321\
Data File : BN016912.D
Acq On : 13 Oct 2021 00:12
Operator : CG/JU
Sample : PB139838BL
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB139838BL

Quant Time: Oct 13 06:19:20 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 13 02:19:57 2021
Response via : Initial Calibration

