

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101321\
 Data File : BN016941.D
 Acq On : 13 Oct 2021 23:30
 Operator : CG/JU
 Sample : M4170-03
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SW-RR-03-(SA)

Manual Integrations
 APPROVED

mohammad
 10/14/2021 10:52:53 AM

Quant Time: Oct 14 02:11:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 13 02:19:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.689	152	20081	0.40	ng	0.00
7) Naphthalene-d8	10.445	136	89631	0.40	ng	#-0.01
13) Acenaphthene-d10	14.294	164	47587	0.40	ng	0.00
19) Phenanthrene-d10	17.030	188	92616	0.40	ng	#-0.01
29) Chrysene-d12	21.228	240	92255	0.40	ng	0.00
35) Perylene-d12	23.471	264	91086	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.310	112	6130	0.13	ng	0.02
5) Phenol-d6	6.877	99	4011	0.08	ng	0.00
8) Nitrobenzene-d5	8.823	82	14548	0.26	ng	0.00
11) 2-Methylnaphthalene-d10	12.039	152	39777	0.32	ng	0.00
14) 2,4,6-Tribromophenol	15.791	330	8593	0.40	ng	0.00
15) 2-Fluorobiphenyl	12.924	172	54333	0.30	ng	0.00
27) Fluoranthene-d10	19.068	212	107192	0.46	ng	0.00
31) Terphenyl-d14	19.679	244	91373	0.45	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.290	88	413m	0.03	ng	
9) Naphthalene	10.496	128	47972	0.21	ng	99
24) Pentachlorophenol	16.689	266	774	0.22	ng	98
34) Bis(2-ethylhexyl)phtha...	21.164	149	22883	0.32	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101321\
Data File : BN016941.D
Acq On : 13 Oct 2021 23:30
Operator : CG/JU
Sample : M4170-03
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SW-RR-03-(SA)

Manual Integrations
APPROVED

mohammad
10/14/2021 10:52:53 AM

Quant Time: Oct 14 02:11:29 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 13 02:19:57 2021
Response via : Initial Calibration

