




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101321\
Data File : BN016920.D
Acq On : 13 Oct 2021 05:36
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 13 06:46:27 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 13 06:46:19 2021
Response via : Initial Calibration



TIC: BN016920.D\data.ms

(70) Pentachlorophenol (C)




16.622min (0.000) 30681.90 ng

response 226096

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	62.61
264.00	61.60	61.11
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101321\
Data File : BN016920.D
Acq On : 13 Oct 2021 05:36
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 13 06:46:27 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 13 06:46:19 2021
Response via : Initial Calibration



TIC: BN016920.D\data.ms

(77) Benzidine

19.245min (0.000) 0.00 ng

response 1443994

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.79
183.00	13.20	11.93
0.00	0.00	0.00