

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101321\
 Data File : BN016932.D
 Acq On : 13 Oct 2021 16:13
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 13 16:47:22 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 13 02:19:57 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	88	0.00
2	1,4-Dioxane	0.400	0.396	1.0	85	0.00
3	n-Nitrosodimethylamine	0.400	0.418	-4.5	94	0.00
4 S	2-Fluorophenol	0.400	0.398	0.5	90	0.00
5 S	Phenol-d6	0.400	0.395	1.3	91	0.00
6	bis(2-Chloroethyl)ether	0.400	0.403	-0.8	88	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	86	-0.01
8 S	Nitrobenzene-d5	0.400	0.376	6.0	90	0.00
9	Naphthalene	0.400	0.397	0.8	86	0.00
10	Hexachlorobutadiene	0.400	0.400	0.0	87	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.392	2.0	86	0.00
12	2-Methylnaphthalene	0.400	0.402	-0.5	87	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	86	0.00
14 S	2,4,6-Tribromophenol	0.400	0.357	10.8	93	0.00
15 S	2-Fluorobiphenyl	0.400	0.403	-0.8	85	0.00
16	Acenaphthylene	0.400	0.379	5.3	85	-0.01
17	Acenaphthene	0.400	0.398	0.5	86	0.00
18	Fluorene	0.400	0.400	0.0	86	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	86	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.305	23.8	53	0.00
21	4-Bromophenyl-phenylether	0.400	0.388	3.0	87	0.00
22	Hexachlorobenzene	0.400	0.398	0.5	86	0.00
23	Atrazine	0.400	0.376	6.0	89	0.00
24	Pentachlorophenol	0.400	0.392	2.0	78	0.00
25	Phenanthrene	0.400	0.397	0.8	86	0.00
26	Anthracene	0.400	0.394	1.5	89	0.00
27 SURR	Fluoranthene-d10	0.400	0.400	0.0	89	0.00
28	Fluoranthene	0.400	0.425	-6.2	90	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	89	0.00
30	Pyrene	0.400	0.391	2.3	89	0.00
31 S	Terphenyl-d14	0.400	0.395	1.3	90	0.00
32	Benzo(a)anthracene	0.400	0.388	3.0	91	0.00
33	Chrysene	0.400	0.417	-4.2	92	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.404	-1.0	90	0.00
35 I	Perylene-d12	0.400	0.400	0.0	79	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.341	14.8	67	0.00
37	Benzo(b)fluoranthene	0.400	0.421	-5.2	86	0.00
38	Benzo(k)fluoranthene	0.400	0.426	-6.5	83	0.00
39 C	Benzo(a)pyrene	0.400	0.400	0.0	82	0.00
40	Dibenzo(a,h)anthracene	0.400	0.348	13.0	68	0.00
41	Benzo(g,h,i)perylene	0.400	0.332	17.0	66	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4