

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101321\
 Data File : BN016903.D
 Acq On : 12 Oct 2021 18:09
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Manual Integrations
 APPROVED

mohammad
 10/14/2021 10:51:57 AM

Quant Time: Oct 13 01:52:52 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 13 01:50:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.688	152	19240	0.40	ng	0.00
7) Naphthalene-d8	10.457	136	84963	0.40	ng	0.00
13) Acenaphthene-d10	14.294	164	44284	0.40	ng	0.00
19) Phenanthrene-d10	17.042	188	82846	0.40	ng	0.00
29) Chrysene-d12	21.227	240	74484	0.40	ng	# 0.00
35) Perylene-d12	23.477	264	78493	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.291	112	4103	0.08	ng	0.00
5) Phenol-d6	6.868	99	4422	0.08	ng	0.00
8) Nitrobenzene-d5	8.835	82	4522	0.08	ng	0.01
11) 2-Methylnaphthalene-d10	12.043	152	11573	0.09	ng	0.00
14) 2,4,6-Tribromophenol	0.000	330	0d	0.00	ng	
15) 2-Fluorobiphenyl	12.924	172	16660	0.10	ng	0.00
27) Fluoranthene-d10	19.073	212	18618	0.09	ng	0.00
31) Terphenyl-d14	19.683	244	15974	0.09	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.281	88	1218	0.08	ng	95
3) n-Nitrosodimethylamine	3.613	42	1252	0.08	ng	# 96
6) bis(2-Chloroethyl)ether	7.126	93	5214	0.10	ng	99
9) Naphthalene	10.495	128	22782	0.10	ng	99
10) Hexachlorobutadiene	10.787	225	4498	0.10	ng	# 100
12) 2-Methylnaphthalene	12.118	142	14131	0.10	ng	99
16) Acenaphthylene	14.015	152	14804	0.08	ng	100
17) Acenaphthene	14.357	154	12136	0.10	ng	99
18) Fluorene	15.347	166	13763	0.09	ng	100
20) 4,6-Dinitro-2-methylph...	15.436	198	432	0.04	ng	# 57
21) 4-Bromophenyl-phenylether	16.238	248	4330	0.09	ng	# 91
22) Hexachlorobenzene	16.348	284	5529	0.10	ng	99
23) Atrazine	16.518	200	2351	0.06	ng	96
25) Phenanthrene	17.078	178	23212	0.10	ng	99
26) Anthracene	17.176	178	16880	0.08	ng	100
28) Fluoranthene	19.102	202	23198	0.09	ng	99
30) Pyrene	19.465	202	23372	0.10	ng	100
32) Benzo(a)anthracene	21.217	228	20317	0.09	ng	99
33) Chrysene	21.270	228	24374	0.10	ng	99
36) Indeno(1,2,3-cd)pyrene	25.750	276	27413	0.09	ng	99
37) Benzo(b)fluoranthene	22.800	252	23234	0.09	ng	97
38) Benzo(k)fluoranthene	22.848	252	23706	0.09	ng	98
39) Benzo(a)pyrene	23.378	252	21480m	0.09	ng	
40) Dibenzo(a,h)anthracene	25.767	278	21633	0.09	ng	98
41) Benzo(g,h,i)perylene	26.442	276	24909	0.09	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101321\
Data File : BN016903.D
Acq On : 12 Oct 2021 18:09
Operator : CG/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Manual Integrations
APPROVED


mohammad
10/14/2021 10:51:57 AM

Quant Time: Oct 13 01:52:52 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 13 01:50:22 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.688 min Scan# 500
Delta R.T. -0.001 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09





Tgt Ion:152 Resp: 19240
Ion Ratio Lower Upper
152 100
150 153.7 122.6 183.8
115 58.1 46.1 69.1



#2
1,4-Dioxane
Concen: 0.08 ng
RT: 3.281 min Scan# 22
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09




Tgt Ion: 88 Resp: 1218
Ion Ratio Lower Upper
88 100
43 36.1 24.3 36.5
58 84.5 65.2 97.8



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED

mohammad
10/14/2021 10:51:57 AM







#3
n-Nitrosodimethylamine
Concen: 0.08 ng
RT: 3.613 min Scan# 58
Delta R.T. 0.009 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09

Tgt Ion: 42 Resp: 1252
Ion Ratio Lower Upper
42 100
74 147.1 114.4 171.6
44 3.7 5.7 8.5#

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
10/14/2021 10:51:57 AM



#4
2-Fluorophenol
Concen: 0.08 ng
RT: 5.291 min Scan# 240
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09

Tgt Ion: 112 Resp: 4103
Ion Ratio Lower Upper
112 100
64 55.6 45.0 67.4
63 28.3 22.6 33.8









#5
Phenol-d6
Concen: 0.08 ng
RT: 6.868 min Scan# 411
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09

Tgt Ion:	99	Resp:	4422
Ion	Ratio	Lower	Upper
99	100		
42	17.5	14.2	21.4
71	30.5	24.5	36.7

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1




Manual Integrations
APPROVED


mohammad
10/14/2021 10:51:57 AM



#6
bis(2-Chloroethyl)ether
Concen: 0.10 ng
RT: 7.126 min Scan# 439
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09


Tgt Ion:	93	Resp:	5214
Ion	Ratio	Lower	Upper
93	100		
63	76.4	60.3	90.5
95	33.8	26.2	39.4





#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.457 min Scan# 737
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:	136	Resp:	84963
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.7	13.1
54	4.8	3.7	5.5
68	4.1	3.2	4.8

Manual Integrations
APPROVED



mohammad
10/14/2021 10:51:57 AM




#8
Nitrobenzene-d5
Concen: 0.08 ng
RT: 8.835 min Scan# 609
Delta R.T. 0.012 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09



Tgt Ion:	82	Resp:	4522
Ion	Ratio	Lower	Upper
82	100		
128	45.9	28.3	42.5#
54	47.2	43.0	64.4









#9
Naphthalene
Concen: 0.10 ng
RT: 10.495 min Scan# 740
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09

Tgt Ion:	128	Resp:	22782
Ion Ratio	Lower	Upper	
128	100		
129	11.3	8.7	13.1
127	13.3	10.6	16.0

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
10/14/2021 10:51:57 AM



#10
Hexachlorobutadiene
Concen: 0.10 ng
RT: 10.787 min Scan# 763
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09

Tgt Ion:	225	Resp:	4498
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	64.1	51.1	76.7









#11
2-Methylnaphthalene-d10
Concen: 0.09 ng
RT: 12.043 min Scan# 940
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09

Tgt Ion:152 Resp: 11573
Ion Ratio Lower Upper
152 100
151 20.5 16.2 24.4

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1




Manual Integrations
APPROVED


mohammad
10/14/2021 10:51:57 AM



#12
2-Methylnaphthalene
Concen: 0.10 ng
RT: 12.118 min Scan# 957
Delta R.T. 0.004 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09


Tgt Ion:142 Resp: 14131
Ion Ratio Lower Upper
142 100
141 87.3 70.5 105.7
115 33.0 26.0 39.0





#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.294 min Scan# 1185
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Tgt Ion:164 Resp: 44284
Ion Ratio Lower Upper
164 100
162 102.2 81.4 122.0
160 46.5 37.0 55.4

Manual Integrations
APPROVED


mohammad
10/14/2021 10:51:57 AM




#14
2,4,6-Tribromophenol
Concen: 0.00 ng
Expected RT: 15.79 min

Lab File: BN016903.D
Acq: 12 Oct 2021 18:09


Tgt Ion: 330
Sig Exp Ratio
330 100
332 0.0
141 30.5





#15
2-Fluorobiphenyl
Concen: 0.10 ng
RT: 12.924 min Scan# 1077
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.3	27.4	41.0
170	23.4	18.1	27.1

Manual Integrations
APPROVED



mohammad
10/14/2021 10:51:57 AM




#16
Acenaphthylene
Concen: 0.08 ng
RT: 14.015 min Scan# 1163
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09




Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.4	15.5	23.3
153	13.0	10.3	15.5





#17
Acenaphthene
Concen: 0.10 ng
RT: 14.357 min Scan# 1190
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1




Tgt Ion:154 Resp: 12136
Ion Ratio Lower Upper
154 100
153 113.3 91.2 136.8
152 55.8 44.4 66.6

Manual Integrations
APPROVED



mohammad
10/14/2021 10:51:57 AM




#18
Fluorene
Concen: 0.09 ng
RT: 15.347 min Scan# 1268
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09



Tgt Ion:166 Resp: 13763
Ion Ratio Lower Upper
166 100
165 98.0 78.1 117.1
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.042 min Scan# 1405
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09

Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	6.7	4.9	7.3

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
10/14/2021 10:51:57 AM



#20
4,6-Dinitro-2-methylphenol
Concen: 0.04 ng
RT: 15.436 min Scan# 1275
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09

Tgt Ion	Ratio	Lower	Upper
198	100		
182	31.4	11.0	16.6
77	0.0	0.0	0.0









#21
4-Bromophenyl-phenylether
Concen: 0.09 ng
RT: 16.238 min Scan# 1339
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09

Tgt Ion:248 Resp: 4330
Ion Ratio Lower Upper
248 100
250 93.7 78.2 117.2
141 67.5 43.1 64.7#

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
10/14/2021 10:51:57 AM




#22
Hexachlorobenzene
Concen: 0.10 ng
RT: 16.348 min Scan# 1348
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09

Tgt Ion:284 Resp: 5529
Ion Ratio Lower Upper
284 100
142 31.8 25.1 37.7
249 29.4 23.2 34.8









#23
Atrazine
Concen: 0.06 ng
RT: 16.518 min Scan# 1362
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09





Tgt Ion:200 Resp: 2351
Ion Ratio Lower Upper
200 100
173 26.8 18.0 27.0
215 51.7 40.0 60.0



#25
Phenanthrene
Concen: 0.10 ng
RT: 17.078 min Scan# 1408
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09




Tgt Ion:178 Resp: 23212
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.4
179 15.9 12.3 18.5




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED

mohammad
10/14/2021 10:51:57 AM




#26
 Anthracene
 Concen: 0.08 ng
 RT: 17.176 min Scan# 1416
 Delta R.T. 0.012 min
 Lab File: BN016903.D
 Acq: 12 Oct 2021 18:09



Tgt Ion:178 Resp: 16880


Ion	Ratio	Lower	Upper
178	100		
176	18.2	14.6	21.8
179	15.1	12.2	18.4




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Manual Integrations
 APPROVED

mohammad
 10/14/2021 10:51:57 AM





#27
 Fluoranthene-d10
 Concen: 0.09 ng
 RT: 19.073 min Scan# 1698
 Delta R.T. -0.000 min
 Lab File: BN016903.D
 Acq: 12 Oct 2021 18:09




Tgt Ion:212 Resp: 18618

Ion	Ratio	Lower	Upper
212	100		
106	13.3	10.7	16.1
104	7.3	5.8	8.8









#28
Fluoranthene
Concen: 0.09 ng
RT: 19.102 min Scan# 1704
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09





Tgt Ion:202 Resp: 23198
Ion Ratio Lower Upper
202 100
101 11.3 9.0 13.4
203 17.0 13.8 20.8



#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.227 min Scan# 2001
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09




Tgt Ion:240 Resp: 74484
Ion Ratio Lower Upper
240 100
120 14.2 8.6 12.8#
236 26.9 21.0 31.4



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED

mohammad
10/14/2021 10:51:57 AM







#30
Pyrene
Concen: 0.10 ng
RT: 19.465 min Scan# 1777
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09

Tgt Ion:	202	Resp:	23372
Ion Ratio	Lower	Upper	
202	100		
200	20.7	16.5	24.7
203	17.4	14.0	21.0

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1




Manual Integrations
APPROVED


mohammad
10/14/2021 10:51:57 AM




#31
Terphenyl-d14
Concen: 0.09 ng
RT: 19.683 min Scan# 1821
Delta R.T. 0.005 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09

Tgt Ion:	244	Resp:	15974
Ion Ratio	Lower	Upper	
244	100		
212	7.4	5.7	8.5
122	12.0	9.8	14.6









#32
Benzo(a)anthracene
Concen: 0.09 ng
RT: 21.217 min Scan# 2000
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09





Tgt Ion:228 Resp: 20317
Ion Ratio Lower Upper
228 100
226 26.6 20.6 31.0
229 19.8 15.8 23.8



#33
Chrysene
Concen: 0.10 ng
RT: 21.270 min Scan# 2005
Delta R.T. -0.000 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09




Tgt Ion:228 Resp: 24374
Ion Ratio Lower Upper
228 100
226 29.3 22.6 34.0
229 19.8 16.0 24.0



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Manual Integrations
APPROVED

mohammad
10/14/2021 10:51:57 AM



#35
Perylene-d12
Concen: 0.40 ng
RT: 23.477 min Scan# 2425
Delta R.T. 0.003 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:264 Resp: 78493
Ion Ratio Lower Upper
264 100
260 24.8 19.4 29.2
265 23.8 18.6 27.8

Manual Integrations
APPROVED



mohammad
10/14/2021 10:51:57 AM




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.09 ng
RT: 25.750 min Scan# 3089
Delta R.T. 0.007 min
Lab File: BN016903.D
Acq: 12 Oct 2021 18:09



Tgt Ion:276 Resp: 27413
Ion Ratio Lower Upper
276 100
138 24.8 20.2 30.4
277 25.4 20.2 30.4





#37

Benzo(b)fluoranthene

Concen: 0.09 ng

RT: 22.800 min Scan# 2227

Delta R.T. -0.000 min

Lab File: BN016903.D

Acq: 12 Oct 2021 18:09

Tgt Ion:	252	Resp:	23234
Ion Ratio	Lower	Upper	
252	100		
253	23.3	17.7	26.5
125	11.1	7.8	11.8

Instrument :





BNA_N

ClientSampleId :

SSTDIC0.1

Manual Integrations
APPROVED

mohammad
10/14/2021 10:51:57 AM



#38

Benzo(k)fluoranthene

Concen: 0.09 ng




RT: 22.848 min Scan# 2241


Delta R.T. 0.007 min

Lab File: BN016903.D

Acq: 12 Oct 2021 18:09

Tgt Ion:	252	Resp:	23706
Ion Ratio	Lower	Upper	
252	100		
253	24.0	18.1	27.1
125	12.0	9.8	14.8





#39

Benzo(a)pyrene

Concen: 0.09 ng m

RT: 23.378 min Scan# 2396

Delta R.T. 0.003 min

Lab File: BN016903.D

Acq: 12 Oct 2021 18:09

Tgt Ion:	252	Resp:	21480
Ion Ratio	Lower	Upper	
252	100		
253	24.6	18.1	27.1
125	13.1	9.0	13.4

Instrument :





BNA_N

ClientSampleId :

SSTDICC0.1

Manual Integrations
APPROVED

mohammad
10/14/2021 10:51:57 AM



#40

Dibenzo(a,h)anthracene

Concen: 0.09 ng




RT: 25.767 min Scan# 3094


Delta R.T. 0.003 min

Lab File: BN016903.D


Acq: 12 Oct 2021 18:09

Tgt Ion:	278	Resp:	21633
Ion Ratio	Lower	Upper	
278	100		
139	17.6	13.8	20.6
279	25.0	19.0	28.6







#41
 Benzo(g,h,i)perylene
 Concen: 0.09 ng
 RT: 26.442 min Scan# 3291
 Delta R.T. 0.007 min
 Lab File: BN016903.D
 Acq: 12 Oct 2021 18:09



Tgt Ion:	276	Resp:	24909
Ion Ratio	Lower	Upper	
276	100		
277	24.3	19.1	28.7
138	21.5	16.6	24.8



Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Manual Integrations
 APPROVED

mohammad
 10/14/2021 10:51:57 AM