

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN101419\
 Data File : BN008157.D
 Acq On : 15 Oct 2019 05:21
 Operator : JU
 Sample : K5097-08
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Oct 15 06:51:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN101019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 15 05:17:53 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	2089	0.40	ng/ul	0.00
2) Naphthalene-d8	10.38	136	10654	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.25	164	7110	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.10	188	910187	0.40	ng/ul	0.09
16) Chrysene-d12	0.00	240	0	0.00	ng/ul	-21.21
20) Perylene-d12	23.38	264	17658	0.40	ng/ul	-0.02
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.98	152	3977	0.22	ng/ul	0.00
14) Fluoranthene-d10	0.00	212	0	0.00	ng/ul	
Target Compounds						
3) Naphthalene	10.43	128	1194	0.039	ng/ul#	82
5) 2-Methylnaphthalene	12.06	142	1298	0.062	ng/ul	99
23) Benzo(a)pyrene	23.24	252	3231	0.053	ng/ul#	56


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN101419\
Data File : BN008157.D
Acq On : 15 Oct 2019 05:21
Operator : JU
Sample : K5097-08
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Quant Time: Oct 15 06:51:20 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN101019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 15 05:17:53 2019
Response via : Initial Calibration





#3

Naphthalene

Concen: 0.039 ng/ul

RT: 10.43 min Scan# 2329

Delta R.T. -0.01 min

Lab File: BN008157.D

Acq: 15 Oct 2019 05:21

Instrument :

BNA_N

ClientSampleId :


Tot Ion:128 Resp: 1194

Ion Ratio Lower Upper

128 100

129 21.5 9.8 14.8#




127 19.1 11.1 16.7#



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#5

2-Methylnaphthalene

Concen: 0.062 ng/ul

RT: 12.06 min Scan# 2625

Delta R.T. -0.01 min

Lab File: BN008157.D


Acq: 15 Oct 2019 05:21

Tgt Ion:142 Resp: 1298

Ion Ratio Lower Upper



142 100


141 90.6 71.4 107.2



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E





#23

Benzo(a)pyrene

Concen: 0.053 ng/ul

RT: 23.24 min Scan# 5390

Delta R.T. -0.06 min


Lab File: BN008157.D

Acq: 15 Oct 2019 05:21

Instrument :

BNA_N

ClientSampleId :




Tot Ion: 252 Resp: 3231

Ion Ratio Lower Upper

252 100

253 48.6 22.8 34.2#

125 47.5 18.4 27.6#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E

