


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101419\
 Data File : BN008133.D
 Acq On : 14 Oct 2019 14:26
 Operator : JU
 Sample : DFTPP03
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP03

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN101019.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Oct 14 17:19:29 2019



AutoFind: Scans 2460, 2461, 2462; Background Corrected with Scan 2454

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.0	64082	PASS
68	69	0.00	2	1.5	1207	PASS
69	198	0.00	100	45.0	82293	PASS
70	69	0.00	2	0.5	447	PASS
127	198	10	80	45.9	83973	PASS
197	198	0.00	2	0.3	532	PASS
198	198	100	100	100.0	182997	PASS
199	198	5	9	6.9	12572	PASS
275	198	10	60	27.1	49602	PASS
365	198	1	100	3.3	6084	PASS
441	198	0.01	100	10.7	19661	PASS
442	198	50	100	66.9	122384	PASS
443	442	15	24	18.9	23184	PASS

m/z	Abundance
40.00	1390.0
41.00	347.0
43.00	392.0
43.90	4576.0
49.00	324.0
73.00	920.0
76.90	383.0
83.80	304.0
96.00	413.0
132.90	362.0
147.00	374.0
191.00	426.0
193.10	357.0
207.00	3108.0
208.00	590.0
208.90	466.0
281.00	893.0

Instrument :

BNA_N

ClientSampleId :

DFTPP03

m/z	Abundance
36.90	371.0
38.00	883.0
39.00	4836.0
40.00	1443.0
41.00	420.0
43.10	333.0
44.00	4324.0
49.00	923.0
50.00	14916.0
51.00	59048.0
52.00	3672.0
55.00	734.0
56.00	2064.0
57.00	4029.0
61.00	994.0
61.90	905.0
63.00	2645.0
64.00	435.0
65.10	1347.0
68.00	1639.0
69.00	74096.0
69.90	445.0
73.00	1743.0
74.00	7445.0
75.00	11959.0
76.00	4212.0
77.00	80016.0
78.00	6061.0
79.00	5339.0
80.00	4039.0
81.00	5712.0
82.00	1168.0
83.00	1599.0
84.00	555.0
84.90	1231.0
85.90	1679.0
87.00	831.0
88.00	396.0
91.00	1259.0
92.00	1778.0
93.00	9065.0
94.00	700.0
95.00	429.0
96.00	522.0
98.00	6798.0
99.00	6750.0
99.90	721.0
100.90	2703.0
103.00	1168.0
104.00	1839.0
105.00	1838.0
106.20	602.0
107.00	21208.0
108.00	3475.0
108.90	862.0
110.00	40528.0
111.00	6622.0
112.00	729.0
116.10	1185.0
117.00	17296.0
118.00	1720.0
119.00	306.0
122.00	1595.0
123.00	2170.0
124.00	978.0
125.00	985.0
127.00	71216.0
128.00	5650.0
129.00	27680.0
130.00	1977.0
130.90	787.0
132.90	607.0
134.00	648.0
135.00	2527.0
135.90	672.0
137.00	1726.0
140.00	431.0
141.00	3444.0
141.90	1265.0
142.90	740.0
146.00	724.0
146.90	1904.0
148.00	4126.0
149.00	1318.0
149.90	374.0
150.90	488.0
151.70	333.0
152.20	355.0
152.90	1094.0
153.90	910.0
155.00	2169.0
156.00	2915.0

Instrument :
BNA_N
ClientSampleId :
DFTPP03

156.90	633.0
157.90	860.0
158.90	467.0
160.00	1145.0
161.00	1696.0
161.90	535.0
165.00	1477.0
165.90	1502.0
167.00	7209.0
168.00	5948.0
168.90	945.0
170.20	311.0
170.80	407.0
171.80	516.0
172.90	775.0
174.00	1143.0
175.00	2398.0
175.90	566.0
176.90	1341.0
178.00	438.0
179.00	5667.0
180.00	4140.0
181.00	1743.0
182.80	328.0
183.80	429.0
185.00	2561.0
186.00	19800.0
187.00	5748.0
188.10	908.0
188.90	1330.0
190.80	739.0
191.90	1734.0
193.00	2391.0
193.90	530.0
196.00	3734.0
197.90	141952.0
198.90	9651.0
199.90	662.0
201.70	555.0
203.00	1082.0
204.00	5014.0
205.00	8635.0
206.00	38640.0
207.00	7995.0
208.00	1754.0
209.00	984.0
210.00	615.0
211.00	1666.0
215.00	578.0
216.00	931.0
217.00	9983.0
217.90	1486.0
221.00	6203.0
221.90	1248.0
223.00	2278.0
224.00	20064.0
225.00	5097.0
226.00	1112.0
227.00	8392.0
228.00	1152.0
229.00	1752.0
231.10	744.0
232.90	319.0
234.10	491.0
235.00	552.0
236.00	452.0
237.00	1532.0
238.90	395.0
239.90	360.0
241.00	514.0
242.00	1044.0
242.90	1141.0
244.00	14788.0
245.00	2314.0
246.00	3083.0
246.90	788.0
248.80	430.0
252.90	430.0
255.00	81000.0
256.00	11545.0
257.00	1255.0
257.90	4320.0
258.80	553.0
265.00	2110.0
273.00	2597.0
274.00	5536.0
275.00	34784.0
276.00	4679.0
277.00	2791.0
277.90	653.0
280.90	851.0
282.00	322.0
283.00	641.0
285.10	478.0
293.00	789.0
296.00	8814.0

Instrument :
BNA_N
ClientSampleId :
DFTPP03

297.00	1536.0
301.80	319.0
303.00	1004.0
303.90	337.0
314.10	437.0
314.90	1010.0
315.90	586.0
323.10	2661.0
324.00	580.0
326.90	679.0
334.00	1580.0
335.00	522.0
340.80	317.0
341.10	329.0
346.00	870.0
351.90	722.0
353.00	680.0
353.90	693.0
355.00	361.0
364.90	4349.0
365.90	562.0
371.90	1219.0
373.00	356.0
382.90	415.0
402.00	650.0
402.90	844.0
403.90	304.0
421.00	584.0
421.90	567.0
423.00	4265.0
423.90	1648.0
424.90	370.0
441.00	10456.0
442.00	68144.0
443.00	13041.0
443.90	1076.0

Instrument :
BNA_N
ClientSampleId :
DFTPP03

m/z	Abundance
38.00	906.0
39.10	5929.0
40.00	1498.0
41.00	624.0
44.00	4143.0
45.00	370.0
49.10	748.0
50.00	19072.0
51.00	76824.0
52.00	3934.0
53.00	318.0
55.00	764.0
55.90	2582.0
57.00	5861.0
61.00	1151.0
62.00	1361.0
63.00	3377.0
64.10	510.0
65.00	1713.0
66.90	320.0
68.10	1983.0
69.00	98952.0
70.00	490.0
73.00	2187.0
74.00	9806.0
75.00	15918.0
76.00	5238.0
77.00	107720.0
78.00	8013.0
79.00	7736.0
80.00	5535.0
81.00	7792.0
82.00	2018.0
83.00	1918.0
83.90	568.0
85.00	1242.0
86.00	2594.0
86.90	974.0
88.00	324.0
91.00	1784.0
92.00	1934.0
93.00	11847.0
94.00	826.0
96.00	1224.0
97.00	441.0
98.00	9314.0
99.00	8243.0
99.90	667.0
100.90	3835.0
103.00	1400.0
104.00	2641.0
105.00	2553.0
106.00	804.0
107.00	29504.0
108.00	4860.0
109.10	1059.0
110.00	55896.0
111.00	8622.0
111.90	985.0
113.00	327.0
116.00	1706.0
117.00	22072.0
118.00	2076.0
120.10	362.0
122.00	2159.0
123.00	3000.0
124.00	1487.0
125.00	1174.0
126.20	823.0
127.00	102800.0
128.00	7648.0
129.00	38744.0
130.00	3201.0
131.00	669.0
132.00	444.0
132.90	460.0
134.10	1007.0
135.00	3205.0
136.00	1296.0
137.00	2265.0
139.90	362.0
141.00	5492.0
141.90	1941.0
143.00	888.0
144.00	376.0
144.80	327.0
145.90	1003.0
147.00	2973.0
148.00	5423.0
149.00	1624.0
150.10	435.0
151.10	754.0

Instrument :
BNA_N
ClientSampleId :
DFTPP03

153.00	1471.0
153.90	1147.0
155.00	2527.0
156.00	4857.0
157.00	712.0
157.90	1048.0
159.00	853.0
160.00	1529.0
161.00	2519.0
162.00	728.0
164.10	330.0
165.00	2156.0
166.00	1490.0
167.00	10124.0
168.00	6883.0
169.00	1144.0
170.10	383.0
170.80	492.0
172.00	1460.0
172.90	1429.0
173.90	2258.0
175.00	4206.0
176.10	838.0
176.90	1737.0
178.00	805.0
178.90	7878.0
180.00	5424.0
181.00	2050.0
181.90	443.0
182.90	339.0
184.00	745.0
184.90	3993.0
186.00	29592.0
187.00	8810.0
188.00	1041.0
189.00	1949.0
189.90	353.0
190.90	1357.0
192.00	2369.0
193.00	3306.0
194.00	717.0
195.00	406.0
196.00	5868.0
198.00	215872.0
198.90	14753.0
200.00	1382.0
201.60	1227.0
202.80	1673.0
204.00	8005.0
205.00	13035.0
206.00	58744.0
207.00	11533.0
208.00	2773.0
208.90	948.0
209.90	820.0
210.90	2135.0
215.00	519.0
216.20	1279.0
217.00	16315.0
218.00	1995.0
220.30	575.0
221.00	9474.0
221.80	2358.0
223.00	3391.0
224.00	33032.0
225.00	8490.0
226.20	1082.0
227.00	12776.0
227.90	2021.0
228.90	2840.0
230.00	347.0
231.00	1301.0
233.00	371.0
234.00	778.0
235.00	1376.0
235.90	435.0
237.00	1088.0
238.90	424.0
240.00	567.0
240.90	591.0
242.00	1802.0
243.00	1760.0
244.00	25512.0
245.00	3545.0
246.00	5183.0
247.00	1205.0
249.00	1169.0
251.00	455.0
252.90	583.0
255.00	134208.0
256.00	19240.0
257.00	2017.0
258.00	7607.0
259.00	1451.0
265.00	2986.0
265.90	570.0

Instrument :
BNA_N
ClientSampleId :
DFTPP03

271.00	400.0
272.00	382.0
273.00	3743.0
274.00	10335.0
275.00	60632.0
276.00	7219.0
277.00	4862.0
278.00	1117.0
278.90	322.0
281.10	907.0
281.90	485.0
283.00	898.0
284.00	378.0
285.10	821.0
292.10	381.0
293.00	1153.0
294.90	490.0
296.00	15919.0
297.00	1852.0
301.90	383.0
303.00	1766.0
303.90	441.0
313.90	708.0
315.00	1973.0
315.90	1195.0
320.80	545.0
321.90	307.0
323.00	4466.0
324.00	663.0
327.00	1119.0
327.90	465.0
332.00	371.0
333.00	569.0
334.00	3705.0
335.00	774.0
340.90	729.0
345.90	1445.0
352.00	1324.0
353.00	1030.0
354.00	1979.0
355.00	306.0
364.90	6947.0
365.90	1153.0
371.00	548.0
372.00	2647.0
372.90	512.0
382.90	727.0
389.90	411.0
401.90	1012.0
402.90	1573.0
403.90	405.0
420.90	1430.0
422.00	1235.0
423.00	9096.0
424.00	2929.0
424.90	609.0
441.00	22312.0
442.00	143296.0
443.00	27800.0
444.10	2365.0

Instrument :
BNA_N
ClientSampleId :
DFTPP03

m/z	Abundance
37.00	330.0
38.00	1035.0
39.00	4414.0
40.00	1465.0
41.00	476.0
43.10	403.0
44.00	4395.0
49.10	736.0
50.00	14063.0
51.00	56376.0
52.00	2880.0
55.00	603.0
56.00	1678.0
57.00	4101.0
60.90	896.0
62.00	1051.0
63.10	2604.0
64.00	392.0
65.00	1322.0
67.10	351.0
69.00	73832.0
70.00	407.0
73.00	1670.0
74.00	7615.0
75.00	11835.0
76.00	3975.0
77.00	83168.0
78.10	5360.0
79.00	5600.0
80.00	4481.0
81.00	5370.0
82.00	1622.0
83.00	1225.0
83.90	507.0
85.00	971.0
85.90	1806.0
87.00	986.0
87.90	319.0
91.00	1515.0
91.90	1518.0
93.00	9470.0
93.90	801.0
95.00	418.0
95.90	811.0
97.00	394.0
98.00	6969.0
99.00	5506.0
99.90	354.0
101.00	3150.0
103.00	1375.0
104.00	1998.0
105.00	2378.0
106.00	794.0
107.00	22984.0
108.00	3629.0
109.10	729.0
110.00	43376.0
111.00	6427.0
111.90	539.0
116.00	1372.0
117.00	19248.0
118.00	1619.0
119.10	433.0
119.90	327.0
122.00	1758.0
123.00	2691.0
124.10	1325.0
124.90	1126.0
127.00	77904.0
128.00	5584.0
129.00	30920.0
129.90	2560.0
131.00	500.0
132.90	378.0
134.00	1163.0
135.00	2362.0
136.00	1429.0
136.90	1505.0
137.70	324.0
139.90	477.0
141.00	3684.0
142.00	1461.0
142.90	1075.0
145.00	305.0
146.00	839.0
146.90	2415.0
148.00	5347.0
148.90	881.0
151.10	451.0
151.70	359.0
153.00	1692.0
153.90	957.0

Instrument :
BNA_N
ClientSampleId :
DFTPP03

155.00	2417.0
156.00	3155.0
157.00	697.0
158.00	884.0
159.00	805.0
160.00	1226.0
161.00	1751.0
162.00	650.0
162.90	344.0
164.10	387.0
165.00	1789.0
166.00	1308.0
167.00	9016.0
168.00	5397.0
169.00	894.0
169.90	334.0
170.90	355.0
172.00	811.0
173.00	1046.0
174.00	1610.0
175.00	3601.0
176.00	1072.0
176.80	1669.0
177.90	846.0
179.00	6443.0
180.00	5087.0
180.90	1638.0
181.90	553.0
183.90	540.0
185.00	3209.0
186.00	24752.0
187.00	7474.0
187.90	760.0
188.90	1690.0
189.90	337.0
190.90	1193.0
192.00	2397.0
192.90	2463.0
193.90	681.0
196.00	5644.0
196.80	1598.0
197.90	191168.0
199.00	13314.0
200.00	984.0
201.30	857.0
202.90	1478.0
204.00	6798.0
205.00	11880.0
206.00	50336.0
207.00	9373.0
207.90	2192.0
209.00	1064.0
210.30	1224.0
211.00	2182.0
215.00	815.0
215.90	1088.0
217.00	14010.0
218.00	1690.0
221.00	8735.0
221.90	2439.0
222.90	3064.0
224.00	28968.0
225.00	6305.0
226.00	1067.0
227.00	12813.0
228.00	1703.0
229.00	2551.0
229.80	466.0
231.10	1048.0
234.00	976.0
234.90	701.0
235.90	798.0
237.00	1281.0
239.00	506.0
239.90	459.0
240.90	846.0
242.00	1681.0
243.00	1803.0
244.00	21192.0
245.00	3025.0
246.00	4568.0
246.90	938.0
248.00	306.0
249.00	980.0
252.90	652.0
255.00	119400.0
256.00	17144.0
257.00	1776.0
258.00	7200.0
259.00	1055.0
261.00	318.0
265.00	2680.0
266.00	384.0
266.90	342.0
270.90	350.0
272.00	344.0

Instrument :
BNA_N
ClientSampleId :
DFTPP03

273.00	3599.0
274.00	9207.0
275.00	53392.0
276.00	6567.0
277.00	4510.0
278.00	902.0
281.00	1159.0
283.00	691.0
284.00	388.0
284.90	916.0
293.00	1119.0
293.90	321.0
296.00	14151.0
296.90	1767.0
302.90	1839.0
304.00	363.0
314.10	490.0
315.00	1635.0
316.00	1283.0
320.90	449.0
322.10	327.0
323.00	4871.0
324.00	853.0
327.00	1037.0
327.80	509.0
332.00	402.0
333.00	477.0
334.00	3362.0
334.90	1066.0
341.00	588.0
345.90	938.0
352.00	1504.0
353.00	1250.0
353.90	1421.0
354.90	611.0
364.90	6957.0
366.00	1202.0
371.00	640.0
372.00	2387.0
373.00	630.0
383.00	850.0
383.80	340.0
390.00	310.0
400.90	332.0
402.00	821.0
403.00	1638.0
404.00	596.0
420.90	1140.0
421.90	1508.0
423.00	9659.0
424.00	3222.0
424.90	392.0
441.00	26216.0
442.00	155712.0
443.00	28712.0
444.00	2770.0

Instrument :
BNA_N
ClientSampleId :
DFTPP03