

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101623\
 Data File : BN028300.D
 Acq On : 16 Oct 2023 21:06
 Operator : MA/JU
 Sample : 04900-05
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AT7

Quant Time: Oct 17 02:11:14 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 16 16:16:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.887	152	2177	0.400	ng/ul	0.00
4) Naphthalene-d8	10.696	136	7383	0.400	ng/ul	-0.02
9) Acenaphthene-d10	14.536	164	4579	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.295	188	10416	0.400	ng/ul	0.00
17) Chrysene-d12	21.498	240	9727	0.400	ng/ul	0.00
23) Perylene-d12	23.948	264	11757	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.258	96	10077	3.685	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.297	152	4083	0.343	ng/ul	0.00
18) Fluoranthene-d10	19.329	212	10808	0.350	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101623\
 Data File : BN028300.D
 Acq On : 16 Oct 2023 21:06
 Operator : MA/JU
 Sample : 04900-05
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AT7

Quant Time: Oct 17 02:11:14 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 16 16:16:04 2023
 Response via : Initial Calibration

