

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101623\
Data File : BN028302.D
Acq On : 16 Oct 2023 22:19
Operator : MA/JU
Sample : 04900-07
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AW3

Quant Time: Oct 17 02:11:28 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 16 16:16:04 2023
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.886	152	2707	0.400	ng/u1	0.00
4) Naphthalene-d8	10.696	136	9157	0.400	ng/u1	-0.02
9) Acenaphthene-d10	14.536	164	5735	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.295	188	13063	0.400	ng/u1	0.00
17) Chrysene-d12	21.498	240	11934	0.400	ng/u1	0.00
23) Perylene-d12	23.945	264	14081	0.400	ng/u1	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.258	96	11834	3.480	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.308	152	4984	0.338	ng/u1	0.00
18) Fluoranthene-d10	19.329	212	14138	0.373	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101623\
Data File : BN028302.D
Acq On : 16 Oct 2023 22:19
Operator : MA/JU
Sample : 04900-07
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AW3

Quant Time: Oct 17 02:11:28 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 16 16:16:04 2023
Response via : Initial Calibration

