

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN101718\
 Data File : BN003193.D
 Acq On : 17 Oct 2018 22:58
 Operator : MJ/SJ
 Sample : J5364-04
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 KY038PS036-181008

Manual Integrations
 APPROVED

Sohil
 10/18/2018 4:41:07 PM

Quant Time: Oct 18 16:14:33 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN101718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 18 15:26:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.68	152	663	0.40	ng	0.04
7) Naphthalene-d8	10.49	136	3280	0.40	ng	0.06
13) Acenaphthene-d10	14.29	164	2728	0.40	ng	0.01
19) Phenanthrene-d10	17.04	188	7474	0.40	ng	-0.01
27) Chrysene-d12	21.25	240	10357	0.40	ng	0.00
34) Perylene-d12	23.47	264	11007	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.34	112	213	0.15	ng	0.04
5) Phenol-d6	7.02	99	180m	0.09	ng	0.08
8) Nitrobenzene-d5	9.03	82	734m	0.35	ng	0.17
11) 2-Methylnaphthalene-d10	12.09	152	1929	0.35	ng	0.05
14) 2,4,6-Tribromophenol	15.83	330	376	0.22	ng	0.00
15) 2-Fluorobiphenyl	12.94	172	3166	0.29	ng	0.01
25) Fluoranthene-d10	19.08	212	9105	0.39	ng	0.00
29) Terphenyl-d14	19.66	244	10595	0.47	ng	-0.01
Target Compounds						
2) 1,4-Dioxane	3.22	88	114	0.066	ng	# 46
23) Phenanthrene	17.09	178	1188	0.061	ng	# 98
32) Bis(2-ethylhexyl)phthalate	21.12	149	1204	0.081	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN101718\
Data File : BN003193.D
Acq On : 17 Oct 2018 22:58
Operator : MJ/SJ
Sample : J5364-04
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
KY038PS036-181008

Manual Integrations
APPROVED

Sohil
10/18/2018 4:41:07 PM

Quant Time: Oct 18 16:14:33 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN101718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 18 15:26:23 2018
Response via : Initial Calibration

