

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN101718\
 Data File : BN003179.D
 Acq On : 17 Oct 2018 13:23
 Operator : MJ/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN101718

Quant Time: Oct 18 15:30:51 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN101718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 18 15:26:23 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.658	0.594	9.7	109	0.00
3	n-Nitrosodimethylamine	0.194	0.147	24.2	70	0.02
4 S	2-Fluorophenol	0.883	0.876	0.8	92	0.00
5 S	Phenol-d6	1.259	1.144	9.1	94	0.00
6	bis(2-Chloroethyl)ether	1.056	0.949	10.1	82	0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
8 S	Nitrobenzene-d5	0.258	0.235	8.9	98	0.01
9	Naphthalene	0.996	0.955	4.1	97	0.00
10	Hexachlorobutadiene	0.195	0.187	4.1	97	0.00
11 SURR	2-Methylnaphthalene-d10	0.677	0.651	3.8	98	0.00
12	2-Methylnaphthalene	0.681	0.663	2.6	99	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
14 S	2,4,6-Tribromophenol	0.255	0.221	13.3	99	0.00
15 S	2-Fluorobiphenyl	1.575	1.481	6.0	99	0.00
16	Acenaphthylene	1.818	1.700	6.5	98	0.01
17	Acenaphthene	1.210	1.171	3.2	101	0.00
18	Fluorene	1.575	1.497	5.0	99	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
20	4-Bromophenyl-phenylether	0.223	0.209	6.3	99	0.00
21	Hexachlorobenzene	0.247	0.238	3.6	100	0.01
22	Pentachlorophenol	0.044	0.035	20.5	90	0.01
23	Phenanthrene	1.047	0.994	5.1	99	0.00
24	Anthracene	0.959	0.897	6.5	98	0.01
25 SURR	Fluoranthene-d10	1.240	1.166	6.0	99	0.00
26	Fluoranthene	1.342	1.269	5.4	99	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
28	Pvrene	1.119	1.048	6.3	99	0.00
29 S	Terphenyl-d14	0.867	0.794	8.4	99	0.00
30	Benzo(a)anthracene	1.075	0.910	15.3	96	0.00
31	Chrysene	1.229	1.203	2.1	99	0.00
32	Bis(2-ethylhexyl)phthalate	0.706	0.581	17.7	93	0.00
33	Indeno(1,2,3-cd)pyrene	1.453	1.279	12.0	94	0.00
34 I	Perylene-d12	1.000	1.000	0.0	97	0.00
35	Benzo(b)fluoranthene	1.080	1.023	5.3	100	0.00
36	Benzo(k)fluoranthene	1.275	1.206	5.4	96	0.00
37 C	Benzo(a)pyrene	1.114	1.033	7.3	96	0.00
38	Dibenzo(a,h)anthracene	1.138	1.023	10.1	94	0.00
39	Benzo(a,h,i)perylene	1.149	1.046	9.0	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN101718\
Data File : BN003179.D
Acq On : 17 Oct 2018 13:23
Operator : MJ/SJ
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN101718

Quant Time: Oct 18 15:30:51 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN101718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 18 15:26:23 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			