

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN101718\
 Data File : BN003181.D
 Acq On : 17 Oct 2018 15:50
 Operator : MJ/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 18 15:50:16 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN101718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 18 15:26:23 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	52	0.02
2	1,4-Dioxane	0.658	0.573	12.9	57	0.00
3	n-Nitrosodimethylamine	0.194	0.049	74.7#	13#	0.03
4 S	2-Fluorophenol	0.883	0.711	19.5	40#	0.04
5 S	Phenol-d6	1.259	1.021	18.9	45#	0.07
6	bis(2-Chloroethyl)ether	1.056	1.611	-52.6#	76	0.05
7 I	Naphthalene-d8	1.000	1.000	0.0	58	0.04
8 S	Nitrobenzene-d5	0.258	0.224	13.2	55	0.08
9	Naphthalene	0.996	0.927	6.9	56	0.04
10	Hexachlorobutadiene	0.195	0.186	4.6	57	0.00
11 SURR	2-Methylnaphthalene-d10	0.677	0.624	7.8	56	0.05
12	2-Methylnaphthalene	0.681	0.615	9.7	55	0.04
13 I	Acenaphthene-d10	1.000	1.000	0.0	63	0.01
14 S	2,4,6-Tribromophenol	0.255	0.205	19.6	58	0.02
15 S	2-Fluorobiphenyl	1.575	1.329	15.6	56	0.03
16	Acenaphthylene	1.818	1.553	14.6	56	0.02
17	Acenaphthene	1.210	1.106	8.6	60	0.01
18	Fluorene	1.575	1.414	10.2	59	0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	65	0.02
20	4-Bromophenyl-phenylether	0.223	0.181	18.8	56	0.02
21	Hexachlorobenzene	0.247	0.216	12.6	59	0.01
22	Pentachlorophenol	0.044	0.038	13.6	63	0.04
23	Phenanthrene	1.047	0.923	11.8	60	0.01
24	Anthracene	0.959	0.814	15.1	58	0.03
25 SURR	Fluoranthene-d10	1.240	1.153	7.0	63	0.01
26	Fluoranthene	1.342	1.244	7.3	63	0.02
27 I	Chrysene-d12	1.000	1.000	0.0	72	0.01
28	Pvrene	1.119	0.921	17.7	64	0.01
29 S	Terphenyl-d14	0.867	0.699	19.4	64	0.01
30	Benzo(a)anthracene	1.075	0.860	20.0	66	0.01
31	Chrysene	1.229	1.130	8.1	68	0.01
32	Bis(2-ethylhexyl)phthalate	0.706	0.519	26.5#	61	0.00
33	Indeno(1,2,3-cd)pyrene	1.453	1.419	2.3	77	0.02
34 I	Perylene-d12	1.000	1.000	0.0	78	0.01
35	Benzo(b)fluoranthene	1.080	0.912	15.6	72	0.00
36	Benzo(k)fluoranthene	1.275	1.159	9.1	74	0.00
37 C	Benzo(a)pyrene	1.114	0.987	11.4	74	0.01
38	Dibenzo(a,h)anthracene	1.138	1.048	7.9	77	0.02
39	Benzo(a,h,i)perylene	1.149	1.071	6.8	78	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN101718\
Data File : BN003181.D
Acq On : 17 Oct 2018 15:50
Operator : MJ/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 18 15:50:16 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN101718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 18 15:26:23 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			