

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN101718\
 Data File : BN003196.D
 Acq On : 18 Oct 2018 00:43
 Operator : MJ/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 18 15:41:21 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN101718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 18 15:26:23 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
2	1,4-Dioxane	0.658	0.687	-4.4	105	0.00
3	n-Nitrosodimethylamine	0.194	0.101	47.9#	40#	0.02
4 S	2-Fluorophenol	0.883	0.845	4.3	74	0.02
5 S	Phenol-d6	1.259	1.168	7.2	80	0.02
6	bis(2-Chloroethyl)ether	1.056	1.016	3.8	73	0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	86	0.01
8 S	Nitrobenzene-d5	0.258	0.241	6.6	89	0.03
9	Naphthalene	0.996	1.064	-6.8	95	0.01
10	Hexachlorobutadiene	0.195	0.211	-8.2	97	0.00
11 SURR	2-Methylnaphthalene-d10	0.677	0.725	-7.1	97	0.00
12	2-Methylnaphthalene	0.681	0.739	-8.5	98	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	92	0.00
14 S	2,4,6-Tribromophenol	0.255	0.251	1.6	104	0.00
15 S	2-Fluorobiphenyl	1.575	1.565	0.6	96	0.00
16	Acenaphthylene	1.818	1.887	-3.8	100	0.01
17	Acenaphthene	1.210	1.282	-6.0	102	0.00
18	Fluorene	1.575	1.663	-5.6	101	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
20	4-Bromophenyl-phenylether	0.223	0.223	0.0	99	0.00
21	Hexachlorobenzene	0.247	0.259	-4.9	103	0.01
22	Pentachlorophenol	0.044	0.043	2.3	103	0.01
23	Phenanthrene	1.047	1.099	-5.0	103	0.00
24	Anthracene	0.959	0.994	-3.6	103	0.01
25 SURR	Fluoranthene-d10	1.240	1.308	-5.5	104	0.00
26	Fluoranthene	1.342	1.419	-5.7	104	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	90	0.00
28	Pvrene	1.119	1.205	-7.7	105	0.00
29 S	Terphenyl-d14	0.867	0.900	-3.8	102	0.00
30	Benzo(a)anthracene	1.075	1.053	2.0	102	0.00
31	Chrysene	1.229	1.342	-9.2	101	0.00
32	Bis(2-ethylhexyl)phthalate	0.706	0.623	11.8	92	0.00
33	Indeno(1,2,3-cd)pyrene	1.453	1.438	1.0	97	0.00
34 I	Perylene-d12	1.000	1.000	0.0	90	0.00
35	Benzo(b)fluoranthene	1.080	1.160	-7.4	106	0.00
36	Benzo(k)fluoranthene	1.275	1.368	-7.3	101	0.00
37 C	Benzo(a)pyrene	1.114	1.171	-5.1	101	0.00
38	Dibenzo(a,h)anthracene	1.138	1.140	-0.2	97	0.00
39	Benzo(a,h,i)perylene	1.149	1.159	-0.9	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN101718\
Data File : BN003196.D
Acq On : 18 Oct 2018 00:43
Operator : MJ/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 18 15:41:21 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN101718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 18 15:26:23 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			