

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN101818\
 Data File : BN003199.D
 Acq On : 18 Oct 2018 10:03
 Operator : MJ/SJ
 Sample : J5317-17DL 2X
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE-109D1-20181005DL

Manual Integrations
 APPROVED

Sohil
 10/19/2018 3:55:34 PM

Quant Time: Oct 19 02:20:56 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 18 15:26:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.66	152	954	0.40	ng	0.02
7) Naphthalene-d8	10.52	136	3947m	0.40	ng	0.09
13) Acenaphthene-d10	14.29	164	3260	0.40	ng	0.01
19) Phenanthrene-d10	17.07	188	8852	0.40	ng	0.02
27) Chrysene-d12	21.26	240	11111	0.40	ng	0.01
34) Perylene-d12	23.47	264	11219	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.37	112	134m	0.06	ng	0.07
5) Phenol-d6	6.89	99	160	0.05	ng	-0.05
8) Nitrobenzene-d5	9.08	82	305m	0.12	ng	0.22
11) 2-Methylnaphthalene-d10	12.13	152	1264	0.19	ng	0.09
14) 2,4,6-Tribromophenol	15.86	330	350	0.17	ng	0.03
15) 2-Fluorobiphenyl	12.96	172	2277	0.18	ng	0.03
25) Fluoranthene-d10	19.09	212	5348	0.19	ng	0.01
29) Terphenyl-d14	19.68	244	4968	0.21	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.22	88	2127	1.740	ng	94
32) Bis(2-ethylhexyl)phthalate	21.13	149	1388	0.086	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN101818\
Data File : BN003199.D
Acq On : 18 Oct 2018 10:03
Operator : MJ/SJ
Sample : J5317-17DL 2X
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE-109D1-20181005DL

Manual Integrations
APPROVED

Sohil
10/19/2018 3:55:34 PM

Quant Time: Oct 19 02:20:56 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 18 15:26:23 2018
Response via : Initial Calibration

