

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101818\
 Data File : BN003198.D
 Acq On : 18 Oct 2018 09:26
 Operator : MJ/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 19 02:18:12 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 17 13:21:03 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
2	1,4-Dioxane	0.658	0.606	7.9	94	0.00
3	n-Nitrosodimethylamine	0.194	0.178	8.2	71	0.00
4 S	2-Fluorophenol	0.883	0.883	0.0	79	0.02
5 S	Phenol-d6	1.259	1.175	6.7	82	0.03
6	bis(2-Chloroethyl)ether	1.056	1.004	4.9	74	0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	83	0.01
8 S	Nitrobenzene-d5	0.258	0.257	0.4	91	0.03
9	Naphthalene	0.996	1.093	-9.7	95	0.01
10	Hexachlorobutadiene	0.195	0.220	-12.8	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.677	0.703	-3.8	90	0.02
12	2-Methylnaphthalene	0.681	0.736	-8.1	94	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	86	0.00
14 S	2,4,6-Tribromophenol	0.255	0.250	2.0	96	0.00
15 S	2-Fluorobiphenyl	1.575	1.671	-6.1	96	0.00
16	Acenaphthylene	1.818	1.894	-4.2	94	0.01
17	Acenaphthene	1.210	1.300	-7.4	96	0.00
18	Fluorene	1.575	1.691	-7.4	96	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.00
20	4-Bromophenyl-phenylether	0.223	0.220	1.3	92	0.01
21	Hexachlorobenzene	0.247	0.263	-6.5	98	0.01
22	Pentachlorophenol	0.044	0.044	0.0	98	0.01
23	Phenanthrene	1.047	1.112	-6.2	97	0.00
24	Anthracene	0.959	0.984	-2.6	95	0.01
25 SURR	Fluoranthene-d10	1.240	1.295	-4.4	96	0.00
26	Fluoranthene	1.342	1.438	-7.2	99	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	86	0.00
28	Pyrene	1.119	1.202	-7.4	99	0.00
29 S	Terphenyl-d14	0.867	0.906	-4.5	98	0.00
30	Benzo(a)anthracene	1.075	1.037	3.5	95	0.00
31	Chrysene	1.229	1.406	-14.4	101	0.00
32	Bis(2-ethylhexyl)phthalate	0.706	0.612	13.3	86	0.00
33	Indeno(1,2,3-cd)pyrene	1.453	1.439	1.0	93	0.00
34 I	Perylene-d12	1.000	1.000	0.0	85	0.00
35	Benzo(b)fluoranthene	1.080	1.164	-7.8	100	0.00
36	Benzo(k)fluoranthene	1.275	1.395	-9.4	97	0.00
37 C	Benzo(a)pyrene	1.114	1.184	-6.3	97	0.00
38	Dibenzo(a,h)anthracene	1.138	1.149	-1.0	92	0.00
39	Benzo(g,h,i)perylene	1.149	1.173	-2.1	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101818\
Data File : BN003198.D
Acq On : 18 Oct 2018 09:26
Operator : MJ/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 19 02:18:12 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 17 13:21:03 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		