

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101818\
 Data File : BN003203.D
 Acq On : 18 Oct 2018 13:30
 Operator : MJ/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 19 02:27:46 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 18 15:26:23 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	-0.02
2	1,4-Dioxane	0.658	0.501	23.9	98	0.00
3	n-Nitrosodimethylamine	0.194	0.248	-27.8#	125	0.00
4 S	2-Fluorophenol	0.883	1.018	-15.3	114	-0.02
5 S	Phenol-d6	1.259	1.310	-4.1	114	-0.05
6	bis(2-Chloroethyl)ether	1.056	1.360	-28.8#	126	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	101	-0.01
8 S	Nitrobenzene-d5	0.258	0.279	-8.1	120	-0.01
9	Naphthalene	0.996	1.101	-10.5	116	-0.01
10	Hexachlorobutadiene	0.195	0.215	-10.3	116	-0.01
11 SURR	2-Methylnaphthalene-d10	0.677	0.699	-3.2	109	0.00
12	2-Methylnaphthalene	0.681	0.736	-8.1	114	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	102	-0.01
14 S	2,4,6-Tribromophenol	0.255	0.266	-4.3	121	0.00
15 S	2-Fluorobiphenyl	1.575	1.695	-7.6	115	-0.01
16	Acenaphthylene	1.818	1.952	-7.4	114	0.00
17	Acenaphthene	1.210	1.335	-10.3	117	-0.01
18	Fluorene	1.575	1.710	-8.6	115	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	104	-0.01
20	4-Bromophenyl-phenylether	0.223	0.229	-2.7	113	0.00
21	Hexachlorobenzene	0.247	0.257	-4.0	113	0.00
22	Pentachlorophenol	0.044	0.052	-18.2	139	0.00
23	Phenanthrene	1.047	1.126	-7.5	117	-0.01
24	Anthracene	0.959	1.019	-6.3	116	0.00
25 SURR	Fluoranthene-d10	1.240	1.320	-6.5	117	0.00
26	Fluoranthene	1.342	1.459	-8.7	119	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
28	Pyrene	1.119	1.168	-4.4	120	0.00
29 S	Terphenyl-d14	0.867	0.890	-2.7	120	0.00
30	Benzo(a)anthracene	1.075	1.077	-0.2	123	0.00
31	Chrysene	1.229	1.348	-9.7	120	0.00
32	Bis(2-ethylhexyl)phthalate	0.706	0.595	15.7	104	0.00
33	Indeno(1,2,3-cd)pyrene	1.453	1.632	-12.3	130	0.00
34 I	Perylene-d12	1.000	1.000	0.0	112	0.00
35	Benzo(b)fluoranthene	1.080	1.176	-8.9	133	0.00
36	Benzo(k)fluoranthene	1.275	1.359	-6.6	124	0.00
37 C	Benzo(a)pyrene	1.114	1.188	-6.6	128	0.00
38	Dibenzo(a,h)anthracene	1.138	1.233	-8.3	130	0.00
39	Benzo(g,h,i)perylene	1.149	1.274	-10.9	133	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101818\
Data File : BN003203.D
Acq On : 18 Oct 2018 13:30
Operator : MJ/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 19 02:27:46 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 18 15:26:23 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		