

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN101918\
 Data File : BN003215.D
 Acq On : 19 Oct 2018 20:12
 Operator : MJ/SJ
 Sample : PB114024BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK24

Manual Integrations
 APPROVED

Sohil
 10/22/2018 4:52:47 PM

Quant Time: Oct 20 05:25:05 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN101918MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 20 05:04:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.61	152	85487	20.00	ng/ul	0.00
18) Naphthalene-d8	10.38	136	410373	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.25	164	263862	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.00	188	617410	20.00	ng/ul	0.00
77) Chrysene-d12	21.22	240	715528	20.00	ng/ul	0.01
85) Perylene-d12	23.43	264	774317	20.00	ng/ul	0.01

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.18	96	10997	6.64	ng/uL	0.00
5) Phenol-d5	6.80	99	221288	27.33	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.95	67	148780	31.40	ng/ul	0.00
9) 2-Chlorophenol-d4	7.15	132	190917	30.22	ng/ul	0.00
13) 4-Methylphenol-d8	8.33	113	190951	28.97	ng/ul	0.00
19) Nitrobenzene-d5	8.76	128	97505	30.62	ng/ul	0.00
22) 2-Nitrophenol-d4	9.48	143	116199	31.45	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.02	165	184484	27.54	ng/ul	0.00
29) 4-Chloroaniline-d4	10.53	131	269127	49.83	ng/ul	0.00
43) Dimethylphthalate-d6	13.66	166	767249	34.12	ng/ul	0.00
46) Acenaphthylene-d8	13.94	160	900064	32.42	ng/ul	0.00
51) 4-Nitrophenol-d4	14.51	143	100786m	32.53	ng/ul	0.00
57) Fluorene-d10	15.25	176	652810	33.14	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.41	200	112829	27.21	ng/ul	0.00
70) Anthracene-d10	17.10	188	1042316	34.10	ng/ul	0.00
78) Pyrene-d10	19.41	212	1179292	34.01	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.29	264	1453350	34.72	ng/ul	0.01

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN101918\
Data File : BN003215.D
Acq On : 19 Oct 2018 20:12
Operator : MJ/SJ
Sample : PB114024BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK24

Quant Time: Oct 20 05:25:05 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN101918MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Oct 20 05:04:06 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

Sohil
10/22/2018 4:52:47 PM

