


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN101918\
Data File : BN003204.D
Acq On : 19 Oct 2018 13:45
Operator : MJ/SJ
Sample : DFTPP39
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP39

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN101918MA.M
Title : SVOA CALIBRATION
Last Update : Mon Oct 22 04:05:35 2018



AutoFind: Scans 2458, 2459, 2460; Background Corrected with Scan 2452

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.5	46581	PASS
68	69	0.00	2	1.9	947	PASS
69	198	0.00	100	35.1	50242	PASS
70	69	0.00	2	0.5	232	PASS
127	198	10	80	44.5	63709	PASS
197	198	0.00	2	0.3	388	PASS
198	198	100	100	100.0	143141	PASS
199	198	5	9	6.8	9684	PASS
275	198	10	60	28.5	40853	PASS
365	198	1	100	3.6	5180	PASS
441	198	0.01	100	14.6	20834	PASS
442	198	50	100	94.0	134509	PASS
443	442	15	24	18.9	25482	PASS

m/z	Abundance
37.90	876.0
39.10	4067.0
39.90	490.0
44.00	1412.0
49.10	380.0
50.00	14310.0
51.00	59224.0
52.00	3373.0
56.00	1857.0
56.90	4206.0
60.90	766.0
61.90	917.0
63.00	2554.0
63.90	485.0
65.00	1198.0
68.00	1191.0
69.00	61464.0
70.10	322.0
73.10	554.0
74.00	6313.0
75.00	9810.0
76.00	3196.0
77.00	70984.0
78.00	5054.0
79.00	4091.0
79.90	3315.0
81.00	5481.0
81.90	1407.0
83.00	980.0
84.90	1154.0
85.90	1536.0
87.00	580.0
90.90	1159.0
92.00	1276.0
93.00	7480.0
94.00	372.0
95.90	390.0
97.90	4971.0
99.00	4462.0
101.00	2661.0
103.00	871.0
103.90	1729.0
105.00	1320.0
106.00	551.0
107.00	19720.0
108.00	3241.0
110.00	38528.0
111.00	5656.0
111.90	595.0
116.00	1169.0
117.00	14680.0
117.90	1453.0
120.00	355.0
121.90	1368.0
123.00	1988.0
123.90	973.0
124.90	850.0
127.00	75664.0
128.00	5390.0
129.00	28840.0
129.90	2336.0
131.00	478.0
132.00	339.0
133.90	705.0
135.00	2425.0
136.00	780.0
137.00	1104.0
137.80	396.0
140.90	3012.0
141.90	1119.0
143.00	825.0
145.80	589.0
147.00	1430.0
148.00	4371.0
149.00	697.0
151.10	404.0
151.90	401.0
152.90	1096.0
153.80	857.0
155.00	1745.0
156.00	2696.0
157.20	475.0
157.90	739.0
158.90	643.0
160.00	1281.0
160.90	1495.0
161.90	340.0
165.00	1359.0
166.00	946.0
167.00	6798.0
168.00	3974.0

Instrument :
BNA_N
ClientSampleId :
DFTPP39

168.90	683.0
170.80	365.0
171.90	602.0
172.90	754.0
173.90	1458.0
175.00	3035.0
176.00	1103.0
176.80	1263.0
178.00	384.0
178.90	5384.0
180.00	4108.0
181.00	1819.0
181.90	446.0
183.80	353.0
185.00	2377.0
186.00	19440.0
187.00	6428.0
187.90	624.0
188.90	1312.0
190.90	540.0
192.00	1655.0
193.00	1969.0
194.00	476.0
196.00	4386.0
197.90	159808.0
198.90	11223.0
199.90	692.0
201.40	850.0
202.90	1226.0
204.00	5064.0
205.00	8434.0
206.00	41160.0
207.00	5987.0
207.90	1332.0
209.00	430.0
210.90	1241.0
212.10	358.0
216.00	769.0
216.90	9329.0
218.00	1559.0
221.00	7698.0
221.80	1461.0
223.00	2352.0
224.00	21480.0
225.00	5455.0
226.00	837.0
227.00	8946.0
227.90	1470.0
229.00	2018.0
231.10	1100.0
233.90	358.0
234.00	356.0
234.90	709.0
235.90	426.0
236.90	974.0
238.90	310.0
241.00	503.0
242.00	1239.0
243.00	1092.0
244.00	18256.0
245.00	2350.0
245.90	2960.0
246.90	555.0
248.80	673.0
253.10	450.0
253.90	588.0
255.00	88368.0
256.00	13489.0
257.00	1073.0
258.00	4531.0
259.00	827.0
264.90	1782.0
273.00	2454.0
274.00	7151.0
275.00	42176.0
276.00	5941.0
276.90	3361.0
277.90	458.0
282.90	370.0
283.90	312.0
285.10	636.0
293.00	738.0
295.90	9594.0
297.00	1508.0
303.00	1272.0
303.90	406.0
314.10	530.0
314.80	940.0
316.00	589.0
321.00	469.0
323.00	3904.0
324.00	596.0
327.00	632.0
327.90	329.0
332.00	402.0

Instrument :
BNA_N
ClientSampleId :
DFTPP39

333.00	370.0
334.00	2380.0
335.00	625.0
341.00	558.0
346.00	572.0
352.00	1023.0
353.00	913.0
354.00	832.0
365.00	4895.0
366.00	690.0
370.80	348.0
372.00	2251.0
372.90	625.0
383.00	433.0
401.90	846.0
403.00	1289.0
403.80	436.0
421.00	890.0
422.00	987.0
423.00	7361.0
424.00	2067.0
424.90	368.0
441.00	17320.0
442.10	114560.0
443.10	22904.0
444.00	1710.0

Instrument :
BNA_N
ClientSampleId :
DFTPP39

m/z	Abundance
38.00	538.0
39.00	3870.0
40.10	519.0
44.00	1715.0
49.00	339.0
50.00	12498.0
51.00	53624.0
52.00	3088.0
54.90	359.0
55.90	1941.0
57.00	4150.0
61.00	659.0
61.90	721.0
63.00	2283.0
65.10	1267.0
68.00	1043.0
69.00	58120.0
70.00	376.0
73.00	540.0
74.00	5501.0
75.00	9393.0
76.00	3110.0
77.00	66192.0
78.00	4119.0
79.00	4169.0
80.00	3262.0
81.00	4427.0
81.90	1066.0
83.00	1297.0
85.00	895.0
86.00	1295.0
87.00	439.0
90.90	845.0
91.90	1229.0
93.00	6975.0
94.00	541.0
96.00	409.0
98.00	5352.0
99.00	4422.0
99.90	563.0
101.00	2636.0
102.90	739.0
103.90	1590.0
105.00	1307.0
106.00	576.0
107.00	19256.0
108.00	2927.0
109.00	649.0
110.00	36568.0
111.00	5161.0
111.90	717.0
116.00	1142.0
117.00	13649.0
118.00	965.0
121.90	1092.0
122.90	1925.0
123.90	905.0
125.00	779.0
127.00	74648.0
128.00	5589.0
129.00	27232.0
130.00	1867.0
131.00	387.0
134.00	948.0
135.00	2320.0
135.90	929.0
136.90	1043.0
137.60	350.0
140.00	425.0
141.00	3602.0
142.00	1142.0
143.00	800.0
146.00	748.0
146.90	2092.0
148.00	3899.0
148.90	1110.0
150.00	328.0
151.00	633.0
152.90	1205.0
154.00	872.0
154.90	2349.0
156.00	2816.0
157.00	576.0
157.80	641.0
159.00	575.0
159.90	1206.0
161.00	2081.0
161.90	592.0
163.90	312.0
164.90	1395.0
166.00	1189.0

Instrument :
BNA_N
ClientSampleId :
DFTPP39

167.00	7244.0
168.00	3645.0
169.00	539.0
170.90	393.0
171.90	871.0
172.90	896.0
173.90	1653.0
175.00	2874.0
176.10	1081.0
176.90	1357.0
177.80	599.0
179.00	5901.0
180.00	4372.0
180.90	1955.0
183.90	525.0
185.00	3026.0
186.00	21104.0
187.00	6255.0
188.00	743.0
189.00	1352.0
190.90	617.0
191.90	1794.0
193.00	2030.0
194.10	470.0
195.10	353.0
196.00	4690.0
197.00	1165.0
197.90	169600.0
198.90	11291.0
199.90	919.0
201.50	1084.0
203.00	954.0
204.00	5692.0
205.00	10137.0
206.00	45800.0
207.00	5980.0
208.00	1295.0
209.00	480.0
210.00	584.0
211.00	1527.0
214.90	483.0
216.00	957.0
216.90	10945.0
217.90	1358.0
221.00	8266.0
221.80	1797.0
222.90	2694.0
224.00	24688.0
225.00	6046.0
225.90	690.0
227.00	8530.0
227.90	1186.0
229.00	1932.0
231.00	626.0
233.90	553.0
234.90	821.0
235.80	513.0
236.90	1050.0
238.90	419.0
239.90	355.0
241.00	381.0
241.90	1077.0
243.00	1287.0
244.00	19880.0
245.00	2357.0
246.00	3714.0
246.90	652.0
248.90	669.0
255.00	105976.0
256.00	14717.0
257.00	1523.0
257.90	5464.0
258.90	1080.0
264.90	2275.0
273.00	3246.0
274.00	7400.0
275.00	48736.0
276.00	7073.0
277.00	3475.0
278.00	833.0
282.90	382.0
284.90	591.0
292.90	1036.0
296.00	12369.0
296.90	1613.0
303.00	1784.0
304.00	374.0
314.00	691.0
314.90	1248.0
315.90	821.0
320.80	327.0
323.00	4573.0
324.00	749.0
326.90	713.0
327.90	413.0

Instrument :
BNA_N
ClientSampleId :
DFTPP39

332.00	361.0
333.00	533.0
334.00	3062.0
335.00	802.0
341.00	461.0
346.10	999.0
352.00	1280.0
353.00	1080.0
354.00	1315.0
355.00	462.0
364.90	5709.0
365.90	824.0
370.90	441.0
372.00	2407.0
372.90	660.0
382.90	714.0
390.00	352.0
402.00	1358.0
403.00	1747.0
404.00	453.0
421.00	1390.0
422.00	1163.0
423.00	9064.0
424.00	3063.0
424.90	362.0
441.00	26232.0
442.00	164736.0
443.00	31472.0
444.00	2602.0

Instrument :
BNA_N
ClientSampleId :
DFTPP39

m/z	Abundance
38.10	311.0
39.00	1996.0
39.90	416.0
44.00	1527.0
50.00	7109.0
51.00	26896.0
52.00	1559.0
55.90	967.0
57.00	1994.0
61.00	364.0
62.10	439.0
63.00	1188.0
65.00	749.0
68.10	609.0
69.00	31144.0
74.00	2749.0
75.00	4590.0
76.00	1515.0
77.00	34520.0
78.10	2095.0
79.00	2182.0
80.00	1621.0
81.00	2842.0
82.00	542.0
82.90	535.0
84.90	517.0
86.00	733.0
86.90	338.0
90.90	516.0
91.90	489.0
93.00	3708.0
94.00	352.0
97.90	3046.0
99.00	2304.0
100.90	1376.0
102.90	479.0
103.90	664.0
104.90	818.0
107.00	10863.0
107.90	1505.0
109.10	426.0
110.00	21632.0
111.00	3045.0
111.80	501.0
115.90	731.0
117.00	7308.0
118.00	444.0
122.00	662.0
123.00	1407.0
124.00	629.0
124.80	368.0
127.00	40816.0
128.00	2854.0
129.00	15364.0
130.00	1427.0
134.90	941.0
135.90	522.0
137.00	694.0
141.00	2026.0
142.00	660.0
142.90	435.0
146.00	383.0
146.90	873.0
147.90	1750.0
148.90	408.0
150.90	408.0
153.00	710.0
153.90	568.0
155.10	1043.0
156.00	1476.0
157.10	325.0
158.00	379.0
160.00	713.0
161.00	981.0
165.00	561.0
165.90	556.0
167.00	4528.0
167.90	2162.0
169.00	388.0
171.90	357.0
172.90	510.0
174.00	778.0
174.90	1721.0
176.00	423.0
176.90	808.0
179.00	3122.0
179.90	2268.0
180.90	1099.0
185.00	1435.0
186.00	12791.0
187.00	3655.0

Instrument :
BNA_N
ClientSampleId :
DFTPP39

188.90	619.0
191.00	452.0
191.90	1004.0
193.00	999.0
194.10	308.0
195.90	3267.0
197.90	100016.0
198.90	6539.0
199.90	619.0
202.90	653.0
204.00	3297.0
205.00	6448.0
206.00	27760.0
207.00	3632.0
208.00	890.0
210.10	590.0
211.00	968.0
214.90	344.0
216.00	563.0
216.90	6972.0
217.90	769.0
221.00	5174.0
221.60	1047.0
222.90	1742.0
224.00	13526.0
225.00	4197.0
226.10	475.0
227.00	6118.0
227.90	835.0
228.90	1212.0
230.90	534.0
233.80	382.0
234.80	369.0
237.00	579.0
238.80	350.0
240.90	337.0
241.90	867.0
242.90	727.0
244.00	12107.0
245.00	1704.0
246.00	2219.0
246.80	350.0
248.90	539.0
254.10	519.0
255.00	67528.0
256.00	8805.0
257.00	707.0
258.00	3213.0
258.90	632.0
264.90	1384.0
273.00	1832.0
274.00	4942.0
275.00	31648.0
276.00	4052.0
276.90	2562.0
278.00	496.0
282.90	329.0
284.90	512.0
292.90	642.0
296.00	8447.0
297.00	1106.0
303.00	948.0
314.00	436.0
314.90	946.0
315.90	531.0
321.00	522.0
323.00	3368.0
324.10	545.0
327.00	715.0
327.90	324.0
334.00	1633.0
335.10	585.0
346.00	651.0
351.90	771.0
352.90	654.0
353.90	1097.0
354.90	343.0
364.90	4937.0
366.00	665.0
372.00	2111.0
373.00	422.0
383.00	513.0
389.80	424.0
401.90	836.0
403.00	1043.0
404.00	587.0
421.00	927.0
422.00	1057.0
423.00	7443.0
424.00	1875.0
425.00	328.0
441.00	18952.0
442.00	124232.0
443.00	22072.0
444.00	1944.0

Instrument :
BNA_N
ClientSampleId :
DFTPP39

Scan 2452 (17.410 min): BN003204.D
DFTPP39

m/z	Abundance
39.90	356.0
43.90	1864.0
206.90	371.0

Instrument :
BNA_N
ClientSampleId :
DFTPP39