

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102023\
 Data File : BN028335.D
 Acq On : 20 Oct 2023 11:24
 Operator : MA/JU
 Sample : PB156467BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK467

Quant Time: Oct 20 12:00:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 16 16:16:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.891	152	3298	0.400	ng/ul	0.00
4) Naphthalene-d8	10.719	136	11330	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.541	164	7231	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.304	188	17235	0.400	ng/ul	0.00
17) Chrysene-d12	21.495	240	17812	0.400	ng/ul	0.00
23) Perylene-d12	23.948	264	17868	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.263	96	2859	0.690	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.308	152	5678	0.311	ng/ul	0.00
18) Fluoranthene-d10	19.330	212	18313	0.324	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102023\
Data File : BN028335.D
Acq On : 20 Oct 2023 11:24
Operator : MA/JU
Sample : PB156467BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK467

Quant Time: Oct 20 12:00:45 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN100323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 16 16:16:04 2023
Response via : Initial Calibration

