




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102025\
Data File : BN038100.D
Acq On : 20 Oct 2025 16:41
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 20 18:28:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 20 18:27:35 2025
Response via : Initial Calibration



TIC: BN038100.D\data.ms

(70) Pentachlorophenol (C)




16.851min (0.000) 277.76 ng

response 700210

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	62.64
264.00	61.60	54.39
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102025\
Data File : BN038100.D
Acq On : 20 Oct 2025 16:41
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 20 18:28:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 20 18:27:35 2025
Response via : Initial Calibration



TIC: BN038100.D\data.ms

(77) Benzidine

19.451min (0.000) 0.00 ng

response 2449641

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	15.05
183.00	13.20	12.03
0.00	0.00	0.00