

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102025\
 Data File : BN038109.D
 Acq On : 20 Oct 2025 22:10
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 21 02:30:38 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 21 02:23:15 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	135	0.00
2	1,4-Dioxane	0.400	0.479	-19.7	161	0.00
3	n-Nitrosodimethylamine	0.400	0.401	-0.3	142	0.00
4 S	2-Fluorophenol	0.400	0.399	0.3	144	0.00
5 S	Phenol-d6	0.400	0.339	15.3	123	0.00
6	bis(2-Chloroethyl)ether	0.400	0.387	3.3	135	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	132	0.00
8 S	Nitrobenzene-d5	0.400	0.358	10.5	125	0.00
9	Naphthalene	0.400	0.391	2.3	136	0.00
10	Hexachlorobutadiene	0.400	0.415	-3.7	142	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.374	6.5	137	0.00
12	2-Methylnaphthalene	0.400	0.372	7.0	133	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	133	0.00
14 S	2,4,6-Tribromophenol	0.400	0.333	16.8	124	0.00
15 S	2-Fluorobiphenyl	0.400	0.373	6.8	130	0.00
16	Acenaphthylene	0.400	0.380	5.0	138	0.00
17	Acenaphthene	0.400	0.374	6.5	134	0.00
18	Fluorene	0.400	0.386	3.5	138	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	124	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.379	5.3	114	0.00
21	4-Bromophenyl-phenylether	0.400	0.387	3.3	128	0.00
22	Hexachlorobenzene	0.400	0.393	1.8	127	0.00
23	Atrazine	0.400	0.329	17.8	118	-0.01
24	Pentachlorophenol	0.400	0.330	17.5	121	0.00
25	Phenanthrene	0.400	0.366	8.5	122	0.00
26	Anthracene	0.400	0.358	10.5	124	0.00
27 SURR	Fluoranthene-d10	0.400	0.354	11.5	125	0.00
28	Fluoranthene	0.400	0.347	13.3	119	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	104	0.00
30	Pyrene	0.400	0.440	-10.0	119	0.00
31 S	Terphenyl-d14	0.400	0.429	-7.2	119	0.00
32	Benzo(a)anthracene	0.400	0.370	7.5	104	0.00
33	Chrysene	0.400	0.413	-3.2	111	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.475	-18.7	141	0.00
35 I	Perylene-d12	0.400	0.400	0.0	105	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.391	2.3	113	-0.01
37	Benzo(b)fluoranthene	0.400	0.371	7.3	102	0.00
38	Benzo(k)fluoranthene	0.400	0.377	5.8	107	0.00
39 C	Benzo(a)pyrene	0.400	0.383	4.3	111	0.00
40	Dibenzo(a,h)anthracene	0.400	0.354	11.5	102	0.00
41	Benzo(g,h,i)perylene	0.400	0.427	-6.7	120	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4